

Appendix C for KSCR230500078901

Calibration Certificate

Object	Apply	No	Model	SN	Calibration Date	Due date of calibration
Dipole	<input type="checkbox"/>	1	CLA150	4025	2021/04/26	2024/04/25
	<input type="checkbox"/>	2	D450V3	1103	2021/04/21	2024/04/20
	<input type="checkbox"/>	3	D750V3	1188	2022/03/29	2025/03/28
	<input type="checkbox"/>	4	D835V2	4d114	2022/03/31	2025/03/30
	<input type="checkbox"/>	5	D900V2	1d079	2022/06/07	2025/06/06
	<input type="checkbox"/>	6	D1800V2	2d170	2022/03/31	2025/03/30
	<input type="checkbox"/>	7	D1900V2	5d136	2022/06/07	2025/06/06
	<input type="checkbox"/>	8	D2000V2	1041	2022/06/06	2025/06/05
	<input type="checkbox"/>	9	D2300V2	1096	2022/03/31	2025/03/30
	<input checked="" type="checkbox"/>	10	D2450V2	817	2022/04/01	2025/03/31
	<input type="checkbox"/>	11	D2600V2	1158	2022/03/31	2025/03/30
	<input type="checkbox"/>	12	D5GHzV2	1095	2022/06/01	2025/05/31
DAE	<input checked="" type="checkbox"/>	13	DAE4	910	2022/07/14	2023/07/13
Probe	<input checked="" type="checkbox"/>	14	EX3DV4	7767	2022/10/28	2023/10/27



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/Terms-and-Conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10 Weiye Road, Development Zone, Kunshan, Jiangsu, China 215300 t(86-512) 5735 5888 f(86-512) 57370818 www.sgsgroup.com.cn
 中国·江苏·昆山开发区伟业路10号 邮编: 215300 t(86-512) 57355888 f(86-512) 57370818 sgs.china@sgs.com

1 Dipole

1.1 CLA150 - SN 4025

Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland		S Schweizerischer Kalibrierdienst Service suisse d'étalonnage C Servizio svizzero di taratura S Swiss Calibration Service	
Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates		Accreditation No.: SCS 0108	
Client: SGS-CN (Auden)		Certificate No: CLA150-4025_Apr21	
CALIBRATION CERTIFICATE			
Object: CLA150 - SN: 4025			
Calibration procedure(s): QA CAL-15-v9 Calibration Procedure for SAR Validation Sources below 700 MHz			
Calibration date: April 26, 2021			
This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.			
All calibrations have been conducted in the closed laboratory facility: environment temperature $22 \pm 3^\circ\text{C}$ and humidity $< 70\%$.			
Calibration Equipment used (M&TE critical for calibration)			
Primary Standards	ID #	Cal Date (Certificate No.)	Scheduled Calibration
Power meter NRP	SN: 10476	09-Apr-21 (No. 217-03201/03202)	Apr-22
Power sensor NRP Z91	SN: 10344	09-Apr-21 (No. 217-03201)	Apr-22
Power sensor NRP Z91	SN: 10345	09-Apr-21 (No. 217-03202)	Apr-22
Reference 20 dB Attenuator	SN: C22862 (203)	09-Apr-21 (No. 217-03343)	Apr-22
Type-N mismatch combination	SN: 310952 / 00357	09-Apr-21 (No. 217-03344)	Apr-22
Reference Probe EX30N4	SN: 3877	30-Dec-20 (No. EX3-3877_Dec20)	Dec-21
CNE4	SN: 684	26-Jan-20 (No. D458-684_Jan20)	Jun-21
Secondary Standards	ID #	Check Date (in house)	Scheduled Check
Power meter S44135	SN: G841203074	06-Apr-15 (in house check Jun-20)	In house check Jun-22
Power sensor E4113A	SN: MY41490807	06-Apr-15 (in house check Jun-20)	In house check Jun-22
Power sensor E4113A	SN: 00010010	06-Apr-15 (in house check Jun-20)	In house check Jun-22
RF generator HP 8440D	SN: US484001709	04-Apr-15 (in house check Jun-20)	In house check Jun-22
Network Analyser Agilent E8363A	SN: US41000477	31-Mar-14 (in house check Oct-20)	In house check Oct-21
Calibrated by:	Name: Jeffrey Katsman	Function: Laboratory Technician	Signature:
Approved by:	Name: Kaja Polovic	Function: Technical Manager	Signature:
This calibration certificate shall not be reproduced except in full without written approval of the laboratory.			
Issued: April 26, 2021			
Certificate No: CLA150-4025_Apr21		Page 1 of 6	

Measurement Conditions	
DASY system configuration, as far as not given on page 1.	
DASY Version	V52.10.4
Extrapolation	Advanced Extrapolation
Phantom	ELIA Flat Phantom
EUT Positioning	Touch Position
Zoom Scan Resolution	$d_x, d_y = 4.0 \text{ mm}$, $d_z = 1.4 \text{ mm}$
Frequency	$150 \text{ MHz} \pm 1 \text{ MHz}$
Graded Ratio	$1.4 \text{ (Z direction)}$

Head TSL parameters	
The following parameters and calculations were applied.	
Nominal Head TSL parameters	22.0°C
Measured Head TSL parameters	$(22.0 \pm 0.2)^\circ\text{C}$
Head TSL temperature change during test	$< 0.5^\circ\text{C}$
Permittivity	52.3
Conductivity	0.75 mho/m
Permittivity	$51.1 \pm 6\%$
Conductivity	$0.75 \text{ mho/m} \pm 6\%$

SAR result with Head TSL	
SAR averaged over 1 cm^3 (1 g) of Head TSL	Condition
SAR measured	1 W input power
SAR for nominal Head TSL parameters	normalized to 1W
SAR measured	3.90 W/kg
SAR for nominal Head TSL parameters	3.88 W/kg $\pm 19.4\%$ (k=2)
SAR averaged over 10 cm^3 (10 g) of Head TSL	condition
SAR measured	1 W input power
SAR for nominal Head TSL parameters	normalized to 1W
SAR measured	2.60 W/kg
SAR for nominal Head TSL parameters	2.59 W/kg $\pm 18.0\%$ (k=2)

Appendix (Additional assessments outside the scope of SCS 0108)	
Antenna Parameters with Head TSL	
Impedance, transformed to feed point	$47.9 \Omega \pm 1.5 \Omega$
Return Loss	-31.4 dB
Additional EUT Data	
Manufactured by	SPEAG

Certificate No: CLA150-4025_Apr21	
Page 2 of 6	

Certificate No: CLA150-4025_Apr21	
Page 3 of 6	


Certificate No: CLA150-4025_Apr21	
Page 4 of 6	



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/Terms-and-Conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10 Weiye Road, Development Zone, Kunshan, Jiangsu, China 215300 t(86-512) 5735 5888 f(86-512) 5737 0818 www.sgs.com.cn
中国·江苏·昆山开发区伟业路10号 邮编: 215300 t(86-512) 5735 5888 f(86-512) 5737 0818 sgs.china@sgs.com



1.2 D450V3 - SN 1103

Calibration Laboratory of Schmid & Partner Engineering AG
Zweigschulstrasse 43, 8004 Zurich, Switzerland

Accredited by the Swiss Accreditation Service (SAS)
The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Client: **SGS-CN (Aude)** Certificate No: **D450V3-1103_Apr21**

CALIBRATION CERTIFICATE

Object: **D450V3 - SN: 1103**

Calibration procedure(s): **QA CAL-15_v9**
Calibration Procedure for SAR Validation Sources below 700 MHz

Calibration date: **April 21, 2021**

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature (20 ± 2) °C and humidity < 70%.

Calibration Equipment used (MPE: critical for calibration)

Primary Standards	ID #	Cal Date (Certificate No.)	Scheduled Calibration
Power meter NRP	SN: 104778	09-Apr-21 (No. 217-03201/03202)	Apr-22
Power sensor NRP-291	SN: 103244	09-Apr-21 (No. 217-03201)	Apr-22
Power sensor NRP-291	SN: 103245	09-Apr-21 (No. 217-03202)	Apr-22
Reference 20 dB Attenuator	SN: CC2852 (200)	09-Apr-21 (No. 217-03345)	Apr-22
Type-N mismatch combination	SN: 310852 / 06327	09-Apr-21 (No. 217-03344)	Apr-22
Reference Probe E30309	SN: 3877	30-Dec-20 (No. E30-3877 Dec20)	Dec-21
DAEA	SN: 654	26-Jun-20 (No. D454-654_Jun20)	Jun-21

Secondary Standards	ID #	Check Date (In House)	Scheduled Check
Power meter E44198	SN: GB41200274	06-Apr-16 (In house check Jun-20)	In house check Jun-22
Power sensor E4412A	SN: MY41496047	06-Apr-16 (In house check Jun-20)	In house check Jun-22
Power sensor E4412A	SN: 00018010	06-Apr-16 (In house check Jun-20)	In house check Jun-22
RF generator HP 8548C	SN: L053460101700	06-Aug-99 (In house check Jun-20)	In house check Jun-22
Network Analyzer Agilent E8358A	SN: U841080477	31-Mar-14 (In house check Oct-20)	In house check Oct-21

Calibrated by: **Christoph Leubner** Function: **Laboratory Technician**

Approved by: **Krista Polovic** Technical Manager

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: D450V3-1103_Apr21 Page 1 of 6

Calibration Laboratory of Schmid & Partner Engineering AG
Zweigschulstrasse 43, 8004 Zurich, Switzerland

Accredited by the Swiss Accreditation Service (SAS)
The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Accreditation No: **SCS 0106**

Glossary:

TSL: Issue simulating liquid sensitivity in TSL / NORM x,y,z

ConvF: not applicable or not measured

N/A: not applicable or not measured

Calibration is Performed According to the Following Standards:

- IEC 61010-1:2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
- IEC 62209-1, "Measurement procedure for the assessment of Specific Absorption Rate (SAR) from hand-held and body-mounted devices used next to the ear (frequency range of 300 MHz to 6 GHz)", July 2016
- IEC 62209-2, "Procedure to determine the Specific Absorption Rate (SAR) for wireless communication devices used in close proximity to the human body (frequency range of 30 MHz to 6 GHz)", March 2010
- KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

Additional Documentation:

- DASY4/5 System Handbook

Methods Applied and Interpretation of Parameters:

- Measurement Conditions:** Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL:** The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss:** These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay:** One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured:** SAR measured at the stated antenna input power.
- SAR normalized:** SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters:** The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor $k=2$, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Certificate No: D450V3-1103_Apr21 Page 2 of 6



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/Terms-and-Conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10 Weiye Road, Development Zone, Kunshan, Jiangsu, China 215300 t(86-512) 5735 5888 f(86-512) 57370818 www.sgsgroup.com.cn
中国·江苏·昆山开发区伟业路10号 邮编: 215300 t(86-512) 57355888 f(86-512) 57370818 sgs.china@sgs.com

Measurement Conditions		
DASY system configuration, as far as not given on page 1.		
DASY Version	DASY5	V82.10.4
Extrapolation	Advanced Extrapolation	
Phantom	ELJ Flat Phantom	Shell thickness: 2 ± 0.2 mm
Distance Dipole Center - TSL	15 mm	with Spacer
Zoom Scan Resolution	dx, dy, dz = 5 mm	
Frequency	450 MHz \pm 1 MHz	

Head TSL parameters

The following parameters and calculations were applied:

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	43.5	0.57 mho/m
Measured Head TSL parameters	(22.0 \pm 0.2) °C	43.1 \pm 6 %	0.07 mho/m \pm 8 %
Head TSL temperature change during test	< 0.5 °C	---	---

SAR result with Head TSL

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	250 mW input power	1.14 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	4.55 W/kg \pm 18.1 % (k=2)

SAR averaged over 10 cm ³ (10 g) of Head TSL	condition	
SAR measured	250 mW input power	0.757 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	3.06 W/kg \pm 17.6 % (k=2)

Certificate No: D450V3-1103_Apr21

Page 5 of 6

DASY5 Validation Report for Head TSL

Test Laboratory: SPEAG, Zurich, Switzerland

Date: 21.04.2021

DUT: Dipole 450 MHz; Type: D450V3; Serial: D450V3 - SN:1103

Communication System: UTD 0 - CW; Frequency: 450 MHz

Medium parameters used: $f = 450$ MHz; $\alpha = 0.87$ S/m; $\epsilon_r = 43.1$; $\rho = 1000$ kg/m³

Phantom section: Flat Section

Measurement Standard: DASY5 (IEE/IEC/ANSI C63.19-2011)

DASY52 Configuration:

- Probe: EX3DV4 - SN3877; ConvF(10.64, 10.64, 10.64) @ 450 MHz; Calibrated: 30.12.2020
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn654; Calibrated: 26.06.2020
- Phantom: ELJ v4.0; Type: QDOVA001BB; Serial: TP:1003
- DASY52 52.10.4(1527); SEMCAD X 14.6.14(7483)

Dipole Calibration for Head Tissue(d=15mm, Pin=250mW/Zoom Scan (7x7x7)/Cube 0:

Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 39.18 V/m; Power Drift = -0.08 dB

Peak SAR (extrapolated) = 1.76 W/kg

SAR(1 g) = 1.14 W/kg; SAR(10 g) = 0.767 W/kg

Smallest distance from peaks to all points 3 dB below: Larger than measurement grid

Ratio of SAR at M2 to SAR at M1 = 64.9%

Maximum value of SAR (measured) = 1.53 W/kg

Certificate No: D450V3-1103_Apr21

Appendix (Additional assessments outside the scope of SCS 0108)

Antenna Parameters with Head TSL

Impedance, transformed to feed point	57.1 Ω - j2.8 jΩ
Return Loss	-23.0 dB

General Antenna Parameters and Design

Electrical Delay (one direction)	1.346 ns
----------------------------------	----------

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard overlight coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.

No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

Additional EUT Data

Manufactured by	SPEAG
-----------------	-------

Certificate No: D450V3-1103_Apr21

Page 4 of 6

Certificate No: D450V3-1103_Apr21

Page 4 of 6

1.3 D750V3 - SN 1188


TTL Speag CALIBRATION LABORATORY		CAICT																					
Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2512 Fax: +86-10-42304633-2504 E-mail: cti@chinaetl.com http://www.chinaetl.cn		Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@chinaetl.com http://www.chinaetl.cn																					
Client: SGS-CN		Certificate No: Z22-60103																					
CALIBRATION CERTIFICATE																							
Object: D750V3 - SN: 1188																							
Calibration Procedure(s): FF-Z11-003-01 Calibration Procedures for dipole validation kits																							
Calibration date: March 28, 2022																							
This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.																							
All calibrations have been conducted in the closed laboratory facility: environment temperature (22±3)°C and humidity<70%.																							
Calibration Equipment used (M&TE critical for calibration)																							
<table border="1"><thead><tr><th>Primary Standards</th><th>ID #</th><th>Cal Date (Calibrated by Certificate No.)</th><th>Scheduled Calibration</th></tr></thead><tbody><tr><td>Power Meter NRP2</td><td>104277</td><td>24-Sep-21 (CTTL No.J21X08326)</td><td>Sep-22</td></tr><tr><td>Power sensor NRP88</td><td>104291</td><td>24-Sep-21 (CTTL No.J21X08326)</td><td>Sep-22</td></tr><tr><td>Reference Probe EX30V4</td><td>SN 7307</td><td>26-May-21(SPEAG.No.EX3-7307_May21)</td><td>May-22</td></tr><tr><td>D4E4</td><td>SN 1556</td><td>12-Jan-22(CTTL-SPEAG.No.Z22-60007)</td><td>Jan-23</td></tr></tbody></table>				Primary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration	Power Meter NRP2	104277	24-Sep-21 (CTTL No.J21X08326)	Sep-22	Power sensor NRP88	104291	24-Sep-21 (CTTL No.J21X08326)	Sep-22	Reference Probe EX30V4	SN 7307	26-May-21(SPEAG.No.EX3-7307_May21)	May-22	D4E4	SN 1556	12-Jan-22(CTTL-SPEAG.No.Z22-60007)	Jan-23
Primary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration																				
Power Meter NRP2	104277	24-Sep-21 (CTTL No.J21X08326)	Sep-22																				
Power sensor NRP88	104291	24-Sep-21 (CTTL No.J21X08326)	Sep-22																				
Reference Probe EX30V4	SN 7307	26-May-21(SPEAG.No.EX3-7307_May21)	May-22																				
D4E4	SN 1556	12-Jan-22(CTTL-SPEAG.No.Z22-60007)	Jan-23																				
<table border="1"><thead><tr><th>Secondary Standards</th><th>ID #</th><th>Cal Date (Calibrated by Certificate No.)</th><th>Scheduled Calibration</th></tr></thead><tbody><tr><td>Signal Generator S4438C</td><td>MY48071430</td><td>13-Jan-22 (CTTL No.J22X00409)</td><td>Jan-23</td></tr><tr><td>Network Analyzer E5071C</td><td>MY46110673</td><td>14-Jan-22 (CTTL No.J22X00409)</td><td>Jan-23</td></tr></tbody></table>				Secondary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration	Signal Generator S4438C	MY48071430	13-Jan-22 (CTTL No.J22X00409)	Jan-23	Network Analyzer E5071C	MY46110673	14-Jan-22 (CTTL No.J22X00409)	Jan-23								
Secondary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration																				
Signal Generator S4438C	MY48071430	13-Jan-22 (CTTL No.J22X00409)	Jan-23																				
Network Analyzer E5071C	MY46110673	14-Jan-22 (CTTL No.J22X00409)	Jan-23																				
Calibrated by: Zhao Jing SAR Test Engineer		Signature																					
Reviewed by: Lin Hao SAR Test Engineer		Signature																					
Approved by: Qi Dianyan SAR Project Leader		Signature																					
Issued: April 3, 2022																							
This calibration certificate shall not be reproduced except in full without written approval of the laboratory.																							
Certificate No: Z22-60103		Page 1 of 6																					

TTL Speag CALIBRATION LABORATORY		CAICT																			
Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@chinaetl.com http://www.chinaetl.cn		Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@chinaetl.com http://www.chinaetl.cn																			
Measurement Conditions DASY system configuration, as far as not given on page 1:																					
<table border="1"><thead><tr><th>DASY Version</th><th>DASY52</th><th>V52.10.4</th></tr></thead><tbody><tr><td>Extrapolation</td><td>Advanced Extrapolation</td><td></td></tr><tr><td>Phantom</td><td>Triple Flat Phantom 5.1C</td><td></td></tr><tr><td>Distance Dipole Center - TSL</td><td>15 mm</td><td>with Spacer</td></tr><tr><td>Zoom Scan Resolution</td><td>dx, dy, dz = 5 mm</td><td></td></tr><tr><td>Frequency</td><td>750 MHz ± 1 MHz</td><td></td></tr></tbody></table>				DASY Version	DASY52	V52.10.4	Extrapolation	Advanced Extrapolation		Phantom	Triple Flat Phantom 5.1C		Distance Dipole Center - TSL	15 mm	with Spacer	Zoom Scan Resolution	dx, dy, dz = 5 mm		Frequency	750 MHz ± 1 MHz	
DASY Version	DASY52	V52.10.4																			
Extrapolation	Advanced Extrapolation																				
Phantom	Triple Flat Phantom 5.1C																				
Distance Dipole Center - TSL	15 mm	with Spacer																			
Zoom Scan Resolution	dx, dy, dz = 5 mm																				
Frequency	750 MHz ± 1 MHz																				
Head TSL parameters The following parameters and calculations were applied:																					
<table border="1"><thead><tr><th></th><th>Temperature</th><th>Permittivity</th><th>Conductivity</th></tr></thead><tbody><tr><td>Nominal Head TSL parameters</td><td>22.0 °C</td><td>42.0</td><td>0.90 mho/m</td></tr><tr><td>Measured Head TSL parameters</td><td>(22.0 ± 0.2) °C</td><td>41.4 ± 6 %</td><td>0.89 mho/m ± 6 %</td></tr><tr><td>Head TSL temperature change during test</td><td><1.0 °C</td><td>---</td><td>---</td></tr></tbody></table>					Temperature	Permittivity	Conductivity	Nominal Head TSL parameters	22.0 °C	42.0	0.90 mho/m	Measured Head TSL parameters	(22.0 ± 0.2) °C	41.4 ± 6 %	0.89 mho/m ± 6 %	Head TSL temperature change during test	<1.0 °C	---	---		
	Temperature	Permittivity	Conductivity																		
Nominal Head TSL parameters	22.0 °C	42.0	0.90 mho/m																		
Measured Head TSL parameters	(22.0 ± 0.2) °C	41.4 ± 6 %	0.89 mho/m ± 6 %																		
Head TSL temperature change during test	<1.0 °C	---	---																		
SAR result with Head TSL																					
<table border="1"><thead><tr><th>SAR averaged over 1 cm² (1 g) of Head TSL</th><th>Condition</th><th></th></tr></thead><tbody><tr><td>SAR measured</td><td>250 mW input power</td><td>2.07 W/kg</td></tr><tr><td>SAR for nominal Head TSL parameters</td><td>normalized to 1W</td><td>8.27 W/kg ± 18.8 % (k=2)</td></tr><tr><td>SAR averaged over 10 cm² (10 g) of Head TSL</td><th>Condition</th><th></th></tr><tr><td>SAR measured</td><td>250 mW input power</td><td>1.37 W/kg</td></tr><tr><td>SAR for nominal Head TSL parameters</td><td>normalized to 1W</td><td>5.48 W/kg ± 18.7 % (k=2)</td></tr></tbody></table>				SAR averaged over 1 cm² (1 g) of Head TSL	Condition		SAR measured	250 mW input power	2.07 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	8.27 W/kg ± 18.8 % (k=2)	SAR averaged over 10 cm² (10 g) of Head TSL	Condition		SAR measured	250 mW input power	1.37 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	5.48 W/kg ± 18.7 % (k=2)
SAR averaged over 1 cm² (1 g) of Head TSL	Condition																				
SAR measured	250 mW input power	2.07 W/kg																			
SAR for nominal Head TSL parameters	normalized to 1W	8.27 W/kg ± 18.8 % (k=2)																			
SAR averaged over 10 cm² (10 g) of Head TSL	Condition																				
SAR measured	250 mW input power	1.37 W/kg																			
SAR for nominal Head TSL parameters	normalized to 1W	5.48 W/kg ± 18.7 % (k=2)																			

TTL Speag CALIBRATION LABORATORY		CAICT	
Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@chinaetl.com http://www.chinaetl.cn		Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@chinaetl.com http://www.chinaetl.cn	
Glossary:			
TSL	tissue simulating liquid		
ConvF	sensitivity in TSL / NORMx.y.z		
N/A	not applicable or not measured		
Calibration is Performed According to the Following Standards:			
a) IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices-Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020			
b) KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"			
Additional Documentation:			
c) DASY4/5 System Handbook			
Methods Applied and Interpretation of Parameters:			
• Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.			
• Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.			
• Feed Point Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.			
• Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required.			
• SAR measured: SAR measured at the stated antenna input power.			
• SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector.			
• MY for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.			
The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.			
Certificate No: Z22-60103		Page 2 of 6	

TTL Speag CALIBRATION LABORATORY		CAICT					
Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@chinaetl.com http://www.chinaetl.cn		Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@chinaetl.com http://www.chinaetl.cn					
Appendix (Additional assessments outside the scope of CNAS L0570)							
Antenna Parameters with Head TSL							
<table border="1"><thead><tr><th>Impedance, transformed to feed point</th><th>53.60-1.13jΩ</th></tr><tr><th>Return Loss</th><th>-28.7dB</th></tr></thead></table>				Impedance, transformed to feed point	53.60-1.13jΩ	Return Loss	-28.7dB
Impedance, transformed to feed point	53.60-1.13jΩ						
Return Loss	-28.7dB						
General Antenna Parameters and Design							
<table border="1"><thead><tr><th>Electrical Delay (one direction)</th><th>0.947 ns</th></tr></thead></table>				Electrical Delay (one direction)	0.947 ns		
Electrical Delay (one direction)	0.947 ns						
After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.							
The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.							
No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.							
Additional EUT Data							
<table border="1"><thead><tr><th>Manufactured by</th><th>SPEAG</th></tr></thead></table>				Manufactured by	SPEAG		
Manufactured by	SPEAG						
Certificate No: Z22-60103		Page 4 of 6					





1.4 D835V2 - SN 4d114

TTS Calibration Laboratory		CAICT	
Add: No.52 HuaYuanbei Road, Haidian District, Beijing, 100191, China Tel: +86-10-6204631-2079 Fax: +86-10-6204631-2504 E-mail: cti@china.tts.com.cn http://www.china.tts.com.cn		Add: No.52 HuaYuanbei Road, Haidian District, Beijing, 100191, China Tel: +86-10-6204631-2079 Fax: +86-10-6204631-2504 E-mail: cti@china.tts.com.cn http://www.china.tts.com.cn	
Client: SGS-CN		Certificate No: Z22-60104	
CALIBRATION CERTIFICATE			
Object: D835V2 - SN: 4d114			
Calibration Procedure(s): FF-Z11-003-01 Calibration Procedures for dipole validation kits			
Calibration date: March 31, 2022			
This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.			
All calibrations have been conducted in the closed laboratory facility: environment temperature (22±3)°C and humidity <70%.			
Calibration Equipment used (M&TE critical for calibration)			
Primary Standards		ID # Cal Date (Calibrated by, Certificate No.) Scheduled Calibration	
Power Meter NRP2		106277 24-Sep-21 (CTTL No.J21X08326) Sep-22	
Power sensor NRPBS		104261 24-Sep-21 (CTTL No.J21X08326) Sep-22	
Reference Probe EX3DV4		SN 7307 26-May-21 (SPEAG No.EX3-7307_May21) May-22	
DAE4		SN 1556 12-Jan-22 (CTTL-SPEAG No.Z22-60007) Jan-23	
Secondary Standards		ID # Cal Date (Calibrated by, Certificate No.) Scheduled Calibration	
Signal Generator E4438C		MY49071430 13-Jan-22 (CTTL No.J22X00409) Jan-23	
Network Analyzer E5071C		MY46110673 14-Jan-22 (CTTL No.J22X00409) Jan-23	
Calibrated by:		Name Function Signature	
Reviewed by:		Zhao Jing SAR Test Engineer	
Approved by:		Lin Hao SAR Test Engineer	
		Qi Diyan SAR Project Leader	
Issued: April 6, 2022			
This calibration certificate shall not be reproduced except in full without written approval of the laboratory.			



Glossary:
TSL: tissue simulating liquid
ConvF: sensitivity in TSL / NORMx.y.z
N/A: not applicable or not measured

Calibration is Performed According to the Following Standards:
a) IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices- Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020
b) KDB 665864, "SAR Measurement Requirements for 100 MHz to 6 GHz"

Additional Documentation:
c) DASY4/5 System Handbook

Methods Applied and Interpretation of Parameters:
• **Measurement Conditions:** Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
• **Antenna Parameters with TSL:** The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
• **Feed Point Impedance and Return Loss:** These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
• **Electrical Delay:** One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
• **SAR measured:** SAR measured at the stated antenna input power.
• **SAR normalized:** SAR as measured, normalized to an input power of 1 W at the antenna connector.
• **SAR for nominal TSL parameters:** The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution Corresponds to a coverage probability of approximately 95%.

TTL Calibration Laboratory		CAICT																			
<p>Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com http://www.china.ttl.com</p>																					
<p>Measurement Conditions DASY system configuration, as far as not given on page 1.</p> <table border="1"> <tr> <td>DASY Version</td> <td>DASY52</td> <td>VS2 10.4</td> </tr> <tr> <td>Extrapolation</td> <td>Advanced Extrapolation</td> <td></td> </tr> <tr> <td>Phantom</td> <td>Triple Flat Phantom 5.1C</td> <td></td> </tr> <tr> <td>Distance Dipole Center - TSL</td> <td>15 mm</td> <td>with Spacer</td> </tr> <tr> <td>Zoom Scan Resolution</td> <td>dx, dy, dz = 5 mm</td> <td></td> </tr> <tr> <td>Frequency</td> <td>835 MHz ± 1 MHz</td> <td></td> </tr> </table>				DASY Version	DASY52	VS2 10.4	Extrapolation	Advanced Extrapolation		Phantom	Triple Flat Phantom 5.1C		Distance Dipole Center - TSL	15 mm	with Spacer	Zoom Scan Resolution	dx, dy, dz = 5 mm		Frequency	835 MHz ± 1 MHz	
DASY Version	DASY52	VS2 10.4																			
Extrapolation	Advanced Extrapolation																				
Phantom	Triple Flat Phantom 5.1C																				
Distance Dipole Center - TSL	15 mm	with Spacer																			
Zoom Scan Resolution	dx, dy, dz = 5 mm																				
Frequency	835 MHz ± 1 MHz																				
<p>Head TSL parameters The following parameters and calculations were applied.</p> <table border="1"> <tr> <td></td> <td>Temperature</td> <td>Permittivity</td> <td>Conductivity</td> </tr> <tr> <td>Nominal Head TSL parameters</td> <td>22.0 °C</td> <td>41.5</td> <td>0.90 mho/m</td> </tr> <tr> <td>Measured Head TSL parameters</td> <td>(22.0 ± 0.2) °C</td> <td>41.0 ± 5 %</td> <td>0.91 mho/m ± 6 %</td> </tr> <tr> <td>Head TSL temperature change during test</td> <td><1.0 °C</td> <td>---</td> <td>---</td> </tr> </table>					Temperature	Permittivity	Conductivity	Nominal Head TSL parameters	22.0 °C	41.5	0.90 mho/m	Measured Head TSL parameters	(22.0 ± 0.2) °C	41.0 ± 5 %	0.91 mho/m ± 6 %	Head TSL temperature change during test	<1.0 °C	---	---		
	Temperature	Permittivity	Conductivity																		
Nominal Head TSL parameters	22.0 °C	41.5	0.90 mho/m																		
Measured Head TSL parameters	(22.0 ± 0.2) °C	41.0 ± 5 %	0.91 mho/m ± 6 %																		
Head TSL temperature change during test	<1.0 °C	---	---																		
<p>SAR result with Head TSL</p> <table border="1"> <tr> <td>SAR averaged over 1 cm³ (1 g) of Head TSL</td> <td>Condition</td> <td></td> </tr> <tr> <td>SAR measured</td> <td>250 mW input power</td> <td>2.37 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>9.40 W/kg ± 18.8 % (k=2)</td> </tr> <tr> <td>SAR averaged over 10 cm³ (10 g) of Head TSL</td> <td>Condition</td> <td></td> </tr> <tr> <td>SAR measured</td> <td>250 mW input power</td> <td>1.54 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>6.12 W/kg ± 18.7 % (k=2)</td> </tr> </table>				SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition		SAR measured	250 mW input power	2.37 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	9.40 W/kg ± 18.8 % (k=2)	SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition		SAR measured	250 mW input power	1.54 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	6.12 W/kg ± 18.7 % (k=2)
SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition																				
SAR measured	250 mW input power	2.37 W/kg																			
SAR for nominal Head TSL parameters	normalized to 1W	9.40 W/kg ± 18.8 % (k=2)																			
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition																				
SAR measured	250 mW input power	1.54 W/kg																			
SAR for nominal Head TSL parameters	normalized to 1W	6.12 W/kg ± 18.7 % (k=2)																			
<p>Certificate No: Z22-60104 Page 3 of 6</p>																					
<p>Appendix (Additional assessments outside the scope of CNAS L0570)</p> <p>Antenna Parameters with Head TSL</p> <table border="1"> <tr> <td>Impedance, transformed to feed point</td> <td>48.70 - j22.0Q</td> </tr> <tr> <td>Return Loss</td> <td>-25.3dB</td> </tr> </table> <p>General Antenna Parameters and Design</p> <table border="1"> <tr> <td>Electrical Delay (one direction)</td> <td>1.307 ns</td> </tr> </table> <p>After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.</p> <p>The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.</p> <p>No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.</p> <p>Additional EUT Data</p> <table border="1"> <tr> <td>Manufactured by</td> <td>SPEAG</td> </tr> </table>				Impedance, transformed to feed point	48.70 - j22.0Q	Return Loss	-25.3dB	Electrical Delay (one direction)	1.307 ns	Manufactured by	SPEAG										
Impedance, transformed to feed point	48.70 - j22.0Q																				
Return Loss	-25.3dB																				
Electrical Delay (one direction)	1.307 ns																				
Manufactured by	SPEAG																				
<p>Certificate No: Z22-60104 Page 4 of 6</p>																					
<p>DASY Validation Report for Head TSL Test Laboratory: CTTL, Beijing, China Date: 2022-03-31</p> <p>DUT: Dipole 835 MHz; Type: D835V2; Serial: D835V2 - SN: 4d114 Communication System: UID 0, CW; Frequency: 835 MHz; Duty Cycle: 1:1 Medium parameters used: f = 835 MHz; σ = 0.907 S/m; ε_r = 40.98; ρ = 1000 kg/m³ Phantom section: Right Section Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007) DASY5 Configuration:</p> <ul style="list-style-type: none"> Probe: EX3DV4 - SN7307; ConvF(10.13, 10.13, 10.13) @ 835 MHz; Calibrated: 2021-05-26 Sensor Surface: 1.4mm (Mechanical Surface Detection) Electronics: DAE4 Sn1556; Calibrated: 2022-01-12 Phantom: MFP_V5.1C (20deg probe tilt); Type: QD 000 P51 Cx; Serial: 1062 DASY52 52.10.4(1535); SEMCAD X 14.6.14(7501) <p>Dipole Calibration/Zoom Scan (7x7x7) (7x7x7) Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm Reference Value = 57.88 V/m; Power Drift = 0.04 dB Peak SAR (extrapolated) = 3.56 W/kg SAR(1 g) = 2.37 W/kg; SAR(10 g) = 1.54 W/kg Smallest distance from peaks to all points 3 dB below = 15.8 mm Ratio of SAR at M2 to SAR at M1 = 66.2% Maximum value of SAR (measured) = 3.17 W/kg</p> 																					
<p>Certificate No: Z22-60104 Page 5 of 6</p>																					
<p>Impedance Measurement Plot for Head TSL</p> 																					
<p>Certificate No: Z22-60104 Page 6 of 6</p>																					

1.5 D900V2 - SN 1d079

In Collaboration with TTL CALIBRATION LABORATORY		In Collaboration with CAICT CALIBRATION LABORATORY	
Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191 Tel: +86-10-4204633-2117 E-mail: cti@china.com.cn		Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-4204633-2117 E-mail: cti@china.com.cn	
Client: SGS-CN		Certificate No: Z22-60184	
CALIBRATION CERTIFICATE			
Object: D900V2 - SN: 1d079			
Calibration Procedure(s): FF-Z11-003-01			
Calibration date: June 7, 2022			
This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.			
All calibrations have been conducted in the closed laboratory facility: environment temperature (23±1)°C and humidity <70%.			
Calibration Equipment used (MATE critical for calibration)			
Primary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration
Power Meter NRP2	106277	24-Sep-21 (CTTL No.J21X08326)	Sep-22
Power sensor NRP8S	104291	24-Sep-21 (CTTL No.J21X08326)	Sep-22
Reference Probe EX3DV4	SN 7464	28-Jan-22 (SPEAG No EX3-7464_Jan22)	Jan-23
DAE4	SN 1556	12-Jan-22 (CTTL-SPEAG No Z22-60007)	Jan-23
Secondary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration
Signal Generator E4438C	MY48071450	13-Jan-22 (CTTL No.Z22X00409)	Jan-23
Network Analyzer E5071C	MY48110673	14-Jan-22 (CTTL No.Z22X00409)	Jan-23
Calibrated by: Zhao Jing SAR Test Engineer			
Reviewed by: Lin Hao SAR Test Engineer			
Approved by: Qi Dianyuan SAR Project Leader			
Issued: June 13, 2022			
This calibration certificate shall not be reproduced except in full without written approval of the laboratory.			
Certificate No: Z22-60184		Page 1 of 6	

In Collaboration with TTL CALIBRATION LABORATORY		In Collaboration with CAICT CALIBRATION LABORATORY	
Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-4204633-2117 E-mail: cti@china.com.cn		Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-4204633-2117 E-mail: cti@china.com.cn	
Glossary: TSL: tissue simulating liquid ConvF: sensitivity in TSL / NORMx,y,z N/A: not applicable or not measured			
Calibration is Performed According to the Following Standards: a) IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices- Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020 b) KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz" Additional Documentation: c) DASYS4/5 System Handbook			
Methods Applied and Interpretation of Parameters: • Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated. • Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis. • Feed Point Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required. • Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required. • SAR measured: SAR measured at the stated antenna input power. • SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector. • SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.			
The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.			
Certificate No: Z22-60184		Page 2 of 6	

In Collaboration with TTL CALIBRATION LABORATORY		In Collaboration with CAICT CALIBRATION LABORATORY	
Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-4204633-2117 E-mail: cti@china.com.cn		Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-4204633-2117 E-mail: cti@china.com.cn	
Measurement Conditions DASY system configuration, as far as not given on page 1.			
DASY Version	DASY52	52.10.4	
Extrapolation	Advanced Extrapolation		
Phantom	Triple Flat Phantom 5.1C		
Distance Dipole Center - TSL	15 mm	with Spacer	
Zoom Scan Resolution	dx, dy, dz = 5 mm		
Frequency	900 MHz ± 1 MHz		
Head TSL parameters The following parameters and calculations were applied.			
	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	41.5	0.97 nholm
Measured Head TSL parameters	(22.0 ± 0.2) °C	42.1 ± 6 %	0.98 nholm ± 6 %
Head TSL temperature change during test	+1.0 °C	---	---
SAR result with Head TSL			
SAR averaged over 1 cm ² (1 g) of Head TSL	Condition		
SAR measured	250 mW input power	2.70 W/kg	
SAR for nominal Head TSL parameters	normalized to 1W	11.0 W/kg ± 18.8 % (k=2)	
SAR averaged over 10 cm ² (10 g) of Head TSL	Condition		
SAR measured	250 mW input power	1.79 W/kg	
SAR for nominal Head TSL parameters	normalized to 1W	7.09 W/kg ± 18.7 % (k=2)	
Certificate No: Z22-60184		Page 3 of 6	


In Collaboration with TTL CALIBRATION LABORATORY		In Collaboration with CAICT CALIBRATION LABORATORY	
Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-4204633-2117 E-mail: cti@china.com.cn		Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-4204633-2117 E-mail: cti@china.com.cn	
Appendix (Additional assessments outside the scope of CNAS L0570)			
Antenna Parameters with Head TSL			
Impedance, transformed to feed point	48.10 - j8.49Ω		
Return Loss	-23.3 dB		
General Antenna Parameters and Design			
Electrical Delay (one direction)	1.312 ns		
After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.			
The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.			
No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.			
Additional EUT Data			
Manufactured by	SPEAG		
Certificate No: Z22-60184		Page 4 of 6	




Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/terms-and-conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10 Weiye Road, Development Zone, Kunshan, Jiangsu, China 215300 t(86-512) 5735 5888 f(86-512) 57370818 www.sgs.com.cn
中国·江苏·昆山开发区伟业路10号 邮编: 215300 t(86-512) 57355888 f(86-512) 57370818 sgs.china@sgs.com






1.7 D1900V2 - SN 5d136

TTL S p e a g		CAICT	
In Collaboration with CALIBRATION LABORATORY		中国合格评定国家认可委员会 CALIBRATION	
Add: No.52 Huayuanbei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42206033-2117 E-mail: cti@china.com.cn		Add: No.52 Huayuanbei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42206033-2117 E-mail: cti@china.com.cn	
Client: SGS-CN		Certificate No: Z22-60185	
CALIBRATION CERTIFICATE			
Object	D1900V2 - SN: 5d136		
Calibration Procedure(s)	FF-Z11-003-01 Calibration Procedures for dipole validation kits		
Calibration date:	June 7, 2022		
This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.			
All calibrations have been conducted in the closed laboratory facility: environment temperature (23±1)°C and humidity <70%.			
Calibration Equipment used (M&TE critical for calibration)			
Primary Standards	ID #	Cal Date (Calibrated by: Certificate No.)	Scheduled Calibration
Power Meter NRP2	106277	24-Sep-21 (CTTL No.J21X08326)	Sep-22
Power sensor NRP6S	104291	24-Sep-21 (CTTL No.J21X08326)	Sep-22
Reference Probe EXSDV4	SN 7464	28-Jan-22 (SPEAG No EX3-7464_Jan22)	Jan-23
DAE4	SN 1656	12-Jan-22 (CTTL-SPEAG No Z22-60007)	Jan-23
Secondary Standards	ID #	Cal Date (Calibrated by: Certificate No.)	Scheduled Calibration
Signal Generator E4438C	MY48071430	13-Jan-22 (CTTL No.J22X00409)	Jan-23
Network Analyzer E5071C	MY48110073	14-Jan-22 (CTTL No.J22X00409)	Jan-23
Calibrated by:	Name	Function	Signature
	Zhao Jing	SAR Test Engineer	
Reviewed by:	Lin Hao	SAR Test Engineer	
Approved by:	Qi Dianyan	SAR Project Leader	
Issued: June 13, 2022			
This calibration certificate shall not be reproduced except in full without written approval of the laboratory.			
Certificate No: Z22-60185		Page 1 of 6	


TTL S p e a g		CAICT	
In Collaboration with CALIBRATION LABORATORY		中国合格评定国家认可委员会 CALIBRATION	
Add: No.52 Huayuanbei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42206033-2117 E-mail: cti@china.com.cn		Add: No.52 Huayuanbei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42206033-2117 E-mail: cti@china.com.cn	
Glossary:			
TSL	tissue simulating liquid		
ComF	sensitivity in TSL / NORMx.y.z		
N/A	not applicable or not measured		
Calibration is Performed According to the Following Standards:			
a) IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices- Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020			
b) KDB 865984, "SAR Measurement Requirements for 100 MHz to 6 GHz"			
Additional Documentation:			
c) DASY4/5 System Handbook			
Methods Applied and Interpretation of Parameters:			
• Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.			
• Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.			
• Feed Point Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.			
• Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required.			
• SAR measured: SAR measured at the stated antenna input power.			
• SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector.			
• SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.			
The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.			
Certificate No: Z22-60185		Page 2 of 6	


TTL S p e a g		CAICT	
In Collaboration with CALIBRATION LABORATORY		中国合格评定国家认可委员会 CALIBRATION	
Add: No.52 Huayuanbei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42206033-2117 E-mail: cti@china.com.cn		Add: No.52 Huayuanbei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42206033-2117 E-mail: cti@china.com.cn	
Measurement Conditions			
DASY system configuration, as far as not given on page 1.			
DASY Version	DASY52		52.10.4
Extrapolation	Advanced Extrapolation		
Phantom	Triple Flat Phantom 5.1C		
Distance Dipole Center - TSL	10 mm		with Spacer
Zoom Scan Resolution	dx, dy, dz = 5 mm		
Frequency	1900 MHz ± 1 MHz		
Head TSL parameters			
The following parameters and calculations were applied:			
Nominal Head TSL parameters	Temperature	Permittivity	Conductivity
	32.0 °C	40.0	1.40 mS/m
Measured Head TSL parameters	(22.0 ± 0.3) °C	39.9 ± 6 %	1.39 mS/m ± 6 %
Head TSL temperature change during test	<1.0 °C	---	---
SAR result with Head TSL			
SAR averaged over 1 cm ² (1 g) of Head TSL	Condition		
SAR measured	250 mW input power	9.65 W/kg	
SAR for nominal Head TSL parameters	normalized to 1W	40.0 W/kg ± 16.8 % (k=2)	
SAR averaged over 10 cm ² (10 g) of Head TSL	Condition		
SAR measured	250 mW input power	5.18 W/kg	
SAR for nominal Head TSL parameters	normalized to 1W	20.8 W/kg ± 18.7 % (k=2)	
Certificate No: Z22-60185		Page 3 of 6	

TTL S p e a g		CAICT	
In Collaboration with CALIBRATION LABORATORY		中国合格评定国家认可委员会 CALIBRATION	
Add: No.52 Huayuanbei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42206033-2117 E-mail: cti@china.com.cn		Add: No.52 Huayuanbei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42206033-2117 E-mail: cti@china.com.cn	
Appendix (Additional assessments outside the scope of CNAS L0570)			
Antenna Parameters with Head TSL			
Impedance, transformed to feed point	51.02 ± 7.56 Ω		
Return Loss	-22.4 dB		
General Antenna Parameters and Design			
Electrical Delay (one direction)	1.109 ns		
After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.			
The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC signals. On some of the dipoles, small and caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.			
Additional EUT Data			
Manufactured by	SPEAG		
Certificate No: Z22-60185		Page 4 of 6	



1.8 D2000V2 - SN 1041







1.9 D2300V2 - SN 1096

TTL S p e a g CALIBRATION LABORATORY		CAICT	
Add: No.52 HuaYuanbei Road, Haidian District, Beijing, 100191 Tel: +86-10-42304633-2512 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com http://www.china.ttl.com		Add: No.52 HuaYuanbei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com http://www.china.ttl.com	
Client: SGS-CN		Certificate No: Z22-60106	
CALIBRATION CERTIFICATE			
Object	D2300V2 - SN: 1096		
Calibration Procedure(s)	FF-Z11-003-01 Calibration Procedures for dipole validation kits		
Calibration date:	March 31, 2022		
This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.			
All calibrations have been conducted in the closed laboratory facility: environment temperature (22±3)°C and humidity <70%.			
Calibration Equipment used (M&TE critical for calibration)			
Primary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration
Power Meter NRP2	108277	24-Sep-21 (CTTL No.J21X08328)	Sep-22
Power sensor NRP8S	104291	24-Sep-21 (CTTL No.J21X08328)	Sep-22
Reference Probe EX30V4	SN 7307	26-May-21 (SPEAG No.EK3-7307_May21)	May-22
DAE4	SN 1556	12-Jan-22 (CTTL-SPEAG No.Z22-60007)	Jan-23
Secondary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration
Signal Generator E4438C	MY49071430	13-Jan-22 (CTTL No.J22X00406)	Jan-23
Network Analyzer E5071C	MY48110673	14-Jan-22 (CTTL No.J22X00406)	Jan-23
Calibrated by:	Name	Function	Signature
	Zhao Jing	SAR Test Engineer	
Reviewed by:	Lin Hao	SAR Test Engineer	
Approved by:	Qi Dianyan	SAR Project Leader	
Issued: April 6, 2022			
This calibration certificate shall not be reproduced except in full without written approval of the laboratory.			
Certificate No: Z22-60106		Page 1 of 6	








TTL S p e a g CALIBRATION LABORATORY		CAICT	
Add: No.52 HuaYuanbei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com http://www.china.ttl.com		Add: No.52 HuaYuanbei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com http://www.china.ttl.com	
Measurement Conditions DASY system configuration, as far as not given on page 1:			
DASY Version	DASY2	52.10.4	
Extrapolation	Advanced Extrapolation		
Phantom	Triple Flat Phantom 5.1C		
Distance Dipole Center - TSL	10 mm	with Spacer	
Zoom Scan Resolution	dx, dy, dz = 5 mm		
Frequency	2300 MHz ± 1 MHz		
Head TSL parameters The following parameters and calculations were applied:			
	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	39.5	1.67 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	39.8 ± 6 %	1.70 mho/m ± 6 %
Head TSL temperature change during test	<1.0 °C	—	—
SAR result with Head TSL			
SAR averaged over 1 cm ² (1 g) of Head TSL	Condition		
SAR measured	250 mW input power	12.4 W/kg	
SAR for nominal Head TSL parameters	normalized to 1W	49.2 W/kg ± 18.8 % (k=2)	
SAR averaged over 10 cm ² (10 g) of Head TSL	Condition		
SAR measured	250 mW input power	5.88 W/kg	
SAR for nominal Head TSL parameters	normalized to 1W	23.4 W/kg ± 18.7 % (k=2)	
Certificate No: Z22-60106		Page 3 of 6	



TTL S p e a g CALIBRATION LABORATORY		CAICT	
Add: No.52 HuaYuanbei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com http://www.china.ttl.com		Add: No.52 HuaYuanbei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com http://www.china.ttl.com	
Appendix (Additional assessments outside the scope of CNAS L0570)			
Antenna Parameters with Head TSL			
Impedance, transformed to feed point	49.20 - 4.56jΩ		
Return Loss	-26.6dB		
General Antenna Parameters and Design			
Electrical Delay (one direction)	1.083 ns		
After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.			
The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.			
No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.			
Additional EUT Data			
Manufactured by	SPEAG		
Certificate No: Z22-60106		Page 4 of 6	





1.10 D2450V2 - SN 817

   													
In Collaboration with Add: No.32 Huailin Road, Huailin District, Beijing, 100191 Tel: +86-10-62462433-2312 Fax: +86-10-62462433-2304 E-mail: cnas@china.com.cn http://www.china.com.cn													
Client	SGS-CN												
Certificate No: Z22-60107													
CALIBRATION CERTIFICATE													
Object	D2450V2-SN: 817												
Calibration Procedure(s)	FF-Z11-203-Q1 Calibration Procedures for dipole validation kits												
Calibration date:	April 1, 2022												
This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.													
All calibrations have been conducted in the closed laboratory facility, environment temperature (22±1)°C and humidity<70%.													
Calibration Equipment used (M&TE critical for calibration)													
Primary Standards	ID # Cat Date (Calibrated by, Certificate No.) Scheduled Calibration												
Power Meter: NRP2	102677 24-Sep-21 (CTTL, No.ZJ1X08326) Sep-22												
Power sensor: NRPFS	104291 24-Sep-21 (CTTL, No.DJ1X08326) Sep-22												
Reference Probe EX3DV4	SN 7307 26-May-21(SPEAG,NO EX3-7307_May21) May-22												
DAE4	SN 1556 12-Jan-22(CTTL-SPEAG,No Z22-60007) Jan-23												
Secondary Standards	ID # Cat Date (Calibrated by, Certificate No.) Scheduled Calibration												
Signal Generator E4438C	MY49071430 13-Jan-22 (CTTL, No.ZJ2X00469) Jan-23												
Network Analyzer E5071C	MY46110673 14-Jan-22 (CTTL, No.ZJ2X00406) Jan-23												
Calibrated by:	<table border="1"> <tr> <th>Name</th> <th>Function</th> <th>Signature</th> </tr> <tr> <td>Zhao Jing</td> <td>SAR Test Engineer</td> <td></td> </tr> <tr> <td>Reviewed by:</td> <td>Lin Hao</td> <td>SAR Test Engineer</td> </tr> <tr> <td>Approved by:</td> <td>Qi Dianyan</td> <td>SAR Project Leader</td> </tr> </table>	Name	Function	Signature	Zhao Jing	SAR Test Engineer		Reviewed by:	Lin Hao	SAR Test Engineer	Approved by:	Qi Dianyan	SAR Project Leader
Name	Function	Signature											
Zhao Jing	SAR Test Engineer												
Reviewed by:	Lin Hao	SAR Test Engineer											
Approved by:	Qi Dianyan	SAR Project Leader											
Issued: April 6, 2022 This calibration certificate shall not be reproduced except in full without written approval of the laboratory													

	
	CAICT CALIBRATION LABORATORY
	Add: No.53 Huayuanli Road, Haodian District, Beijing, 100191, China Tel: +86-10-62164633-3079 Fax: +86-10-62164633-2504 E-mail: cti@chinaul.com http://www.chinaul.com

Glossary:

TSL	tissue simulating liquid
ConWf	sensitivity in TSL / NORMMe,y,z
N/A	not applicable or not measured

Calibration is Performed According to the Following Standards:

- a) IECIEIIE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices- Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020.
- b) KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

Additional Documentation:

- c) DASV45 System Handbook

Methods Applied and Interpretation of Parameters:

- **Measurement Conditions:** Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- **Antenna Parameters with TSL:** The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- **Feed Point Impedance and Return Loss:** These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- **Electrical Delay:** One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- **SAR measured:** SAR measured at the stated antenna input power.
- **SAR normalized:** SAR as measured, normalized to an input power of 1 W at the antenna connector.
- **SAR for nominal TSL parameters:** The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution Corresponds to a coverage probability of approximately 95%.




Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/Terms-and-Conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted under the applicable laws. The results otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.


Attention: To check the authenticity of testing /inspection report or certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.DCcheck@sgs.com.

China 10 Weiyue Road, Development Zone, Kunshan, Jiangsu, China	215300	(86-512) 57355888	(86-512) 57370818	www.sgs.com
中国·江苏·昆山开发区伟业路10号	邮编: 215300	(86-512) 57355888	(86-512) 57370818	china@sgs.com

TTL Speaq CALIBRATION LABORATORY		CAICT	
Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@chinaul.com http://www.chinaul.cn			
Measurement Conditions DASY system configuration, as far as not given on page 1.			
DASY Version	DASY52	52.10.4	
Extrapolation	Advanced Extrapolation		
Phantom	Triple Flat Phantom 5.1C		
Distance Dipole Center - TSL	10 mm	with Spacer	
Zoom Scan Resolution	dx, dy, dz = 5 mm		
Frequency	2450 MHz ± 1 MHz		
Head TSL parameters The following parameters and calculations were applied.			
Nominal Head TSL parameters	Temperature	Permittivity	Conductivity
	22.0 °C	39.2	1.80 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	39.5 ± 6 %	1.79 mho/m ± 6 %
Head TSL temperature change during test	<1.0 °C	---	---
SAR result with Head TSL			
SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition		
SAR measured	250 mW input power	13.2 W/kg	
SAR for nominal Head TSL parameters	normalized to 1W	53.0 W/kg ± 18.8 % (k=2)	
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition		
SAR measured	250 mW input power	6.15 W/kg	
SAR for nominal Head TSL parameters	normalized to 1W	24.7 W/kg ± 18.7 % (k=2)	
Certificate No: Z22-60107 Page 3 of 6			

TTL Speaq CALIBRATION LABORATORY		CAICT	
Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@chinaul.com http://www.chinaul.cn			
Appendix (Additional assessments outside the scope of CNAS L0570)			
Antenna Parameters with Head TSL			
Impedance, transformed to feed point	52.10 ± 3.20jΩ		
Return Loss	~28.5dB		
General Antenna Parameters and Design			
Electrical Delay (one direction)	1.066 ns		
After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.			
The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.			
Additional EUT Data			
Manufactured by	SPEAG		
Certificate No: Z22-60107 Page 4 of 6			

TTL Speaq CALIBRATION LABORATORY		CAICT	
Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@chinaul.com http://www.chinaul.cn			
DASY5 Validation Report for Head TSL Date: 2022-04-01			
Test Laboratory: CTTL, Beijing, China			
DUT: Dipole 2450 MHz; Type: D2450V2; Serial: D2450V2 - SN: 817			
Communication System: UID 0, CW; Frequency: 2450 MHz; Duty Cycle: 1:1			
Medium parameters used: f = 2450 MHz; σ = 1.79 S/m; ε _r = 39.52; ρ = 1000 kg/m ³			
Phantom section: Right Section			
Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)			
DASY5 Configuration:			
• Probe: EX3DV4 - SN7307; ConvF(7.75, 7.75, 7.75) @ 2450 MHz; Calibrated: 2021-05-26			
• Sensor-Surface: 1.4mm (Mechanical Surface Detection)			
• Electronics: DA154 Sn1556; Calibrated: 2022-01-12			
• Phantom: MFP_V5.1C (20deg probe tilt); Type: QD 000 P51 Cx; Serial: 1062			
• DASY52 52.10.4(1535); SEMCAD X 14.6.14(7501)			
Dipole Calibration/Zoom Scan (7x7x7) (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm			
Reference Value = 104.6 V/m; Power Drift = -0.03 dB			
Peak SAR (extrapolated) = 27.0 W/kg			
SAR(1 g) = 13.2 W/kg; SAR(10 g) = 6.15 W/kg			
Smallest distance from peaks to all points 3 dB below = 8.9 mm			
Ratio of SAR at M2 to SAR at M1 = -49.2%			
Maximum value of SAR (measured) = 22.1 W/kg			
			
Certificate No: Z22-60107 Page 5 of 6			

TTL Speaq CALIBRATION LABORATORY		CAICT	
Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504 E-mail: cti@chinaul.com http://www.chinaul.cn			
Impedance Measurement Plot for Head TSL			
			
Certificate No: Z22-60107 Page 6 of 6			


1.11 D2600V2 - SN 1158

TTL Speaq CALIBRATION LABORATORY		CAICT																																	
Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2512 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com.cn http://www.china.ttl.com.cn		Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2579 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com.cn http://www.china.ttl.com.cn																																	
Client: SGS-CN		Certificate No: Z22-40108																																	
CALIBRATION CERTIFICATE																																			
Object: D2600V2 - SN: 1158																																			
Calibration Procedure(s): FF-Z11-003-01 Calibration Procedures for dipole validation kits																																			
Calibration date: March 31, 2022																																			
This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.																																			
All calibrations have been conducted in the closed laboratory facility: environment temperature (22±3)°C and humidity <70%.																																			
Calibration Equipment used (M&TE critical for calibration)																																			
<table border="1"><thead><tr><th>Primary Standards</th><th>ID #</th><th>Cal Date (Calibrated by Certificate No.)</th><th>Scheduled Calibration</th></tr></thead><tbody><tr><td>Power Meter: NRP2</td><td>106277</td><td>24-Sep-21 (CTTL No.J21X08326)</td><td>Sep-22</td></tr><tr><td>Power sensor: NRPBS</td><td>104291</td><td>24-Sep-21 (CTTL No.J21X08326)</td><td>Sep-22</td></tr><tr><td>Reference Probe EX3DVA</td><td>SN 7307</td><td>26-May-21 (SPEAG No EX3-7307_May21)</td><td>May-22</td></tr><tr><td>DAE4</td><td>SN 1556</td><td>12-Jan-22 (CTTL-SPEAG No Z22-60007)</td><td>Jan-23</td></tr><tr><th>Secondary Standards</th><th>ID #</th><th>Cal Date (Calibrated by Certificate No.)</th><th>Scheduled Calibration</th></tr><tr><td>Signal Generator E4438C</td><td>MY49071430</td><td>13-Jan-22 (CTTL No.Z22X00406)</td><td>Jan-23</td></tr><tr><td>Network Analyzer E5071C</td><td>MY48110673</td><td>14-Jan-22 (CTTL No.Z22X00406)</td><td>Jan-23</td></tr></tbody></table>				Primary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration	Power Meter: NRP2	106277	24-Sep-21 (CTTL No.J21X08326)	Sep-22	Power sensor: NRPBS	104291	24-Sep-21 (CTTL No.J21X08326)	Sep-22	Reference Probe EX3DVA	SN 7307	26-May-21 (SPEAG No EX3-7307_May21)	May-22	DAE4	SN 1556	12-Jan-22 (CTTL-SPEAG No Z22-60007)	Jan-23	Secondary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration	Signal Generator E4438C	MY49071430	13-Jan-22 (CTTL No.Z22X00406)	Jan-23	Network Analyzer E5071C	MY48110673	14-Jan-22 (CTTL No.Z22X00406)	Jan-23
Primary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration																																
Power Meter: NRP2	106277	24-Sep-21 (CTTL No.J21X08326)	Sep-22																																
Power sensor: NRPBS	104291	24-Sep-21 (CTTL No.J21X08326)	Sep-22																																
Reference Probe EX3DVA	SN 7307	26-May-21 (SPEAG No EX3-7307_May21)	May-22																																
DAE4	SN 1556	12-Jan-22 (CTTL-SPEAG No Z22-60007)	Jan-23																																
Secondary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration																																
Signal Generator E4438C	MY49071430	13-Jan-22 (CTTL No.Z22X00406)	Jan-23																																
Network Analyzer E5071C	MY48110673	14-Jan-22 (CTTL No.Z22X00406)	Jan-23																																
Calibrated by: Zhao Jing SAR Test Engineer																																			
Reviewed by: Lin Hao SAR Test Engineer																																			
Approved by: Qi Dianyan SAR Project Leader																																			
Issued: April 6, 2022																																			
This calibration certificate shall not be reproduced except in full without written approval of the laboratory.																																			
Certificate No: Z22-40108		Page 1 of 6																																	

TTL Speaq CALIBRATION LABORATORY		CAICT																						
Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-3079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com.cn http://www.china.ttl.com.cn		Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-3079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com.cn http://www.china.ttl.com.cn																						
Measurement Conditions DASY system configuration, as far as not given on page 1:																								
<table border="1"><thead><tr><th>DASY Version</th><th>DASY52</th><th>52.10.4</th></tr></thead><tbody><tr><td>Extrapolation</td><td>Advanced Extrapolation</td><td></td></tr><tr><td>Phantom</td><td>Triple Flat Phantom 5.1C</td><td></td></tr><tr><td>Distance Dipole Center - TSL</td><td>10 mm</td><td>with Spacer</td></tr><tr><td>Zoom Scan Resolution</td><td>dx, dy, dz = 5 mm</td><td></td></tr><tr><td>Frequency</td><td>2600 MHz ± 1 MHz</td><td></td></tr></tbody></table>				DASY Version	DASY52	52.10.4	Extrapolation	Advanced Extrapolation		Phantom	Triple Flat Phantom 5.1C		Distance Dipole Center - TSL	10 mm	with Spacer	Zoom Scan Resolution	dx, dy, dz = 5 mm		Frequency	2600 MHz ± 1 MHz				
DASY Version	DASY52	52.10.4																						
Extrapolation	Advanced Extrapolation																							
Phantom	Triple Flat Phantom 5.1C																							
Distance Dipole Center - TSL	10 mm	with Spacer																						
Zoom Scan Resolution	dx, dy, dz = 5 mm																							
Frequency	2600 MHz ± 1 MHz																							
Head TSL parameters The following parameters and calculations were applied:																								
<table border="1"><thead><tr><th></th><th>Temperature</th><th>Permittivity</th><th>Conductivity</th></tr></thead><tbody><tr><td>Nominal Head TSL parameters</td><td>22.0 °C</td><td>39.0</td><td>1.96 mho/m</td></tr><tr><td>Measured Head TSL parameters</td><td>(22.0 ± 0.2) °C</td><td>36.7 ± 6 %</td><td>1.96 mho/m ± 6 %</td></tr><tr><td>Head TSL temperature change during test</td><td><1.0 °C</td><td>---</td><td>---</td></tr></tbody></table>					Temperature	Permittivity	Conductivity	Nominal Head TSL parameters	22.0 °C	39.0	1.96 mho/m	Measured Head TSL parameters	(22.0 ± 0.2) °C	36.7 ± 6 %	1.96 mho/m ± 6 %	Head TSL temperature change during test	<1.0 °C	---	---					
	Temperature	Permittivity	Conductivity																					
Nominal Head TSL parameters	22.0 °C	39.0	1.96 mho/m																					
Measured Head TSL parameters	(22.0 ± 0.2) °C	36.7 ± 6 %	1.96 mho/m ± 6 %																					
Head TSL temperature change during test	<1.0 °C	---	---																					
SAR result with Head TSL																								
<table border="1"><thead><tr><th></th><th>Condition</th><th></th></tr></thead><tbody><tr><td>SAR averaged over 1 cm² (1 g) of Head TSL</td><td></td><td></td></tr><tr><td>SAR measured</td><td>250 mW input power</td><td>13.7 W/kg</td></tr><tr><td>SAR for nominal Head TSL parameters</td><td>normalized to 1W</td><td>54.8 W/kg ± 18.8 % (k=2)</td></tr><tr><td>SAR averaged over 10 cm² (10 g) of Head TSL</td><td></td><td></td></tr><tr><td>SAR measured</td><td>250 mW input power</td><td>6.12 W/kg</td></tr><tr><td>SAR for nominal Head TSL parameters</td><td>normalized to 1W</td><td>24.6 W/kg ± 18.7 % (k=2)</td></tr></tbody></table>					Condition		SAR averaged over 1 cm ² (1 g) of Head TSL			SAR measured	250 mW input power	13.7 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	54.8 W/kg ± 18.8 % (k=2)	SAR averaged over 10 cm ² (10 g) of Head TSL			SAR measured	250 mW input power	6.12 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	24.6 W/kg ± 18.7 % (k=2)
	Condition																							
SAR averaged over 1 cm ² (1 g) of Head TSL																								
SAR measured	250 mW input power	13.7 W/kg																						
SAR for nominal Head TSL parameters	normalized to 1W	54.8 W/kg ± 18.8 % (k=2)																						
SAR averaged over 10 cm ² (10 g) of Head TSL																								
SAR measured	250 mW input power	6.12 W/kg																						
SAR for nominal Head TSL parameters	normalized to 1W	24.6 W/kg ± 18.7 % (k=2)																						
Certificate No: Z22-40108		Page 3 of 6																						

TTL Speaq CALIBRATION LABORATORY		CAICT	
Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-3079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com.cn http://www.china.ttl.com.cn		Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-3079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com.cn http://www.china.ttl.com.cn	
Glossary: TSL: tissue simulating liquid ConvF: sensitivity in TSL / NORMx.y.z N/A: not applicable or not measured			
Calibration is Performed According to the Following Standards: a) IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices - Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020 b) KDB 865864, "SAR Measurement Requirements for 100 MHz to 6 GHz"			
Additional Documentation: c) DASY4/S System Handbook			
Methods Applied and Interpretation of Parameters: • Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated. • Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis. • Feed Point Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required. • Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required. • SAR measured: SAR measured at the stated antenna input power. • SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector. • SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.			
The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution Corresponds to a coverage probability of approximately 95%.			
Certificate No: Z22-40108		Page 2 of 6	

TTL Speaq CALIBRATION LABORATORY		CAICT					
Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-3079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com.cn http://www.china.ttl.com.cn		Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-3079 Fax: +86-10-42304633-2504 E-mail: cti@china.ttl.com.cn http://www.china.ttl.com.cn					
Appendix (Additional assessments outside the scope of CNAS L0570)							
Antenna Parameters with Head TSL							
<table border="1"><thead><tr><th>Impedance, transformed to feed point</th><th>49.90-6.49jΩ</th></tr><tr><th>Return Loss</th><th>-23.8dB</th></tr></thead></table>				Impedance, transformed to feed point	49.90-6.49jΩ	Return Loss	-23.8dB
Impedance, transformed to feed point	49.90-6.49jΩ						
Return Loss	-23.8dB						
General Antenna Parameters and Design							
<table border="1"><thead><tr><th>Electrical Delay (one direction)</th><th>1.053 ns</th></tr></thead></table>				Electrical Delay (one direction)	1.053 ns		
Electrical Delay (one direction)	1.053 ns						
After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.							
The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.							
Additional EUT Data							
<table border="1"><thead><tr><th>Manufactured by</th><th>SPEAG</th></tr></thead></table>				Manufactured by	SPEAG		
Manufactured by	SPEAG						
Certificate No: Z22-40108		Page 4 of 6					



1.12 D5GHzV2 - SN 1095

TTL S p e a g CALIBRATION LABORATORY		CAICT	
Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62104633-2079 Fax: +86-10-62104633-2504 E-mail: cti@china.ttl.com http://www.china.ttl.com		Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62104633-2079 Fax: +86-10-62104633-2504 E-mail: cti@china.ttl.com http://www.china.ttl.com	
Client: SGS-CN		Certificate No: Z22-60187	
CALIBRATION CERTIFICATE			
Object: D5GHzV2 - SN: 1095			
Calibration Procedure(s): FF-Z11-003-01 Calibration Procedures for dipole validation kits			
Calibration date: June 1, 2022			
This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.			
All calibrations have been conducted in the closed laboratory facility: environment temperature (23±3)°C and humidity <70%.			
Calibration Equipment used (M&TE critical for calibration)			
Primary Standards	ID #	Cal Date (Calibrated by: Certificate No.)	Scheduled Calibration
Power Meter NRP2	106277	24-Sep-21 (CTTL No.J21008328)	Sep-22
Power sensor NRP8S	104201	24-Sep-21 (CTTL No.J21008328)	Sep-22
Reference Probe EX3DV4	SN 7464	26-Jan-22(SPEAG No EX3-7464, Jan22)	Jan-23
DAE4	SN 1556	12-Jan-22(CTTL-SPEAG No.Z22-60007)	Jan-23
Secondary Standards	ID #	Cal Date (Calibrated by: Certificate No.)	Scheduled Calibration
Signal Generator E4438C	MY48071430	13-Jan-22 (CTTL No. J22000408)	Jan-23
Network Analyzer E5071C	MY48110673	14-Jan-22 (CTTL No. J22000406)	Jan-23
Calibrated by: Zhao Jing SAR Test Engineer			
Reviewed by: Lin Hao SAR Test Engineer			
Approved by: Qi Dianyan SAR Project Leader			
Issued: June 6, 2022			
This calibration certificate shall not be reproduced except in full without written approval of the laboratory.			
Certificate No: Z22-60187		Page 1 of 10	

TTL S p e a g CALIBRATION LABORATORY		CAICT	
Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62104633-2079 Fax: +86-10-62104633-2504 E-mail: cti@china.ttl.com http://www.china.ttl.com		Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62104633-2079 Fax: +86-10-62104633-2504 E-mail: cti@china.ttl.com http://www.china.ttl.com	
Glossary:			
TSL	Issue simulating liquid		
ComF	sensitivity in TSL / NORMx,y,z		
N/A	not applicable or not measured		
Calibration is Performed According to the Following Standards:			
a) IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices-Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020			
b) KDB 665664, "SAR Measurement Requirements for 100 MHz to 6 GHz"			
Additional Documentation:			
c) DASYS4.5 System Handbook			
Methods Applied and Interpretation of Parameters:			
• Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.			
• Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.			
• Feed Point Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.			
• Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required.			
• SAR measured: SAR measured at the stated antenna input power.			
• SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector.			
• SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.			
The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution Corresponds to a coverage probability of approximately 95%.			
Certificate No: Z22-60187		Page 2 of 10	



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/terms-and-conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10 Weiye Road, Development Zone, Kunshan, Jiangsu, China 215300 t(86-512) 57355888 f(86-512) 57370818 www.sgs.com.cn
中国·江苏·昆山开发区伟业路10号 邮编: 215300 t(86-512) 57355888 f(86-512) 57370818 sgs.china@sgs.com

TTL In Collaboration with
s p e a q
CALIBRATION LABORATORY

Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62302117
E-mail: csl@speaq.com.cn http://www.csl.ac.cn

CAICT

Measurement Conditions

DAISY system configuration, as for as not given on page 1.

DAISY Version	DAISY2	52.10.4
Extrapolation	Advanced Extrapolation	
Phantom	Triple Flat Phantom 5.1C	
Distance Dipole Center - TSL	10 mm	with Spacer
Zoom Scan Resolution	dx, dy = 4 mm, dz = 1.4 mm	Graded Ratio = 1.4 (Z direction)
Frequency	5200 MHz ± 1 MHz 5300 MHz ± 1 MHz 5500 MHz ± 1 MHz 5600 MHz ± 1 MHz 5800 MHz ± 1 MHz	

Head TSL parameters at 5200MHz

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	36.0	4.66 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	35.4 ± 6 %	4.62 mho/m ± 6 %
Head TSL temperature change during test	<1.0 °C	---	---

SAR result with Head TSL at 5200MHz

	Condition	
SAR averaged over 1 cm ³ (1 g) of Head TSL	250 mW input power	7.79 W/kg
SAR measured	250 mW input power	77.4 W/kg ± 24.4 % (k=2)
SAR for nominal Head TSL parameters	normalized to 1W	77.4 W/kg ± 24.4 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition	
SAR measured	250 mW input power	2.22 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	22.1 W/kg ± 24.2 % (k=2)

Certificate No: Z22-40187

Page 3 of 10

TTL In Collaboration with
s p e a q
CALIBRATION LABORATORY

Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62302117
E-mail: csl@speaq.com.cn http://www.csl.ac.cn

CAICT

Head TSL parameters at 5300MHz

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	35.9	4.76 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	35.2 ± 6 %	4.73 mho/m ± 6 %
Head TSL temperature change during test	<1.0 °C	---	---

SAR result with Head TSL at 5300MHz

	Condition	
SAR averaged over 1 cm ³ (1 g) of Head TSL	100 mW input power	7.94 W/kg
SAR measured	100 mW input power	79.1 W/kg ± 24.4 % (k=2)
SAR for nominal Head TSL parameters	normalized to 1W	79.1 W/kg ± 24.4 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition	
SAR measured	100 mW input power	2.27 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	22.6 W/kg ± 24.2 % (k=2)

Head TSL parameters at 5500MHz

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	35.8	4.96 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	34.8 ± 6 %	4.94 mho/m ± 6 %
Head TSL temperature change during test	<1.0 °C	---	---

SAR result with Head TSL at 5500MHz

	Condition	
SAR averaged over 1 cm ³ (1 g) of Head TSL	100 mW input power	6.29 W/kg
SAR measured	100 mW input power	62.5 W/kg ± 24.4 % (k=2)
SAR for nominal Head TSL parameters	normalized to 1W	62.5 W/kg ± 24.4 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition	
SAR measured	100 mW input power	2.34 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	23.3 W/kg ± 24.2 % (k=2)

Certificate No: Z22-40187

Page 4 of 10

TTL In Collaboration with
s p e a q
CALIBRATION LABORATORY

Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62302117
E-mail: csl@speaq.com.cn http://www.csl.ac.cn

CAICT

Head TSL parameters at 5600MHz

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	35.5	5.07 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	34.7 ± 6 %	5.05 mho/m ± 6 %
Head TSL temperature change during test	<1.0 °C	---	---

SAR result with Head TSL at 5600MHz

	Condition	
SAR averaged over 1 cm ³ (1 g) of Head TSL	100 mW input power	8.12 W/kg
SAR measured	100 mW input power	80.8 W/kg ± 24.4 % (k=2)
SAR for nominal Head TSL parameters	normalized to 1W	80.8 W/kg ± 24.4 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition	
SAR measured	100 mW input power	2.30 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	22.9 W/kg ± 24.2 % (k=2)

Head TSL parameters at 5800MHz

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	35.3	5.27 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	34.4 ± 6 %	5.25 mho/m ± 6 %
Head TSL temperature change during test	<1.0 °C	---	---

SAR result with Head TSL at 5800MHz

	Condition	
SAR averaged over 1 cm ³ (1 g) of Head TSL	100 mW input power	7.71 W/kg
SAR measured	100 mW input power	76.7 W/kg ± 24.4 % (k=2)
SAR for nominal Head TSL parameters	normalized to 1W	76.7 W/kg ± 24.4 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Head TSL	Condition	
SAR measured	100 mW input power	2.16 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	21.8 W/kg ± 24.2 % (k=2)

Certificate No: Z22-40187

Page 3 of 10

TTL In Collaboration with
s p e a q
CALIBRATION LABORATORY

Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
Tel: +86-10-62302117
E-mail: csl@speaq.com.cn http://www.csl.ac.cn

CAICT

Appendix (Additional assessments outside the scope of CNAS L0570)

Antenna Parameters with Head TSL at 5200MHz

Impedance, transformed to feed point	48.1D-5.03jΩ
Return Loss	-23.6dB

Antenna Parameters with Head TSL at 5300MHz

Impedance, transformed to feed point	47.8D-2.42jΩ
Return Loss	-28.5dB

Antenna Parameters with Head TSL at 5500MHz

Impedance, transformed to feed point	50.3D-4.26jΩ
Return Loss	-27.4dB

Antenna Parameters with Head TSL at 5600MHz



Impedance, transformed to feed point	54.5D-4.80jΩ
Return Loss	-24.0dB

Antenna Parameters with Head TSL at 5800MHz

Impedance, transformed to feed point	51.5D-5.61jΩ
Return Loss	-24.9dB

Certificate No: Z22-40187

Page 5 of 10

<p>In Collaboration with TTL s p e a q CALIBRATION LABORATORY</p> <p>CAICT</p> <p>Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62302117 E-mail: cn@sgs.com http://www.caict.ac.cn</p> <p>General Antenna Parameters and Design</p> <table border="1"><tr><td>Electrical Delay (one direction)</td><td>1.101 ns</td></tr></table> <p>After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.</p> <p>The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.</p> <p>No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.</p> <p>Additional EUT Data</p> <table border="1"><tr><td>Manufactured by</td><td>SPEAG</td></tr></table> <p>Certificate No: Z22-60187 Page 7 of 10</p>	Electrical Delay (one direction)	1.101 ns	Manufactured by	SPEAG	<p>In Collaboration with TTL s p e a q CALIBRATION LABORATORY</p> <p>CAICT</p> <p>Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62302117 E-mail: cn@sgs.com http://www.caict.ac.cn</p> <p>DASY5 Validation Report for Head TSL</p> <p>Test Laboratory: CTTL, Beijing, China Date: 2022-06-01</p> <p>DUT: Dipole 5GHz; Type: D5GHzV2; Serial: D5GHzV2 - SN: 1095</p> <p>Communication System: CW; Frequency: 5200 MHz; Frequency: 5300 MHz; Frequency: 5500 MHz; Frequency: 5600 MHz; Frequency: 5800 MHz; Duty Cycle: 1:1 Medium parameters used: $f = 5200 \text{ MHz}$; $\sigma = 4.62 \text{ S/m}$; $\epsilon = 35.19$; $\rho = 1000 \text{ kg/m}^3$ Medium parameters used: $f = 5300 \text{ MHz}$; $\sigma = 4.73 \text{ S/m}$; $\epsilon = 35.19$; $\rho = 1000 \text{ kg/m}^3$ Medium parameters used: $f = 5500 \text{ MHz}$; $\sigma = 4.939 \text{ S/m}$; $\epsilon = 34.83$; $\rho = 1000 \text{ kg/m}^3$ Medium parameters used: $f = 5600 \text{ MHz}$; $\sigma = 5.051 \text{ S/m}$; $\epsilon = 34.89$; $\rho = 1000 \text{ kg/m}^3$ Medium parameters used: $f = 5800 \text{ MHz}$; $\sigma = 5.247 \text{ S/m}$; $\epsilon = 34.42$; $\rho = 1000 \text{ kg/m}^3$ Phantom section: Right Section Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007) DASY5 Configuration:</p> <ul style="list-style-type: none">Probe: EX3DV4 - SN7484; ConvF(5.6, 5.6, 5.6) @ 5200 MHz; ConvF(5.32, 5.32, 5.32) @ 5300 MHz; ConvF(5.11, 5.11, 5.11) @ 5500 MHz; ConvF(4.91, 4.91, 4.91) @ 5600 MHz; ConvF(5, 5, 5) @ 5800 MHz; Calibrated: 2022-01-26Sensor-Surface: 1.4mm (Mechanical Surface Detection)Electronics: DA64 Sn1556; Calibrated: 2022-01-12Phantom: MPF_V5.1C (20deg probe tilt); Type: QD 000 P51 Cx; Serial: 1062DASY52 52.10.4(1535); SEMCAD X 14.6.14(7601) <p>Dipole Calibration /Pin=100mW, d=10mm, f=5200 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 60.80 V/m; Power Drift = -0.06 dB Peak SAR (extrapolated) = 29.8 W/kg SAR(1 g) = 7.79 W/kg; SAR(10 g) = 2.22 W/kg Smallest distance from peaks to all points 3 dB below = 7.2 mm Ratio of SAR at M2 to SAR at M1 = 66.8% Maximum value of SAR (measured) = 18.3 W/kg</p> <p>Dipole Calibration /Pin=100mW, d=10mm, f=5300 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 61.08 V/m; Power Drift = -0.07 dB Peak SAR (extrapolated) = 31.5 W/kg SAR(1 g) = 7.94 W/kg; SAR(10 g) = 2.27 W/kg Smallest distance from peaks to all points 3 dB below = 7.2 mm Ratio of SAR at M2 to SAR at M1 = 65.5% Maximum value of SAR (measured) = 19.0 W/kg</p> <p>Certificate No: Z22-60187 Page 8 of 10</p>
Electrical Delay (one direction)	1.101 ns				
Manufactured by	SPEAG				
<p>In Collaboration with TTL s p e a q CALIBRATION LABORATORY</p> <p>CAICT</p> <p>Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62302117 E-mail: cn@sgs.com http://www.caict.ac.cn</p> <p>Dipole Calibration /Pin=100mW, d=10mm, f=5500 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 61.92 V/m; Power Drift = -0.08 dB Peak SAR (extrapolated) = 34.7 W/kg SAR(1 g) = 8.20 W/kg; SAR(10 g) = 2.34 W/kg Smallest distance from peaks to all points 3 dB below = 7.2 mm Ratio of SAR at M2 to SAR at M1 = 63.9% Maximum value of SAR (measured) = 20.2 W/kg</p> <p>Dipole Calibration /Pin=100mW, d=10mm, f=5600 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 65.08 V/m; Power Drift = -0.07 dB Peak SAR (extrapolated) = 35.2 W/kg SAR(1 g) = 8.12 W/kg; SAR(10 g) = 2.3 W/kg Smallest distance from peaks to all points 3 dB below = 7.2 mm Ratio of SAR at M2 to SAR at M1 = 62.5% Maximum value of SAR (measured) = 19.1 W/kg</p> <p>Dipole Calibration /Pin=100mW, d=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 62.13 V/m; Power Drift = -0.06 dB Peak SAR (extrapolated) = 34.8 W/kg SAR(1 g) = 7.71 W/kg; SAR(10 g) = 2.16 W/kg Smallest distance from peaks to all points 3 dB below = 7.2 mm Ratio of SAR at M2 to SAR at M1 = 61.6% Maximum value of SAR (measured) = 18.7 W/kg</p> <p>0 dB = 18.7 W/kg = 12.72 dBW/kg</p>  <p>Certificate No: Z22-60187 Page 9 of 10</p>	<p>In Collaboration with TTL s p e a q CALIBRATION LABORATORY</p> <p>CAICT</p> <p>Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62302117 E-mail: cn@sgs.com http://www.caict.ac.cn</p> <p>Impedance Measurement Plot for Head TSL</p>  <p>Certificate No: Z22-60187 Page 10 of 10</p>				

2 DAE4 - SN 910

<p>In Collaboration with TTL s p e a g CALIBRATION LABORATORY</p> <p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62304633-2117 E-mail: emf@caict.ac.cn http://www.caict.ac.cn</p> <p>CAICT 中国计量科学研究院 CALIBRATION CNAS (L357)</p> <p>Client: Auden Certificate No: Z22-60275</p> <p>CALIBRATION CERTIFICATE</p> <p>Object: DAE4 - SN: 910</p> <p>Calibration Procedure(s): FF-Z11-002-01 Calibration Procedure for the Data Acquisition Electronics (DAEx)</p> <p>Calibration date: July 14, 2022</p> <p>This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements(SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.</p> <p>All calibrations have been conducted in the closed laboratory facility, environment temperature(22±3)°C and humidity<70%.</p> <p>Calibration Equipment used (M&TE critical for calibration)</p> <table border="1"> <thead> <tr> <th>Primary Standards</th> <th>ID #</th> <th>Cal Date(Calibrated by, Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>Process Calibrator 753</td> <td>1971018</td> <td>14-Jun-22 (CTTL, No.J22X04180)</td> <td>Jun-23</td> </tr> </tbody> </table> <p>Calibrated by: Yu Zongying SAR Test Engineer</p> <p>Reviewed by: Lin Hao SAR Test Engineer</p> <p>Approved by: Qi Dianyan SAR Project Leader</p> <p>Issued July 20, 2022</p> <p>This calibration certificate shall not be reproduced except in full without written approval of the laboratory.</p> <p>Certificate No: Z22-60275 Page 1 of 3</p>	Primary Standards	ID #	Cal Date(Calibrated by, Certificate No.)	Scheduled Calibration	Process Calibrator 753	1971018	14-Jun-22 (CTTL, No.J22X04180)	Jun-23	<p>In Collaboration with TTL s p e a g CALIBRATION LABORATORY</p> <p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62304633-2117 E-mail: emf@caict.ac.cn http://www.caict.ac.cn</p> <p>CAICT</p> <p>Glossary:</p> <p>DAE: data acquisition electronics</p> <p>Connector angle: information used in DASY system to align probe sensor X to the robot coordinate system.</p> <p>Methods Applied and Interpretation of Parameters:</p> <ul style="list-style-type: none"> DC Voltage Measurement: Calibration Factor assessed for use in DASY system by comparison with a calibrated instrument traceable to national standards. The figure given corresponds to the full scale range of the voltmeter in the respective range. Connector angle: The angle of the connector is assessed measuring the angle mechanically by a tool inserted. Uncertainty is not required. The report provide only calibration results for DAE, it does not contain other performance test results. <p>Certificate No: Z22-60275 Page 2 of 3</p>																
Primary Standards	ID #	Cal Date(Calibrated by, Certificate No.)	Scheduled Calibration																						
Process Calibrator 753	1971018	14-Jun-22 (CTTL, No.J22X04180)	Jun-23																						
<p>In Collaboration with TTL s p e a g CALIBRATION LABORATORY</p> <p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-62304633-2117 E-mail: emf@caict.ac.cn http://www.caict.ac.cn</p> <p>CAICT</p> <p>DC Voltage Measurement</p> <p>A/D Converter Resolution nominal</p> <table border="1"> <thead> <tr> <th>High Range</th> <th>Full Range</th> <th>Low Range</th> <th>Full Range</th> </tr> </thead> <tbody> <tr> <td>1LSB = 6.1μV</td> <td>-100...+300 mV</td> <td>1LSB = 61μV</td> <td>-1...+3mV</td> </tr> </tbody> </table> <p>DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec.</p> <table border="1"> <thead> <tr> <th>Calibration Factors</th> <th>X</th> <th>Y</th> <th>Z</th> </tr> </thead> <tbody> <tr> <td>High Range</td> <td>403.347 ± 0.15% (k=2)</td> <td>402.759 ± 0.15% (k=2)</td> <td>403.242 ± 0.15% (k=2)</td> </tr> <tr> <td>Low Range</td> <td>3.98151 ± 0.7% (k=2)</td> <td>3.94044 ± 0.7% (k=2)</td> <td>3.94827 ± 0.7% (k=2)</td> </tr> </tbody> </table> <p>Connector Angle</p> <table border="1"> <thead> <tr> <th>Connector Angle to be used in DASY system</th> <th></th> </tr> </thead> <tbody> <tr> <td></td> <td>269° ± 1°</td> </tr> </tbody> </table> <p>Certificate No: Z22-60275 Page 3 of 3</p>	High Range	Full Range	Low Range	Full Range	1LSB = 6.1μV	-100...+300 mV	1LSB = 61μV	-1...+3mV	Calibration Factors	X	Y	Z	High Range	403.347 ± 0.15% (k=2)	402.759 ± 0.15% (k=2)	403.242 ± 0.15% (k=2)	Low Range	3.98151 ± 0.7% (k=2)	3.94044 ± 0.7% (k=2)	3.94827 ± 0.7% (k=2)	Connector Angle to be used in DASY system			269° ± 1°	
High Range	Full Range	Low Range	Full Range																						
1LSB = 6.1μV	-100...+300 mV	1LSB = 61μV	-1...+3mV																						
Calibration Factors	X	Y	Z																						
High Range	403.347 ± 0.15% (k=2)	402.759 ± 0.15% (k=2)	403.242 ± 0.15% (k=2)																						
Low Range	3.98151 ± 0.7% (k=2)	3.94044 ± 0.7% (k=2)	3.94827 ± 0.7% (k=2)																						
Connector Angle to be used in DASY system																									
	269° ± 1°																								




3 EX3DV4 - SN 7767

Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland		S Schweizerischer Kalibrierdienst C Service suisse d'étalonnage S Service svizzero di taratura S Swiss Calibration Service																																	
Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates		Accreditation No.: SCS 0108																																	
Client: SGS-CN (Auden)	Certificate No: EX-7767_Oct22																																		
CALIBRATION CERTIFICATE																																			
Object: EX3DV4 - SN:7767																																			
Calibration procedures: QA CAL-01.v9, QA CAL-12.v9, QA CAL-14.v6, QA CAL-23.v5, QA CAL-25.v7 Calibration procedure for dosimetric E-field probes																																			
Calibration date: October 28, 2022																																			
This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate. All calibrations have been conducted in the closed laboratory facility: environment temperature (22 ± 1) °C and humidity < 70%. Calibration Equipment used (MATE critical for calibration)																																			
<table border="1"><thead><tr><th>Primary Standards</th><th>ID</th><th>Cal Date (Certificate No.)</th><th>Scheduled Calibration</th></tr></thead><tbody><tr><td>Power meter NRP</td><td>SN 102778</td><td>14-Apr-22 (No. 217-05858-0584)</td><td>Apr-23</td></tr><tr><td>Power sensor NRP-291</td><td>SN 102844</td><td>04-Apr-22 (No. 217-05858-0584)</td><td>Apr-23</td></tr><tr><td>DOF DAK-EX (imp/imp)</td><td>SN 1246</td><td>28-Oct-22 (DOF-DAK-EX-1246_Oct22)</td><td>Oct-23</td></tr><tr><td>DOF DAK-12</td><td>SN 1018</td><td>28-Oct-22 (DOF-DAK-12-1018_Oct22)</td><td>Oct-23</td></tr><tr><td>Reference 25 dB Attenuator</td><td>SN 02582 (20v)</td><td>04-Apr-22 (No. 217-05857)</td><td>Apr-23</td></tr><tr><td>DAK4</td><td>SN 166</td><td>19-Oct-22 (No. 2464-480_Oct22)</td><td>Oct-23</td></tr><tr><td>Reference Probe E532V2</td><td>SN 3013</td><td>27-Dec-21 (No. E53-3013_Oct21)</td><td>Dec-22</td></tr></tbody></table>				Primary Standards	ID	Cal Date (Certificate No.)	Scheduled Calibration	Power meter NRP	SN 102778	14-Apr-22 (No. 217-05858-0584)	Apr-23	Power sensor NRP-291	SN 102844	04-Apr-22 (No. 217-05858-0584)	Apr-23	DOF DAK-EX (imp/imp)	SN 1246	28-Oct-22 (DOF-DAK-EX-1246_Oct22)	Oct-23	DOF DAK-12	SN 1018	28-Oct-22 (DOF-DAK-12-1018_Oct22)	Oct-23	Reference 25 dB Attenuator	SN 02582 (20v)	04-Apr-22 (No. 217-05857)	Apr-23	DAK4	SN 166	19-Oct-22 (No. 2464-480_Oct22)	Oct-23	Reference Probe E532V2	SN 3013	27-Dec-21 (No. E53-3013_Oct21)	Dec-22
Primary Standards	ID	Cal Date (Certificate No.)	Scheduled Calibration																																
Power meter NRP	SN 102778	14-Apr-22 (No. 217-05858-0584)	Apr-23																																
Power sensor NRP-291	SN 102844	04-Apr-22 (No. 217-05858-0584)	Apr-23																																
DOF DAK-EX (imp/imp)	SN 1246	28-Oct-22 (DOF-DAK-EX-1246_Oct22)	Oct-23																																
DOF DAK-12	SN 1018	28-Oct-22 (DOF-DAK-12-1018_Oct22)	Oct-23																																
Reference 25 dB Attenuator	SN 02582 (20v)	04-Apr-22 (No. 217-05857)	Apr-23																																
DAK4	SN 166	19-Oct-22 (No. 2464-480_Oct22)	Oct-23																																
Reference Probe E532V2	SN 3013	27-Dec-21 (No. E53-3013_Oct21)	Dec-22																																
<table border="1"><thead><tr><th>Secondary Standards</th><th>ID</th><th>Check Date (in house)</th><th>Scheduled Check</th></tr></thead><tbody><tr><td>Power meter E4418B</td><td>SN 0841938814</td><td>05-Apr-18 (in house check Jun-20)</td><td>In house check Jun-24</td></tr><tr><td>Power sensor E4418A</td><td>SN 1414148607</td><td>05-Apr-18 (in house check Jun-20)</td><td>In house check Jun-24</td></tr><tr><td>Power sensor E4415A</td><td>SN 10071210</td><td>05-Apr-18 (in house check Jun-20)</td><td>In house check Jun-24</td></tr><tr><td>RF generator HP 8448C</td><td>SN 1068420170</td><td>01-Apr-99 (in house check Jun-20)</td><td>In house check Jun-24</td></tr><tr><td>Network Analyzer E6396A</td><td>SN 1084108477</td><td>31-Mar-14 (in house check Oct-22)</td><td>In house check Oct-24</td></tr></tbody></table>				Secondary Standards	ID	Check Date (in house)	Scheduled Check	Power meter E4418B	SN 0841938814	05-Apr-18 (in house check Jun-20)	In house check Jun-24	Power sensor E4418A	SN 1414148607	05-Apr-18 (in house check Jun-20)	In house check Jun-24	Power sensor E4415A	SN 10071210	05-Apr-18 (in house check Jun-20)	In house check Jun-24	RF generator HP 8448C	SN 1068420170	01-Apr-99 (in house check Jun-20)	In house check Jun-24	Network Analyzer E6396A	SN 1084108477	31-Mar-14 (in house check Oct-22)	In house check Oct-24								
Secondary Standards	ID	Check Date (in house)	Scheduled Check																																
Power meter E4418B	SN 0841938814	05-Apr-18 (in house check Jun-20)	In house check Jun-24																																
Power sensor E4418A	SN 1414148607	05-Apr-18 (in house check Jun-20)	In house check Jun-24																																
Power sensor E4415A	SN 10071210	05-Apr-18 (in house check Jun-20)	In house check Jun-24																																
RF generator HP 8448C	SN 1068420170	01-Apr-99 (in house check Jun-20)	In house check Jun-24																																
Network Analyzer E6396A	SN 1084108477	31-Mar-14 (in house check Oct-22)	In house check Oct-24																																
Calibrated by: Name: Adonia Georgiadou Function: Laboratory Technician Signature: [Signature]																																			
Approved by: Even Kijhn Technical Manager Signature: [Signature]																																			
The calibration certificate shall not be reproduced except in full without written approval of the laboratory.																																			
Certificate No: EX-7767_Oct22 Page 1 of 9																																			

EX3DV4 - SN:7767		October 28, 2022	
Parameters of Probe: EX3DV4 - SN:7767			
Basic Calibration Parameters			
	Sensor X	Sensor Y	Sensor Z
North (µV/(V/m)) ^A	0.87	5.89	0.56
DCP (mV) ^B	103.4	107.3	105.7
Unc (k = 2)			
Calibration Results for Modulation Responses			
USB ^C Communication System Name	A	B	C
	dB	dB	dB
0 CW	X 0.00	0.00	1.00
	Y 0.00	0.00	1.00
	Z 0.00	0.00	1.00
The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.			
^A The uncertainties of Item X1.2 do not affect the E-field uncertainty inside TSL (see Page 3).			
^B Uncertainty for probe sensitivity for maximum specified field strength.			
^C Uncertainty is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.			
Certificate No: EX-7767_Oct22 Page 3 of 9			

EX3DV4 - SN:7767		October 28, 2022	
Parameters of Probe: EX3DV4 - SN:7767			
Other Probe Parameters			
Sensor Arrangement		Triangular	
Connector Angle		144.8°	
Mechanical Surface Detection Mode		enabled	
Optical Surface Detection Mode		disabled	
Probe Overall Length		337 mm	
Probe Body Diameter		10 mm	
Tip Diameter		8 mm	
Tip Length		2.5 mm	
Tip Distance		1 mm	
Probe Tip to Sensor X Calibration Point		1 mm	
Probe Tip to Sensor Y Calibration Point		1 mm	
Probe Tip to Sensor Z Calibration Point		1 mm	
Recommended Measurement Distance from Surface		1.4 mm	
Note: Measurement distance from surface can be increased to 3-4 mm for all Area Scan applications.			
Certificate No: EX-7767_Oct22 Page 4 of 9			





Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/Terms-and-Conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10 Weiye Road, Development Zone, Kunshan, Jiangsu, China 215300 t(86-512) 5735 5888 f(86-512) 57370818 www.sgsgroup.com.cn
中国·江苏·昆山开发区伟业路10号 邮编: 215300 t(86-512) 57355888 f(86-512) 57370818 sgs.china@sgs.com

4 Impedance and return loss

Dipole CLA150 SN 4025				
Head Liquid				
Date of Measurement	Return Loss(dB)	Δ %	Impedance (Ω)	$\Delta\Omega$
2021/4/26	-31.4	/	47.8	/
2022/4/26	-32.5	-3.5%	47.1	0.7
2023/4/26	-32.3	-2.87%	46.5	1.3
Dipole D450V3 SN 1103				
Head Liquid				
Date of Measurement	Return Loss(dB)	Δ %	Impedance (Ω)	$\Delta\Omega$
2021/4/21	-23	/	57.1	/
2022/4/26	-23.4	-1.74%	56.6	0.5
2023/4/26	-23.9	-3.91%	56.2	0.9
Dipole D750V3 SN 1188				
Head Liquid				
Date of Measurement	Return Loss(dB)	Δ %	Impedance (Ω)	$\Delta\Omega$
2022/3/29	-28.7	/	53.6	/
2023/3/29	-28.3	1.39%	53.2	0.4
Dipole D835V2 SN 4d114				
Head Liquid				
Date of Measurement	Return Loss(dB)	Δ %	Impedance (Ω)	$\Delta\Omega$
2022/3/31	-25.3	/	48.7	/
2023/3/31	-24.6	2.77%	49.1	0.4
Dipole D900V2 SN 1d079				
Head Liquid				
Date of Measurement	Return Loss(dB)	Δ %	Impedance (Ω)	$\Delta\Omega$
2022/6/7	-23.3	/	48.1	/
2023/6/7	-23.6	-1.29%	48.3	0.2
Dipole D1800V2 SN 2d170				
Head Liquid				
Date of Measurement	Return Loss(dB)	Δ %	Impedance (Ω)	$\Delta\Omega$
2022/3/31	-29.4	/	47.9	/
2023/3/31	-28.9	1.70%	47.2	0.7
Dipole D1900V2 SN 5d136				
Head Liquid				
Date of Measurement	Return Loss(dB)	Δ %	Impedance (Ω)	$\Delta\Omega$
2022/6/7	-22.4	/	51.2	/
Dipole D2000V2 SN 1041				
Head Liquid				
Date of Measurement	Return Loss(dB)	Δ %	Impedance (Ω)	$\Delta\Omega$
2022/6/6	-34.9	/	48.4	/
Dipole D2300V2 SN 1096				
Head Liquid				



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/Terms-and-Conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10 Weiye Road, Development Zone, Kunshan, Jiangsu, China 215300 t(86-512) 5735 5888 f(86-512) 57370818 www.sgs.com.cn
中国·江苏·昆山开发区伟业路10号 邮编: 215300 t(86-512) 57355888 f(86-512) 57370818 sgs.china@sgs.com

Date of Measurement	Return Loss(dB)	Δ %	Impedance (Ω)	$\Delta\Omega$
2022/3/31	-26.6	/	49.2	/
2023/3/31	-27.1	-1.88%	49.4	0.2
Dipole D2450V2 SN 817				
Head Liquid				
Date of Measurement	Return Loss(dB)	Δ %	Impedance (Ω)	$\Delta\Omega$
2022/4/1	-28.5	/	52.1	/
2023/4/1	-28.0	1.75%	51.6	0.5
Dipole D2600V2 SN 1158				
Head Liquid				
Date of Measurement	Return Loss(dB)	Δ %	Impedance (Ω)	$\Delta\Omega$
2022/3/31	-23.8	/	49.9	/
2023/3/31	-23.3	2.10%	50.3	0.4
Dipole D5GHzV2 SN 1095 for 5200				
Head Liquid				
Date of Measurement	Return Loss(dB)	Δ %	Impedance (Ω)	$\Delta\Omega$
2022/6/1	-23.6	/	46.1	/
Dipole D5GHzV2 SN 1095 for 5300				
Head Liquid				
Date of Measurement	Return Loss(dB)	Δ %	Impedance (Ω)	$\Delta\Omega$
2022/6/1	-29.5	/	47.8	/
Dipole D5GHzV2 SN 1095 for 5500				
Head Liquid				
Date of Measurement	Return Loss(dB)	Δ %	Impedance (Ω)	$\Delta\Omega$
2022/6/1	-27.4	/	50.3	/
Dipole D5GHzV2 SN 1095 for 5600				
Head Liquid				
Date of Measurement	Return Loss(dB)	Δ %	Impedance (Ω)	$\Delta\Omega$
2022/6/1	-24.0	/	54.5	/
Dipole D5GHzV2 SN 1095 for 5800				
Head Liquid				
Date of Measurement	Return Loss(dB)	Δ %	Impedance (Ω)	$\Delta\Omega$
2022/6/1	-24.9	/	51.5	/



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/Terms-and-Conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10 Weiye Road, Development Zone, Kunshan, Jiangsu, China 215300 t(86-512) 5735 5888 f(86-512) 57370818 www.sgs.com.cn
 中国·江苏·昆山开发区伟业路10号 邮编: 215300 t(86-512) 57355888 f(86-512) 57370818 sgs.china@sgs.com