

To:
Federal Communications Commission
Equipment Authorization Division
7435 Oakland Mills Road
Columbia, MD 21048
U.S.A.

Date: 1 March 2021

FCC Labeling of Widex model MBR3D

FCC ID: 2AXDT-MBR3D

To Whom It May Concern:

The dimensions of the hearing aid are too small to allow the FCC ID to be printed on the device in a font size that is readable. Consequently, the FCC ID is printed in the user instructions together with the FCC regulatory statements.

Justification is provided by 3 photographs below that show the MBR3D in 3 different angles together with a printout of the FCC ID in font size 4 point that is the minimum readable font size as defined in FCC OET Labelling Guidelines.

Bottom view



Side view



Top view



Thank you for your attention to this matter.

Sincerely,
WSAUD A/S

Carsten Quaade Jensen
RF Design Engineer, Wireless system integration