

Measurement Conditions

DASY system configuration, as far as not given on page 1.

DASY Version	DASY5	V52.10.1
Extrapolation	Advanced Extrapolation	
Phantom	Modular Flat Phantom	
Distance Dipole Center - TSL	10 mm	with Spacer
Zoom Scan Resolution	dx, dy, dz = 5 mm	
Frequency	2450 MHz ± 1 MHz	

Head TSL parameters

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	39.2	1.80 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	37.7 ± 6 %	1.86 mho/m ± 6 %
Head TSL temperature change during test	< 0.5 °C		

SAR result with Head TSL

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	250 mW input power	13.3 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	51.9 W/kg ± 17.0 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Head TSL	condition	
SAR averaged over 10 cm ³ (10 g) of Head TSL SAR measured	condition 250 mW input power	6.18 W/kg

Body TSL parameters

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Body TSL parameters	22.0 °C	52.7	1.95 mho/m
Measured Body TSL parameters	(22.0 ± 0.2) °C	51.8 ± 6 %	2.02 mho/m ± 6 %
Body TSL temperature change during test	< 0.5 °C		

SAR result with Body TSL

SAR averaged over 1 cm ³ (1 g) of Body TSL	Condition	
SAR measured	250 mW input power	13.3 W/kg
SAR for nominal Body TSL parameters	normalized to 1W	52.1 W/kg ± 17.0 % (k=2)

SAR averaged over 10 cm ³ (10 g) of Body TSL	condition	
SAR measured	250 mW input power	6.22 W/kg
SAR for nominal Body TSL parameters	normalized to 1W	24.6 W/kg ± 16.5 % (k=2)

Certificate No: D2450V2-920_Aug18



Appendix (Additional assessments outside the scope of SCS 0108)

Antenna Parameters with Head TSL

Impedance, transformed to feed point	57.3 Ω + 1.9 jΩ
Return Loss	- 23.0 dB

Antenna Parameters with Body TSL

Impedance, transformed to feed point	52.1 Ω + 6.2 jΩ
Return Loss	- 23.9 dB

General Antenna Parameters and Design

53 ns

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.

No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

Additional EUT Data

Manufactured by	SPEAG
Manufactured on	December 19, 2012

Page 4 of 8





DASY5 Validation Report for Head TSL

Date: 23.08.2018

Test Laboratory: SPEAG, Zurich, Switzerland

DUT: Dipole 2450 MHz; Type: D2450V2; Serial: D2450V2 - SN:920


Communication System: UID 0 - CW; Frequency: 2450 MHz Medium parameters used: f = 2450 MHz; $\sigma = 1.86$ S/m; $\varepsilon_r = 37.7$; $\rho = 1000$ kg/m³ Phantom section: Flat Section Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011)

DASY52 Configuration:

- Probe: EX3DV4 SN7349; ConvF(7.88, 7.88, 7.88) @ 2450 MHz; Calibrated: 30.12.2017
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 26.10.2017
- Phantom: Flat Phantom 5.0 (front); Type: QD 000 P50 AA; Serial: 1001
- DASY52 52.10.1(1476); SEMCAD X 14.6.11(7439)

Dipole Calibration for Head Tissue/Pin=250 mW, d=10mm/Zoom Scan (7x7x7)/Cube 0:


Measurement grid: dx=5mm, dy=5mm, dz=5mmReference Value = 115.8 V/m; Power Drift = -0.00 dB Peak SAR (extrapolated) = 26.4 W/kg SAR(1 g) = 13.3 W/kg; SAR(10 g) = 6.18 W/kg Maximum value of SAR (measured) = 21.8 W/kg



Page 5 of 8



Impedance Measurement Plot for Head TSL



Certificate No: D2450V2-920_Aug18

Page 6 of 8





DASY5 Validation Report for Body TSL

Date: 24.08.2018

Test Laboratory: SPEAG, Zurich, Switzerland

DUT: Dipole 2450 MHz; Type: D2450V2; Serial: D2450V2 - SN:920


Communication System: UID 0 - CW; Frequency: 2450 MHz Medium parameters used: f = 2450 MHz; $\sigma = 2.02$ S/m; $\varepsilon_r = 51.8$; $\rho = 1000$ kg/m³ Phantom section: Flat Section Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011)

DASY52 Configuration:

- Probe: EX3DV4 SN7349; ConvF(8.01, 8.01, 8.01) @ 2450 MHz; Calibrated: 30.12.2017
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 26.10.2017
- Phantom: Flat Phantom 5.0 (back); Type: QD 000 P50 AA; Serial: 1002
- DASY52 52.10.1(1476); SEMCAD X 14.6.11(7439)

Dipole Calibration for Body Tissue/Pin=250 mW, d=10mm/Zoom Scan (7x7x7)/Cube 0:


Measurement grid: dx=5mm, dy=5mm, dz=5mm Reference Value = 109.0 V/m; Power Drift = -0.06 dB Peak SAR (extrapolated) = 26.2 W/kg SAR(1 g) = 13.3 W/kg; SAR(10 g) = 6.22 W/kg Maximum value of SAR (measured) = 21.6 W/kg



Page 7 of 8



Impedance Measurement Plot for Body TSL



Certificate No: D2450V2-920_Aug18

Page 8 of 8





Client

Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland

Accredited by the Swiss Accreditation Service (SAS)

DT&C (Dymstec)

The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates



S Schweizerischer Kalibrierdienst Service suisse d'étalonnage Servizio svizzero di taratura S swiss Calibration Service

Accreditation No.: SCS 0108

Certificate No: D2600V2-1016_Feb19

Object	D2600V2 - SN:1	016	
Calibration procedure(s)	QA CAL-05.v11 Calibration Proce	edure for SAR Validation Sources	s between 0.7-3 GHz
Calibration date:	February 27, 201	19	
The measurements and the uncert	ainties with confidence p ed in the closed laborato	tional standards, which realize the physical un probability are given on the following pages ar bry facility: environment temperature $(22 \pm 3)^{\circ 4}$	nd are part of the certificate.
Primary Standards	LID #	Cal Date (Certificate No.)	Scheduled Calibration
Power meter NRP	SN: 104778	04-Apr-18 (No. 217-02672/02673)	Apr-19
ower sensor NRP-Z91	SN: 103244	04-Apr-18 (No. 217-02672)	Apr-19
ALTER THE	SN: 103245		
ower sensor NHP-Z91	014. 1002-10	04-Apr-18 (No. 217-02673)	Apr-19
	SN: 5058 (20k)	04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02682)	Apr-19 Apr-19
Reference 20 dB Attenuator			
eference 20 dB Attenuator ype-N mismatch combination	SN: 5058 (20k)	04-Apr-18 (No. 217-02682)	Apr-19
Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4	SN: 5058 (20k) SN: 5047,2 / 06327	04-Apr-18 (No. 217-02682) 04-Apr-18 (No. 217-02683)	Apr-19 Apr-19
Reference 20 dB Attenuator ype-N mismatch combination Reference Probe EX3DV4	SN: 5058 (20k) SN: 5047,2 / 06327 SN: 7349	04-Apr-18 (No. 217-02682) 04-Apr-18 (No. 217-02683) 31-Dec-18 (No. EX3-7349_Dec18)	Apr-19 Apr-19 Dec-19
Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4	SN: 5058 (20k) SN: 5047.2 / 06327 SN: 7349 SN: 601	04-Apr-18 (No. 217-02682) 04-Apr-18 (No. 217-02683) 31-Dec-18 (No. EX3-7349_Dec18) 04-Oct-18 (No. DAE4-601_Oct18)	Apr-19 Apr-19 Dec-19 Oct-19
Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards	SN: 5058 (20k) SN: 5047.2 / 06327 SN: 7349 SN: 601	04-Apr-18 (No. 217-02682) 04-Apr-18 (No. 217-02683) 31-Dec-18 (No. EX3-7349_Dec18) 04-Oct-18 (No. DAE4-601_Oct18) Check Date (in house)	Apr-19 Apr-19 Dec-19 Oct-19 Scheduled Check
Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E4419B	SN: 5058 (20k) SN: 5047.2 / 06327 SN: 7349 SN: 601 ID # SN: GB39512475	04-Apr-18 (No. 217-02682) 04-Apr-18 (No. 217-02683) 31-Dec-18 (No. EX3-7349_Dec18) 04-Oct-18 (No. DAE4-601_Oct18) Check Date (in house) 07-Oct-15 (in house check Feb-19)	Apr-19 Apr-19 Dec-19 Oct-19 Scheduled Check In house check: Oct-20
eference 20 dB Attenuator ype-N mismatch combination teference Probe EX3DV4 AE4 econdary Standards ower meter E4419B ower sensor HP 8481A ower sensor HP 8481A	SN: 5058 (20k) SN: 5047,2 / 06327 SN: 7349 SN: 601 ID # SN: GB39512475 SN: US37292783	04-Apr-18 (No. 217-02682) 04-Apr-18 (No. 217-02683) 31-Dec-18 (No. EX3-7349_Dec18) 04-Oct-18 (No. DAE4-601_Oct18) Check Date (in house) 07-Oct-15 (in house check Feb-19) 07-Oct-15 (in house check Oct-18)	Apr-19 Apr-19 Dec-19 Oct-19 Scheduled Check In house check: Oct-20 In house check: Oct-20
Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Recondary Standards Rower meter E4419B Rower sensor HP 8481A Rower sensor HP 8481A Regenerator R&S SMT-06	SN: 5058 (20k) SN: 5047.2 / 06327 SN: 7349 SN: 601 ID # SN: GB39512475 SN: US37292783 SN: MY41092317	04-Apr-18 (No. 217-02682) 04-Apr-18 (No. 217-02683) 31-Dec-18 (No. EX3-7349_Dec18) 04-Oct-18 (No. DAE4-601_Oct18) Check Date (in house) 07-Oct-15 (in house check Feb-19) 07-Oct-15 (in house check Oct-18) 07-Oct-15 (in house check Oct-18)	Apr-19 Apr-19 Dec-19 Oct-19 Scheduled Check In house check: Oct-20 In house check: Oct-20 In house check: Oct-20
Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Becondary Standards Power meter E4419B Power sensor HP 8481A	SN: 5058 (20k) SN: 5047.2 / 06327 SN: 7349 SN: 601 ID # SN: GB39512475 SN: US37292783 SN: MY41092317 SN: 100972	04-Apr-18 (No. 217-02682) 04-Apr-18 (No. 217-02683) 31-Dec-18 (No. EX3-7349_Dec18) 04-Oct-18 (No. DAE4-601_Oct18) Check Date (in house) 07-Oct-15 (in house check Feb-19) 07-Oct-15 (in house check Oct-18) 07-Oct-15 (in house check Oct-18) 15-Jun-15 (in house check Oct-18)	Apr-19 Apr-19 Dec-19 Oct-19 Scheduled Check In house check: Oct-20 In house check: Oct-20 In house check: Oct-20 In house check: Oct-20 In house check: Oct-20
Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Recondary Standards Rower meter E4419B Rower sensor HP 8481A Rower sensor HP 8481A RF generator R&S SMT-06 Retwork Analyzer Agilent E8358A	SN: 5058 (20k) SN: 5047.2 / 06327 SN: 7349 SN: 601 ID # SN: GB39512475 SN: US37292783 SN: MY41092317 SN: 100972 SN: US41080477	04-Apr-18 (No. 217-02682) 04-Apr-18 (No. 217-02683) 31-Dec-18 (No. EX3-7349_Dec18) 04-Oct-18 (No. DAE4-601_Oct18) Check Date (in house) 07-Oct-15 (in house check Feb-19) 07-Oct-15 (in house check Oct-18) 07-Oct-15 (in house check Oct-18) 15-Jun-15 (in house check Oct-18) 31-Mar-14 (in house check Oct-18)	Apr-19 Apr-19 Dec-19 Oct-19 Scheduled Check In house check: Oct-20 In house check: Oct-20 In house check: Oct-20 In house check: Oct-20 In house check: Oct-19
Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Recondary Standards Power meter E4419B Power sensor HP 8481A Power sensor HP 8481A Regenerator R&S SMT-06	SN: 5058 (20k) SN: 5047.2 / 06327 SN: 7349 SN: 601 ID # SN: GB39512475 SN: US37292783 SN: MY41092317 SN: 100972 SN: US41080477 Name	04-Apr-18 (No. 217-02682) 04-Apr-18 (No. 217-02683) 31-Dec-18 (No. EX3-7349_Dec18) 04-Oct-18 (No. DAE4-601_Oct18) Check Date (in house) 07-Oct-15 (in house check Feb-19) 07-Oct-15 (in house check Oct-18) 07-Oct-15 (in house check Oct-18) 15-Jun-15 (in house check Oct-18) 31-Mar-14 (in house check Oct-18) Function	Apr-19 Apr-19 Dec-19 Oct-19 Scheduled Check In house check: Oct-20 In house check: Oct-19
Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E4419B Power sensor HP 8481A Power sensor HP 8481A RF generator R&S SMT-06 Network Analyzer Agilent E8358A	SN: 5058 (20k) SN: 5047.2 / 06327 SN: 7349 SN: 601 ID # SN: GB39512475 SN: US37292783 SN: MY41092317 SN: 100972 SN: US41080477 Name	04-Apr-18 (No. 217-02682) 04-Apr-18 (No. 217-02683) 31-Dec-18 (No. EX3-7349_Dec18) 04-Oct-18 (No. DAE4-601_Oct18) Check Date (in house) 07-Oct-15 (in house check Feb-19) 07-Oct-15 (in house check Oct-18) 07-Oct-15 (in house check Oct-18) 15-Jun-15 (in house check Oct-18) 31-Mar-14 (in house check Oct-18) Function	Apr-19 Apr-19 Dec-19 Oct-19 Scheduled Check In house check: Oct-20 In house check: Oct-19
Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power sensor HP 8481A Power sensor HP 8481A Power sensor HP 8481A RF generator R&S SMT-06 Network Analyzer Agilent E8358A Calibrated by:	SN: 5058 (20k) SN: 5047.2 / 06327 SN: 7349 SN: 601 ID # SN: GB39512475 SN: US37292783 SN: MY41092317 SN: 100972 SN: US41080477 Name Jeton Kastrati	04-Apr-18 (No. 217-02682) 04-Apr-18 (No. 217-02683) 31-Dec-18 (No. EX3-7349_Dec18) 04-Oct-18 (No. DAE4-601_Oct18) Check Date (In house) 07-Oct-15 (In house check Feb-19) 07-Oct-15 (In house check Oct-18) 07-Oct-15 (In house check Oct-18) 15-Jun-15 (In house check Oct-18) 31-Mar-14 (In house check Oct-18) Function Laboratory Technician	Apr-19 Apr-19 Dec-19 Oct-19 Scheduled Check In house check: Oct-20 In house check: Oct-19

Certificate No: D2600V2-1016_Feb19

Page 1 of 8



Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland



Schweizerischer Kalibrierdienst

C Service suisse d'étalonnage

S

- Servizio svizzero di taratura
- S Swiss Calibration Service

Accreditation No.: SCS 0108

Accredited by the Swiss Accreditation Service (SAS)

The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Glossary:

TSL	tissue simulating liquid
ConvF	sensitivity in TSL / NORM x,y,z
N/A	not applicable or not measured

Calibration is Performed According to the Following Standards:

- a) IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
- b) IEC 62209-1, "Measurement procedure for the assessment of Specific Absorption Rate (SAR) from hand-held and body-mounted devices used next to the ear (frequency range of 300 MHz to 6 GHz)", July 2016
- c) IEC 62209-2, "Procedure to determine the Specific Absorption Rate (SAR) for wireless communication devices used in close proximity to the human body (frequency range of 30 MHz to 6 GHz)", March 2010
- d) KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

Additional Documentation:

e) DASY4/5 System Handbook

Methods Applied and Interpretation of Parameters:

- Measurement Conditions: Further details are available from the Validation Report at the end
 of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed
 point exactly below the center marking of the flat phantom section, with the arms oriented
 parallel to the body axis.
- Feed Point Impedance and Return Loss: These parameters are measured with the dipole
 positioned under the liquid filled phantom. The impedance stated is transformed from the
 measurement at the SMA connector to the feed point. The Return Loss ensures low
 reflected power. No uncertainty required.
- Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured: SAR measured at the stated antenna input power.
- SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Certificate No: D2600V2-1016_Feb19

Page 2 of 8



Measurement Conditions

DASY system configuration, as far as not given on page 1.

DASY Version	DASY5	V52.10.2
Extrapolation	Advanced Extrapolation	
Phantom	Modular Flat Phantom	
Distance Dipole Center - TSL	10 mm	with Spacer
Zoom Scan Resolution	dx, dy, dz = 5 mm	
Frequency	2600 MHz ± 1 MHz	

Head TSL parameters

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	39.0	1.96 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	37.1 ± 6 %	2.02 mho/m ± 6 %
Head TSL temperature change during test	< 0.5 °C		

SAR result with Head TSL

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	250 mW input power	14.5 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	56.6 W/kg ± 17.0 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Head TSL		
SAR averaged over to cill (to g) of head 13L	condition	
SAR measured	250 mW input power	6.41 W/kg

Body TSL parameters

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Body TSL parameters	22.0 °C	52.5	2.16 mho/m
Measured Body TSL parameters	(22.0 ± 0.2) °C	50.9 ± 6 %	2.21 mho/m ± 6 %
Body TSL temperature change during test	< 0.5 °C		

SAR result with Body TSL

SAR averaged over 1 cm ³ (1 g) of Body TSL	Condition	
SAR measured	250 mW input power	13.6 W/kg
SAR for nominal Body TSL parameters	normalized to 1W	53.5 W/kg ± 17.0 % (k=2)

SAR averaged over 10 cm ³ (10 g) of Body TSL	condition	
SAR measured	250 mW input power	6.04 W/kg
SAR for nominal Body TSL parameters	normalized to 1W	23.9 W/kg ± 16.5 % (k=2)

Certificate No: D2600V2-1016_Feb19



Appendix (Additional assessments outside the scope of SCS 0108)

Antenna Parameters with Head TSL

Impedance, transformed to feed point	49.3 Ω - 6.7 jΩ
Return Loss	- 23.4 dB

Antenna Parameters with Body TSL

Impedance, transformed to feed point	46.3 Ω - 4.5 jΩ
Return Loss	- 24.3 dB

General Antenna Parameters and Design

Electrical Delay (one direction)	1 154 ns
	1.101110

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.

No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

Additional EUT Data

Manufactured by	SPEAG
	or Erica



DASY5 Validation Report for Head TSL

Date: 27.02.2019

Test Laboratory: SPEAG, Zurich, Switzerland

DUT: Dipole 2600 MHz; Type: D2600V2; Serial: D2600V2 - SN:1016


Communication System: UID 0 - CW; Frequency: 2600 MHz Medium parameters used: f = 2600 MHz; σ = 2.02 S/m; ϵ_r = 37.1; ρ = 1000 kg/m³ Phantom section: Flat Section Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011)

DASY52 Configuration:

- Probe: EX3DV4 SN7349; ConvF(7.74, 7.74, 7.74) @ 2600 MHz; Calibrated: 31.12.2018
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 04.10.2018
- Phantom: Flat Phantom 5.0 (front); Type: QD 000 P50 AA; Serial: 1001
- DASY52 52.10.2(1495); SEMCAD X 14.6.12(7450)

Dipole Calibration for Head Tissue/Pin=250 mW, d=10mm/Zoom Scan (7x7x7)/Cube 0:

Measurement grid: dx=5mm, dy=5mm, dz=5mm Reference Value = 120.3 V/m; Power Drift = 0.04 dB Peak SAR (extrapolated) = 30.0 W/kg SAR(1 g) = 14.5 W/kg; SAR(10 g) = 6.41 W/kgMaximum value of SAR (measured) = 24.8 W/kg




Certificate No: D2600V2-1016_Feb19

Page 5 of 8



Impedance Measurement Plot for Head TSL



Certificate No: D2600V2-1016_Feb19

Page 6 of 8



DASY5 Validation Report for Body TSL

Date: 26.02.2019

Test Laboratory: SPEAG, Zurich, Switzerland

DUT: Dipole 2600 MHz; Type: D2600V2; Serial: D2600V2 - SN:1016


Communication System: UID 0 - CW; Frequency: 2600 MHz Medium parameters used: f = 2600 MHz; $\sigma = 2.21$ S/m; $\varepsilon_r = 50.9$; $\rho = 1000$ kg/m³ Phantom section: Flat Section Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011)

DASY52 Configuration:

- Probe: EX3DV4 SN7349; ConvF(7.89, 7.89, 7.89) @ 2600 MHz; Calibrated: 31.12.2018
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 04.10.2018
- Phantom: Flat Phantom 5.0 (back); Type: QD 000 P50 AA; Serial: 1002
- DASY52 52.10.2(1495); SEMCAD X 14.6.12(7450)

Dipole Calibration for Body Tissue/Pin=250 mW, d=10mm/Zoom Scan (7x7x7)/Cube 0:


Measurement grid: dx=5mm, dy=5mm, dz=5mm Reference Value = 108.2 V/m; Power Drift = -0.02 dB Peak SAR (extrapolated) = 28.3 W/kg SAR(1 g) = 13.6 W/kg; SAR(10 g) = 6.04 W/kg Maximum value of SAR (measured) = 23.1 W/kg



Certificate No: D2600V2-1016_Feb19

Page 7 of 8

Impedance Measurement Plot for Body TSL



Certificate No: D2600V2-1016_Feb19

Page 8 of 8





Client

Calibration Laboratory of Schmid & Partner Engineering AG

Zeughausstrasse 43, 8004 Zurich, Switzerland

Accredited by the Swiss Accreditation Service (SAS)

DT&C (Dymstec)

The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates



Schweizerischer Kalibrierdienst s Service suisse d'étalonnage С Servizio svizzero di taratura Swiss Calibration Service

Accreditation No.: SCS 0108

Certificate No: D5GHzV2-1103_Feb19

S

Object	D5GHzV2 - SN:1	1103	
Calibration procedure(s)	QA CAL-22.v4 Calibration Proce	edure for SAR Validation Sources	between 3-6 GHz
Calibration date:	February 28, 201	19	
The measurements and the uncert	tainties with confidence p ed in the closed laborato	ional standards, which realize the physical un probability are given on the following pages ar by facility: environment temperature $(22 \pm 3)^{\circ 4}$	nd are part of the certificate.
Primary Standards	10#	Cal Date (Certificate No.)	Scheduled Calibration
Power meter NRP	SN: 104778	04-Apr-18 (No. 217-02672/02673)	Apr-19
Power sensor NRP-Z91	SN: 103244	04-Apr-18 (No. 217-02672)	Apr-19
AN COURSE OF A COMPANY OF A COM	SN: 103245		
ower sensor NRP-Z91	ON. 103245	04-Apr-18 (No. 217-02673)	ADF-19
이번화 관계에서 관계가 즐기게 다니는 것이다.	SN: 5058 (20k)	04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02682)	Apr-19 Apr-19
Reference 20 dB Attenuator	100000 00000 AL		
Reference 20 dB Attenuator Type-N mismatch combination	SN: 5058 (20k)	04-Apr-18 (No. 217-02682)	Apr-19
Reference 20 dB Attenuator ype-N mismatch combination Reference Probe EX3DV4	SN: 5058 (20k) SN: 5047.2 / 06327	04-Apr-18 (No. 217-02682) 04-Apr-18 (No. 217-02683)	Apr-19 Apr-19
Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4	SN: 5058 (20k) SN: 5047.2 / 06327 SN: 3503	04-Apr-18 (No. 217-02682) 04-Apr-18 (No. 217-02683) 31-Dec-18 (No. EX3-3503_Dec18)	Apr-19 Apr-19 Dec-19
Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards	SN: 5058 (20k) SN: 5047.2 / 06327 SN: 3503 SN: 601	04-Apr-18 (No. 217-02682) 04-Apr-18 (No. 217-02683) 31-Dec-18 (No. EX3-3503_Dec18) 04-Oct-18 (No. DAE4-601_Oct18)	Apr-19 Apr-19 Dec-19 Oct-19
Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E4419B	SN: 5058 (20k) SN: 5047.2 / 06327 SN: 3503 SN: 601	04-Apr-18 (No. 217-02682) 04-Apr-18 (No. 217-02683) 31-Dec-18 (No. EX3-3503_Dec18) 04-Oct-18 (No. DAE4-601_Oct18) Check Date (in house)	Apr-19 Apr-19 Dec-19 Oct-19 Scheduled Check
Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E4419B Power sensor HP 8481A	SN: 5058 (20k) SN: 5047.2 / 06327 SN: 3503 SN: 601 ID # SN: GB41293874	04-Apr-18 (No. 217-02682) 04-Apr-18 (No. 217-02683) 31-Dec-18 (No. EX3-3503_Dec18) 04-Oct-18 (No. DAE4-601_Oct18) Check Date (in house) 06-Apr-16 (in house check Jun-18)	Apr-19 Apr-19 Dec-19 Oct-19 Scheduled Check In house check: Jun-20
Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E4419B Power sensor HP 8481A Power sensor HP 8481A	SN: 5058 (20k) SN: 5047.2 / 06327 SN: 3503 SN: 601 ID # SN: GB41293874 SN: US37292783	04-Apr-18 (No. 217-02682) 04-Apr-18 (No. 217-02683) 31-Dec-18 (No. EX3-3503_Dec18) 04-Oct-18 (No. DAE4-601_Oct18) Check Date (in house) 06-Apr-16 (in house check Jun-18) 07-Oct-15 (in house check Oct-18)	Apr-19 Apr-19 Dec-19 Oct-19 Scheduled Check In house check: Jun-20 In house check: Oct-20
Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E4419B Power sensor HP 8481A Power sensor HP 8481A RF generator R&S SMT-06	SN: 5058 (20k) SN: 5047.2 / 06327 SN: 3503 SN: 601 ID # SN: GB41293674 SN: US37292783 SN: MY41092317	04-Apr-18 (No. 217-02682) 04-Apr-18 (No. 217-02683) 31-Dec-18 (No. EX3-3503_Dec18) 04-Oct-18 (No. DAE4-601_Oct18) Check Date (in house) 06-Apr-16 (in house check Jun-18) 07-Oct-15 (in house check Oct-18) 07-Oct-15 (in house check Oct-18)	Apr-19 Apr-19 Dec-19 Oct-19 Scheduled Check In house check: Jun-20 In house check: Oct-20 In house check: Oct-20
Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E4419B Power sensor HP 8481A Power sensor HP 8481A RF generator R&S SMT-06 Network Analyzer Agilent E8358A	SN: 5058 (20k) SN: 5047.2 / 06327 SN: 3503 SN: 601 ID # SN: GB41293874 SN: US37292783 SN: MY41092317 SN: 100972	04-Apr-18 (No. 217-02682) 04-Apr-18 (No. 217-02683) 31-Dec-18 (No. EX3-3503_Dec18) 04-Oct-18 (No. DAE4-601_Oct18) Check Date (in house) 06-Apr-16 (in house check Jun-18) 07-Oct-15 (in house check Oct-18) 07-Oct-15 (in house check Oct-18) 15-Jun-15 (in house check Oct-18)	Apr-19 Apr-19 Dec-19 Oct-19 Scheduled Check In house check: Jun-20 In house check: Oct-20 In house check: Oct-20 In house check: Oct-20
Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E4419B Power sensor HP 8481A Power sensor HP 8481A RF generator R&S SMT-06	SN: 5058 (20k) SN: 5047.2 / 06327 SN: 3503 SN: 601 ID # SN: GB41293874 SN: US37292783 SN: MY41092317 SN: 100972 SN: US41080477	04-Apr-18 (No. 217-02682) 04-Apr-18 (No. 217-02683) 31-Dec-18 (No. 217-02683) 04-Oct-18 (No. DAE4-601_Oct18) Check Date (in house) 06-Apr-16 (in house check Jun-18) 07-Oct-15 (in house check Oct-18) 07-Oct-15 (in house check Oct-18) 15-Jun-15 (in house check Oct-18) 31-Mar-14 (in house check Oct-18)	Apr-19 Apr-19 Dec-19 Oct-19 Scheduled Check In house check: Jun-20 In house check: Oct-20 In house check: Oct-20 In house check: Oct-20 In house check: Oct-19
Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power sensor HP 8481A Power sensor HP 8481A Power sensor HP 8481A RF generator R&S SMT-06 Network Analyzer Agilent E8358A Calibrated by:	SN: 5058 (20k) SN: 5047.2 / 06327 SN: 3503 SN: 601 ID # SN: GB41293874 SN: US37292783 SN: MY41092317 SN: 100972 SN: US41080477 Name Jeton Kastrati	04-Apr-18 (No. 217-02682) 04-Apr-18 (No. 217-02683) 31-Dec-18 (No. EX3-3503_Dec18) 04-Oct-18 (No. DAE4-601_Oct18) Check Date (in house) 06-Apr-16 (in house check Jun-18) 07-Oct-15 (in house check Oct-18) 07-Oct-15 (in house check Oct-18) 15-Jun-15 (in house check Oct-18) 31-Mar-14 (in house check Oct-18) Function Laboratory Technician	Apr-19 Apr-19 Dec-19 Oct-19 Scheduled Check In house check: Jun-20 In house check: Oct-20 In house check: Oct-20 In house check: Oct-20 In house check: Oct-19
Reference 20 dB Attenuator Type-N mismatch combination Reference Probe EX3DV4 DAE4 Secondary Standards Power meter E4419B Power sensor HP 8481A Power sensor HP 8481A RF generator R&S SMT-06 Network Analyzer Agilent E8358A	SN: 5058 (20k) SN: 5047.2 / 06327 SN: 3503 SN: 601 ID # SN: GB41293874 SN: US37292783 SN: MY41092317 SN: 100972 SN: US41080477 Name	04-Apr-18 (No. 217-02682) 04-Apr-18 (No. 217-02683) 31-Dac-18 (No. EX3-3503_Dec18) 04-Oct-18 (No. DAE4-601_Oct18) Check Date (in house) 06-Apr-16 (in house check Jun-18) 07-Oct-15 (in house check Oct-18) 07-Oct-15 (in house check Oct-18) 15-Jun-15 (in house check Oct-18) 31-Mar-14 (in house check Oct-18) Function	Apr-19 Apr-19 Dec-19 Oct-19 Scheduled Check In house check: Jun-20 In house check: Oct-20 In house check: Oct-20 In house check: Oct-20 In house check: Oct-19

Certificate No: D5GHzV2-1103_Feb19

Page 1 of 16



Calibration Laboratory of Schmid & Partner Engineering AG

Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst C Service suisse d'étalonnage Servizio svizzero di taratura

S Swiss Calibration Service

Accreditation No.: SCS 0108

Accredited by the Swiss Accreditation Service (SAS)

The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Glossary:

TSL	tissue simulating liquid
ConvF	sensitivity in TSL / NORM x,y,z
N/A	not applicable or not measured

Calibration is Performed According to the Following Standards:

- a) IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
- b) IEC 62209-1, "Measurement procedure for the assessment of Specific Absorption Rate (SAR) from hand-held and body-mounted devices used next to the ear (frequency range of 300 MHz to 6 GHz)", July 2016
- c) IEC 62209-2, "Procedure to determine the Specific Absorption Rate (SAR) for wireless communication devices used in close proximity to the human body (frequency range of 30 MHz to 6 GHz)", March 2010
- d) KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

Additional Documentation:

e) DASY4/5 System Handbook

Methods Applied and Interpretation of Parameters:

- Measurement Conditions: Further details are available from the Validation Report at the end
 of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed
 point exactly below the center marking of the flat phantom section, with the arms oriented
 parallel to the body axis.
- Feed Point Impedance and Return Loss: These parameters are measured with the dipole
 positioned under the liquid filled phantom. The impedance stated is transformed from the
 measurement at the SMA connector to the feed point. The Return Loss ensures low
 reflected power. No uncertainty required.
- Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured: SAR measured at the stated antenna input power.
- SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Certificate No: D5GHzV2-1103_Feb19

Page 2 of 16

Measurement Conditions

DASY system configuration, as far as not given on page 1.

DASY Version	DASY5	V52.10.2
Extrapolation	Advanced Extrapolation	
Phantom	Modular Flat Phantom V5.0	
Distance Dipole Center - TSL	10 mm	with Spacer
Zoom Scan Resolution	dx, dy = 4.0 mm, dz = 1.4 mm	Graded Ratio = 1.4 (Z direction)
Frequency	5200 MHz ± 1 MHz 5300 MHz ± 1 MHz 5500 MHz ± 1 MHz 5600 MHz ± 1 MHz 5800 MHz ± 1 MHz	

Head TSL parameters at 5200 MHz

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	36.0	4.66 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	36.1 ± 6 %	4.45 mho/m ± 6 %
Head TSL temperature change during test	< 0.5 °C		

SAR result with Head TSL at 5200 MHz

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	100 mW input power	7.94 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	79.4 W/kg ± 19.9 % (k=2)

SAR averaged over 10 cm ³ (10 g) of Head TSL	condition	
SAR measured	100 mW input power	2.29 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	22.9 W/kg ± 19.5 % (k=2)

Head TSL parameters at 5300 MHz The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	35.9	4.76 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	35.9 ± 6 %	4.55 mho/m ± 6 %
Head TSL temperature change during test	< 0.5 °C		

SAR result with Head TSL at 5300 MHz

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	100 mW input power	8.25 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	82.4 W/kg ± 19.9 % (k=2)

SAR averaged over 10 cm ³ (10 g) of Head TSL	condition	
SAR measured	100 mW input power	2.36 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	23.5 W/kg ± 19.5 % (k=2)

Head TSL parameters at 5500 MHz

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	35.6	4.96 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	35.7 ± 6 %	4.76 mho/m ± 6 %
Head TSL temperature change during test	< 0.5 °C		

SAR result with Head TSL at 5500 MHz

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	100 mW input power	8.41 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	84.0 W/kg ± 19.9 % (k=2)
SAR averaged over 10 cm^3 (10 g) of Head TSL	condition	
SAR measured	100 mW input power	2.39 W/kg

Head TSL parameters at 5600 MHz

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	35.5	5.07 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	35.5 ± 6 %	4.86 mho/m ± 6 %
Head TSL temperature change during test	< 0.5 °C		

SAR result with Head TSL at 5600 MHz

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	100 mW input power	8.42 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	84.0 W/kg ± 19.9 % (k=2)

SAR averaged over 10 cm ³ (10 g) of Head TSL	condition	
SAR measured	100 mW input power	2.41 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	24.0 W/kg ± 19.5 % (k=2)

Head TSL parameters at 5800 MHz

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	35.3	5.27 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	35.2 ± 6 %	5.07 mho/m ± 6 %
Head TSL temperature change during test	< 0.5 °C		

SAR result with Head TSL at 5800 MHz

SAR averaged over 1 cm ³ (1 g) of Head TSL	Condition	
SAR measured	100 mW input power	8.16 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	81.4 W/kg ± 19.9 % (k=2)
SAR averaged over 10 cm ³ (10 g) of Head TSL	condition	
SAR averaged over 10 cm ³ (10 g) of Head TSL SAR measured	condition 100 mW input power	2.32 W/kg

Body TSL parameters at 5200 MHz

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Body TSL parameters	22.0 °C	49.0	5.30 mho/m
Measured Body TSL parameters	(22.0 ± 0.2) °C	47.1 ± 6 %	5.40 mho/m ± 6 %
Body TSL temperature change during test	< 0.5 °C		

SAR result with Body TSL at 5200 MHz

SAR averaged over 1 cm ³ (1 g) of Body TSL	Condition	
SAR measured	100 mW input power	7.61 W/kg
SAR for nominal Body TSL parameters	normalized to 1W	75.5 W/kg ± 19.9 % (k=2)

SAR averaged over 10 cm ³ (10 g) of Body TSL	condition	
SAR measured	100 mW input power	2.14 W/kg
SAR for nominal Body TSL parameters	normalized to 1W	21.2 W/kg ± 19.5 % (k=2)

Body TSL parameters at 5300 MHz

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Body TSL parameters	22.0 °C	48.9	5.42 mho/m
Measured Body TSL parameters	(22.0 ± 0.2) °C	46.9 ± 6 %	5.53 mho/m ± 6 %
Body TSL temperature change during test	< 0.5 °C		

SAR result with Body TSL at 5300 MHz

SAR averaged over 1 cm ³ (1 g) of Body TSL	Condition	
SAR measured	100 mW input power	7.50 W/kg
SAR for nominal Body TSL parameters	normalized to 1W	74.4 W/kg ± 19.9 % (k=2)
	111	a contractor a ta conserva-
SAR averaged over 10 cm ³ (10 g) of Body TSL	condition	
SAR averaged over 10 cm ³ (10 g) of Body TSL SAR measured	condition 100 mW input power	2.11 W/kg

Body TSL parameters at 5500 MHz

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Body TSL parameters	22.0 °C	48.6	5.65 mho/m
Measured Body TSL parameters	(22.0 ± 0.2) °C	46.5 ± 6 %	5.80 mho/m ± 6 %
Body TSL temperature change during test	< 0.5 °C		

SAR result with Body TSL at 5500 MHz

SAR averaged over 1 cm ³ (1 g) of Body TSL	Condition	
SAR measured	100 mW input power	8.02 W/kg
SAR for nominal Body TSL parameters	normalized to 1W	79.6 W/kg ± 19.9 % (k=2)

SAR averaged over 10 cm ³ (10 g) of Body TSL	condition	
SAR measured	100 mW input power	2.23 W/kg
SAR for nominal Body TSL parameters	normalized to 1W	22.1 W/kg ± 19.5 % (k=2)

Body TSL parameters at 5600 MHz

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Body TSL parameters	22.0 °C	48.5	5.77 mho/m
Measured Body TSL parameters	(22.0 ± 0.2) °C	46.4 ± 6 %	5.94 mho/m ± 6 %
Body TSL temperature change during test	< 0.5 °C		

SAR result with Body TSL at 5600 MHz

SAR averaged over 1 cm^3 (1 g) of Body TSL	Condition	
SAR measured	100 mW input power	8.03 W/kg
SAR for nominal Body TSL parameters	normalized to 1W	79.7 W/kg ± 19.9 % (k=2)

SAR averaged over 10 cm^3 (10 g) of Body TSL	condition	
SAR measured	100 mW input power	2.25 W/kg
SAR for nominal Body TSL parameters	normalized to 1W	22.3 W/kg ± 19.5 % (k=2)

Body TSL parameters at 5800 MHz

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Body TSL parameters	22.0 °C	48.2	6.00 mho/m
Measured Body TSL parameters	(22.0 ± 0.2) °C	46.0 ± 6 %	6.22 mho/m ± 6 %
Body TSL temperature change during test	< 0.5 °C		

SAR result with Body TSL at 5800 MHz

SAR averaged over 1 cm^3 (1 g) of Body TSL	Condition	
SAR measured	100 mW input power	7.54 W/kg
SAR for nominal Body TSL parameters	normalized to 1W	74.8 W/kg ± 19.9 % (k=2)

SAR averaged over 10 cm ³ (10 g) of Body TSL	condition	
SAR measured	100 mW input power	2.11 W/kg
SAR for nominal Body TSL parameters	normalized to 1W	20.9 W/kg ± 19.5 % (k=2)

Certificate No: D5GHzV2-1103_Feb19

Page 8 of 16



Appendix (Additional assessments outside the scope of SCS 0108)

Antenna Parameters with Head TSL at 5200 MHz

Impedance, transformed to feed point	51.5 Ω - 6.7 jΩ
Return Loss	- 23.4 dB

Antenna Parameters with Head TSL at 5300 MHz

Impedance, transformed to feed point	49.8 Ω + 0.6 jΩ
Return Loss	- 44.4 dB

Antenna Parameters with Head TSL at 5500 MHz

Impedance, transformed to feed point	48.0 Ω - 4.3 jΩ
Return Loss	- 26.3 dB

Antenna Parameters with Head TSL at 5600 MHz

Impedance, transformed to feed point	56.0 Ω + 0.2 jΩ
Return Loss	- 25.0 dB

Antenna Parameters with Head TSL at 5800 MHz

Impedance, transformed to feed point	51.1 Ω + 1.9 jΩ
Return Loss	- 33.2 dB

Antenna Parameters with Body TSL at 5200 MHz

Impedance, transformed to feed point	52.9 Ω - 5.3 jΩ
Return Loss	- 24.6 dB

Antenna Parameters with Body TSL at 5300 MHz

Impedance, transformed to feed point	50.0 Ω + 2.0 jΩ
Return Loss	- 34.0 dB

Antenna Parameters with Body TSL at 5500 MHz

Impedance, transformed to feed point	48.9 Ω - 4.0 jΩ
Return Loss	- 27.6 dB



Antenna Parameters with Body TSL at 5600 MHz

Impedance, transformed to feed point	57.3 Ω + 1.8 jΩ
Return Loss	- 23.1 dB

Antenna Parameters with Body TSL at 5800 MHz

Impedance, transformed to feed point	51.9 Ω + 1.2 jΩ			
Return Loss	- 33.0 dB			

General Antenna Parameters and Design

Electrical Delay (one direction)	1.208 ns
----------------------------------	----------

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.

No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

Additional EUT Data

|--|



DASY5 Validation Report for Head TSL

Date: 28.02.2019

Test Laboratory: SPEAG, Zurich, Switzerland

DUT: Dipole D5GHzV2; Type: D5GHzV2; Serial: D5GHzV2 - SN:1103

Communication System: UID 0 - CW; Frequency: 5200 MHz, Frequency: 5300 MHz, Frequency: 5500 MHz, Frequency: 5600 MHz, Frequency: 5800 MHz Medium parameters used: f = 5200 MHz; $\sigma = 4.45$ S/m; $\varepsilon_r = 36.1$; $\rho = 1000$ kg/m³, Medium parameters used: f = 5300 MHz; $\sigma = 4.55$ S/m; $\varepsilon_r = 35.9$; $\rho = 1000$ kg/m³, Medium parameters used: f = 5500 MHz; $\sigma = 4.76$ S/m; $\varepsilon_r = 35.7$; $\rho = 1000$ kg/m³, Medium parameters used: f = 5600 MHz; $\sigma = 4.86$ S/m; $\varepsilon_r = 35.5$; $\rho = 1000$ kg/m³, Medium parameters used: f = 5800 MHz; $\sigma = 5.07$ S/m; $\varepsilon_r = 35.2$; $\rho = 1000$ kg/m³ Phantom section: Flat Section Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011)

DASY52 Configuration:

- Probe: EX3DV4 SN3503; ConvF(5.69, 5.69, 5.69) @ 5200 MHz, ConvF(5.45, 5.45, 5.45) @ 5300 MHz, ConvF(5.15, 5.15, 5.15) @ 5500 MHz, ConvF(5, 5, 5) @ 5600 MHz, ConvF(4.96, 4.96, 4.96) @ 5800 MHz; Calibrated: 31.12.2018
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 04.10.2018
- Phantom: Flat Phantom 5.0 (front); Type: QD 000 P50 AA; Serial: 1001
- DASY52 52.10.2(1495); SEMCAD X 14.6.12(7450)

```
Dipole Calibration for Head Tissue/Pin=100mW, dist=10mm, f=5200 MHz/Zoom Scan,
dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm
Reference Value = 76.19 V/m; Power Drift = 0.05 dB
Peak SAR (extrapolated) = 28.1 W/kg
SAR(1 g) = 7.94 W/kg; SAR(10 g) = 2.29 W/kg
Maximum value of SAR (measured) = 18.0 W/kg
```

Dipole Calibration for Head Tissue/Pin=100mW, dist=10mm, f=5300 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 77.28 V/m; Power Drift = -0.02 dB Peak SAR (extrapolated) = 29.3 W/kg SAR(1 g) = 8.25 W/kg; SAR(10 g) = 2.36 W/kg Maximum value of SAR (measured) = 18.9 W/kg

Dipole Calibration for Head Tissue/Pin=100mW, dist=10mm, f=5500 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 76.59 V/m; Power Drift = -0.01 dB Peak SAR (extrapolated) = 32.5 W/kg SAR(1 g) = 8.41 W/kg; SAR(10 g) = 2.39 W/kg Maximum value of SAR (measured) = 20.0 W/kg

Certificate No: D5GHzV2-1103_Feb19


Page 11 of 16

Dt&C

Dipole Calibration for Head Tissue/Pin=100mW, dist=10mm, f=5600 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 77.06 V/m; Power Drift = -0.01 dB Peak SAR (extrapolated) = 31.5 W/kg SAR(1 g) = 8.42 W/kg; SAR(10 g) = 2.41 W/kg Maximum value of SAR (measured) = 19.9 W/kg

Dipole Calibration for Head Tissue/Pin=100mW, dist=10mm, f=5800 MHz/Zoom Scan,


dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 74.97 V/m; Power Drift = 0.02 dB Peak SAR (extrapolated) = 32.4 W/kg SAR(1 g) = 8.16 W/kg; SAR(10 g) = 2.32 W/kg Maximum value of SAR (measured) = 19.6 W/kg



0 dB = 18.0 W/kg = 12.55 dBW/kg



Impedance Measurement Plot for Head TSL



Certificate No: D5GHzV2-1103_Feb19

Page 13 of 16



Date: 28.02.2019

DASY5 Validation Report for Body TSL

Test Laboratory: SPEAG, Zurich, Switzerland

DUT: Dipole D5GHzV2; Type: D5GHzV2; Serial: D5GHzV2 - SN:1103

Communication System: UID 0 - CW; Frequency: 5200 MHz, Frequency: 5300 MHz, Frequency: 5500 MHz, Frequency: 5600 MHz, Frequency: 5800 MHz Medium parameters used: f = 5200 MHz; $\sigma = 5.4$ S/m; $\epsilon r = 47.1$; $\rho = 1000$ kg/m3, Medium parameters used: f = 5300 MHz; $\sigma = 5.53$ S/m; $\epsilon r = 46.9$; $\rho = 1000$ kg/m3, Medium parameters used: f = 5500 MHz; $\sigma = 5.8$ S/m; $\epsilon r = 46.5$; $\rho = 1000$ kg/m3, Medium parameters used: f = 5600 MHz; $\sigma = 5.94$ S/m; $\epsilon r = 46.4$; $\rho = 1000$ kg/m3, Medium parameters used: f = 5800 MHz; $\sigma = 6.22$ S/m; $\epsilon r = 46.9$; $\rho = 1000$ kg/m3, Medium parameters used: f = 5800 MHz; $\sigma = 6.22$ S/m; $\epsilon r = 46.9$; $\rho = 1000$ kg/m3, Medium parameters used: f = 5800 MHz; $\sigma = 6.22$ S/m; $\epsilon r = 46.9$; $\rho = 1000$ kg/m3, Medium parameters used: f = 5800 MHz; $\sigma = 6.22$ S/m; $\epsilon r = 46.9$; $\rho = 1000$ kg/m3, Medium parameters used: f = 5800 MHz; $\sigma = 6.22$ S/m; $\epsilon r = 46.9$; $\rho = 1000$ kg/m3, Medium parameters used: f = 5800 MHz; $\sigma = 6.22$ S/m; $\epsilon r = 46.9$; $\rho = 1000$ kg/m3, Medium parameters used: f = 5800 MHz; $\sigma = 6.22$ S/m; $\epsilon r = 46.9$; $\rho = 1000$ kg/m3, Medium parameters used: f = 5800 MHz; $\sigma = 6.22$ S/m; $\epsilon r = 46.9$; $\rho = 1000$ kg/m3, Medium parameters used: f = 5800 MHz; $\sigma = 6.22$ S/m; $\epsilon r = 46.9$; $\rho = 1000$ kg/m3, Medium parameters used: f = 5800 MHz; $\sigma = 6.22$ S/m; $\epsilon r = 46.9$; $\rho = 1000$ kg/m3, Medium parameters used: f = 5800 MHz; $\sigma = 6.22$ S/m; $\epsilon r = 46.9$; $\rho = 1000$ kg/m3, Medium parameters used: f = 5800 MHz; $\sigma = 6.22$ S/m; $\epsilon r = 46.9$; $\rho = 1000$ kg/m3, Medium parameters used: f = 5800 MHz; $\sigma = 6.22$ S/m; $\epsilon r = 46.9$; $\rho = 1000$ kg/m3, Medium parameters used: f = 5800 MHz; $\sigma = 6.22$ S/m; $\epsilon r = 46.9$; $\rho = 1000$ kg/m3, Medium parameters used: $\epsilon = 5800$ MHz; $\sigma = 6.22$ S/m; $\epsilon = 40.9$; $\rho = 1000$ kg/m3, Medium parameters used: $\epsilon = 5800$ MHz; $\sigma = 6.22$ S/m; $\epsilon = 40.9$; $\rho = 1000$ kg/m3, Medium parameters used: $\epsilon = 5800$ MHz; $\sigma = 5.90$ MHz; $\epsilon = 5.900$ MHz; $\epsilon = 5.900$

DASY52 Configuration:

- Probe: EX3DV4 SN3503; ConvF(5.24, 5.24, 5.24) @ 5200 MHz, ConvF(5.15, 5.15, 5.15) @ 5300 MHz, ConvF(4.75, 4.75, 4.75) @ 5500 MHz, ConvF(4.7, 4.7, 4.7) @ 5600 MHz, ConvF(4.58, 4.58, 4.58) @ 5800 MHz; Calibrated: 31.12.2018
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 04.10.2018
- Phantom: Flat Phantom 5.0 (back); Type: QD 000 P50 AA; Serial: 1002
- DASY52 52.10.2(1495); SEMCAD X 14.6.12(7450)

Dipole Calibration for Body Tissue/Pin=100mW, dist=10mm, f=5200 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 69.63 V/m; Power Drift = -0.05 dB Peak SAR (extrapolated) = 28.8 W/kg SAR(1 g) = 7.61 W/kg; SAR(10 g) = 2.14 W/kg Maximum value of SAR (measured) = 17.5 W/kg

Dipole Calibration for Body Tissue/Pin=100mW, dist=10mm, f=5300 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 67.82 V/m; Power Drift = -0.02 dB Peak SAR (extrapolated) = 29.3 W/kg SAR(1 g) = 7.5 W/kg; SAR(10 g) = 2.11 W/kg Maximum value of SAR (measured) = 17.6 W/kg

Dipole Calibration for Body Tissue/Pin=100mW, dist=10mm, f=5500 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 69.31 V/m; Power Drift = -0.00 dB Peak SAR (extrapolated) = 33.2 W/kg SAR(1 g) = 8.02 W/kg; SAR(10 g) = 2.23 W/kg Maximum value of SAR (measured) = 19.0 W/kg


Certificate No: D5GHzV2-1103_Feb19

Page 14 of 16

Dt&C

Dipole Calibration for Body Tissue/Pin=100mW, dist=10mm, f=5600 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 68.57 V/m; Power Drift = -0.03 dB Peak SAR (extrapolated) = 34.5 W/kg SAR(1 g) = 8.03 W/kg; SAR(10 g) = 2.25 W/kg Maximum value of SAR (measured) = 19.5 W/kg

Dipole Calibration for Body Tissue/Pin=100mW, dist=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=1.4mm Reference Value = 66.27 V/m; Power Drift = -0.01 dB Peak SAR (extrapolated) = 32.6 W/kg SAR(1 g) = 7.54 W/kg; SAR(10 g) = 2.11 W/kg Maximum value of SAR (measured) = 18.3 W/kg




0 dB = 17.5 W/kg = 12.43 dBW/kg

Certificate No: D5GHzV2-1103_Feb19

Page 15 of 16



Impedance Measurement Plot for Body TSL



Certificate No: D5GHzV2-1103_Feb19

Page 16 of 16



APPENDIX C. – SAR Tissue Specifications



The brain and muscle mixtures consist of a viscous gel using hydrox-ethylcellulose (HEC) gelling agent and saline solution (see Table C.1). Preservation with a bactericide is added and visual inspection is made to make sure air bubbles are not trapped during the mixing process. The mixture is calibrated to obtain proper dielectric constant (permittivity) and conductivity of the desired tissue. The mixture characterizations used for the brain and muscle tissue simulating liquids are according to the data by C. Gabriel and G. Harts grove.



Figure C.1 Simulated Tissue

Ingredients		Frequency (MHz)										
(% by weight)	83	35	19	000	24	150	5200	~ 5800				
Tissue Type	Head	Body	Head	Body	Head	Body	Head	Body				
Water	40.19	50.75	55.24	70.23	71.88	73.40	65.52	80.00				
Salt (NaCl)	1.480	0.940	0.310	0.290	0.160	0.060	-	-				
Sugar	57.90	48.21	-	-	-	-	-	-				
HEC	0.250	-	-	-	-	-	-	-				
Bactericide	0.180	0.100	-	-	-	-	-	-				
Triton X-100	-	-	-	-	19.97	-	17.24	-				
DGBE	-	-	44.45	29.48	7.990	26.54	-	-				
Diethylene glycol hexyl ether	-	-	-	-	-	-	17.24	-				
Polysorbate (Tween) 80	-	-	-	-	-	-		20.00				
Target for Dielectric Constant	41.5	55.2	40.0	53.3	39.2	52.7	-	-				
Target for Conductivity (S/m)	0.90	0.97	1.40	1.52	1.80	1.95	-	-				

Table C.1 Composition of the Tissue Equivalent Matter

Salt:	99 % Pure Sodium Chloride	Sugar:	98 % Pure Sucrose
Water:	De-ionized, 16M resistivity	HEC:	Hydroxyethyl Cellulose
DGBE:	99 % Di(ethylene glycol) butyl eth	ner,[2-(2-butoxye	ethoxy) ethanol]
Triton X-100(ultra pure):	Polyethylene glycol mono[4-(1,1,	3,3-tetramethylb	outyl)phenyl] ether

Table C.2 HSL/MSL750 (Head and Body liquids for 700 – 800 MHz)

Item	Head Tissue Simulation Liquids HSL750						
item	Muscle (body) Tissue Simulation Liquids MSL750						
Type No	SLAAH 075, SLAAM 075						
Manufacturer	SPEAG						
The item is composed of the following ingredients:							
H ² O	Water, 35 – 58%						
Sucrose	Sucrose, 40 – 60%						
NaCl	Sodium Chloride, 0 – 6%						
Hydroxyethyl-cellulose	Medium Viscosity (CAS# 9004-62-0), < 0.3%						
Preventol-D7	Preservative: aqueous preparation, (CAS# 55965-84-9), containing 5-chloro-2-methyl-3(2H)-isothiazolone and 2-methyyl- 3(2H)-isothiazolone, 0.1 – 0.6%						

Table C.3 HSL/MSL1750 (Head and Body liquids for 1700 – 1800 MHz)

Item	Head Tissue Simulation Liquids HSL1750
item	Muscle (body) Tissue Simulation Liquids MSL1750
Type No	SLAAH 175, SLAAM 175
Manufacturer	SPEAG
The item is composed of the following ingredients:	
H ² O	Water, 52 – 75%
C8H18O3	Diethylene glycol monobutyl ether (DGBE), 25 – 48%
NaCl	Sodium Chloride, < 1.0%



APPENDIX D. – SAR SYSTEM VALIDATION

SAR System Validation

Per FCC KDB 865664 D02v01r02, SAR system validation status should be documented to confirm measurement accuracy. The SAR systems (including SAR probes, system components and software versions) used for this device were validated against its performance specifications prior to the SAR measurements. Reference dipoles were used with the required tissue- equivalent media for system validation, according to the procedures outlined in FCC KDB 865664 D01v01r04 and IEEE 1528-2013.Since SAR probe calibrations are frequency dependent, each probe calibration point was validated at a frequency within the valid frequency range of the probe calibration point, using the system that normally operates with the probe for routine SAR measurements and according to the required tissue-equivalent media.

A tabulated summary of the system validation status including the validation date(s), measurement frequencies, SAR probes and tissue dielectric parameters has been included.

SAR	Freq.	Data	Probe	Probe	Durk o	AL. Point PERM. COND.		1	CW Validatio	on	MOD. Validation				
System	[MHz]	Date	SN	Туре	Probe C.	AL. Point	(εr) (σ)		Sensi- tivity	Probe Linearity	Probe Isortopy	MOD. Type	Duty Factor	PAR	
D	750	2018.10.15	3933	EX3DV4	750	Head	42.212	0.887	PASS	PASS	PASS	N/A	N/A	N/A	
D	835	2018.10.16	3933	EX3DV4	835	Head	40.978	0.919	PASS	PASS	PASS	GMSK	PASS	N/A	
D	1800	2018.10.17	3933	EX3DV4	1800	Head	39.954	1.397	PASS	PASS	PASS	N/A	N/A	N/A	
D	1900	2018.10.18	3933	EX3DV4	1900	Head	39.547	1.369	PASS	PASS	PASS	GMSK	PASS	N/A	
D	2450	2018.10.19	3933	EX3DV4	2450	Head	38.885	1.851	PASS	PASS	PASS	OFDM/TDD	PASS	PASS	
D	2600	2018.10.22	3933	EX3DV4	2600	Head	38.458	1.977	PASS	PASS	PASS	TDD	PASS	N/A	
F	5200	2018.05.14	3916	EX3DV4	5200	Head	34.856	4.725	PASS	PASS	PASS	OFDM	N/A	PASS	
F	5300	2018.05.15	3916	EX3DV4	5300	Head	34.854	4.856	PASS	PASS	PASS	OFDM	N/A	PASS	
F	5500	2018.05.16	3916	EX3DV4	5500	Head	34.625	5.113	PASS	PASS	PASS	OFDM	N/A	PASS	
F	5600	2018.05.17	3916	EX3DV4	5600	Head	34.265	5.221	PASS	PASS	PASS	OFDM	N/A	PASS	
F	5800	2018.05.18	3916	EX3DV4	5800	Head	34.113	5.414	PASS	PASS	PASS	OFDM	N/A	PASS	
D	750	2018.10.15	3933	EX3DV4	750	Body	54.650	0.965	PASS	PASS	PASS	N/A	N/A	N/A	
D	835	2018.10.16	3933	EX3DV4	835	Body	54.597	0.975	PASS	PASS	PASS	GMSK	PASS	N/A	
D	1800	2018.10.17	3933	EX3DV4	1800	Body	52.381	1.553	PASS	PASS	PASS	N/A	N/A	N/A	
D	1900	2018.10.18	3933	EX3DV4	1900	Body	52.289	1.571	PASS	PASS	PASS	GMSK	PASS	N/A	
D	2450	2018.10.19	3933	EX3DV4	2450	Body	51.77	2.015	PASS	PASS	PASS	OFDM/TDD	PASS	PASS	
D	2600	2018.10.22	3933	EX3DV4	2600	Body	51.142	2.221	PASS	PASS	PASS	TDD	PASS	N/A	
А	5200	2018.08.17	3930	EX3DV4	5200	Body	48.845	5.416	PASS	PASS	PASS	OFDM	N/A	PASS	
А	5300	2018.08.20	3930	EX3DV4	5300	Body	48.745	5.526	PASS	PASS	PASS	OFDM	N/A	PASS	
А	5500	2018.08.21	3930	EX3DV4	5500	Body	48.546	5.708	PASS	PASS	PASS	OFDM	N/A	PASS	
А	5600	2018.08.22	3930	EX3DV4	5600	Body	48.325	5.815	PASS	PASS	PASS	OFDM	N/A	PASS	
А	5800	2018.08.23	3930	EX3DV4	5800	Body	47.151	6.137	PASS	PASS	PASS	OFDM	N/A	PASS	

Table D.1 SAR System Validation Summary

NOTE: While the probes have been calibrated for both a CW and modulated signals, all measurements were performed using communication systems calibrated for CW signals only. Modulations in the table above represent test configurations for which the measurement system has been validated per FCC KDB Publication 865664 D01v01r04 for scenarios when CW probe calibrations are used with other signal types. SAR systems were validated for modulated signals with a periodic duty cycle, such as GMSK, or with a high peak to average ratio (>5 dB), such as OFDM according to KDB 865664.



APPENDIX E. – Downlink LTE CA RF Conducted Powers



E.1 LTE Downlink Only Carrier Aggregation Test Reduction Methodology

SAR test exclusion for LTE downlink Carrier Aggregation is determined by power measurements according to the number of component carriers (CCs) supported by the product implementation. Per April 2018 TCBC Workshop Notes, the following test reduction methodology was applied to determine the combinations required for conducted power measurements.

LTE DL CA Test Reduction Methodology:

- (1) Test supported combinations were arranged by the number of component carriers in columns.
- (2) Any limitations on the PCC or SCC for each combination were identified alongside the combination (e.g. CA_2A-2A-4A-12A, but B12 can only be configured as a SCC).
- (3) Power measurements were performed for "supersets" (LTE CA combinations with multiple components carriers) and any "subsets" (LTE CA combinations with fewer component carriers) that were not completely covered by the supersets.
- (4) Only subsets that have the exact same components as a superset were excluded for measurement.
- (5) When there were certain restrictions on component carriers that existed in the superset that were not applied for the subset, the subset configuration was additionally evaluated.
- (6) Both inter-band and intra-band downlink carrier aggregation scenarios were considered.

Index	2CC	Restriction	Completely Covered by Measurement Superset	Index	3CC	Restriction	Completely Covered by Measurement Superset	Index	4CC	Restriction	Completely Covered by Measurement Superset		Index	5CC	Restriction	Completely Covered by Measurement Superset
2CC #1	CA 2C		3CC #7	3CC #1	CA 2A-2A-4A		4CC #1	4CC #1	CA 2A-2A-4A-4A		No	50	C #1	CA 2A-2A-46D	B46 SCC Only	No
2CC #2	CA 2A-2A		4CC #1	3CC #2	CA 2A-2A-5A		4CC #2	4CC #2	CA 2A-2A-4A-5A		No	50	C #2	CA 2A-5B-30A-66A		No
2CC #3	CA_2A-4A (2)		4CC #1	3CC #3	CA 2A-2A-12A		No	4CC #3	CA 2A-2A-4A-12A	B12 SCC Only	No	50	C #3	CA_2A-5B-66A-66A		No
2CC #4	CA_2A-5A		4CC #2	3CC #4	CA_2A-2A-13A		4CC #8	4CC #4	CA 2A-2A-5A-30A		No	50	C #4	CA_2A-46D-66A	B46 SCC Only	No
2CC #5	CA_2A-7A		4CC #16	3CC #5	CA 2A-2A-29A	B29 SCC Only	4CC #9	4CC #5	CA 2A-2A-5A-66A		No	50	C #5	CA_2A-46A-46C-66A	B46 SCC Only	No
2CC #6	CA_2A-12A(1)		3CC #3	3CC #6	CA_2A-2A-30A		4CC #9	4CC #6	CA 2A-2A-12A-30A	B12 SCC Only	No	50	C #6	CA_41C-41D		No
2CC #7	CA_2A-13A		4CC #31	3CC #7	CA_2C-66A		No	4CC #7	CA 2A-2A-12A-66A	B12 SCC Only	No	50	C #7	CA_46D-66A-66A	B46 SCC Only	No
2CC #8	CA 2A-14A		3CC #27	3CC #8	CA_2A-2A-66A		4CC #7	4CC #8	CA 2A-2A-13A-66A		No	50	C #8			
2CC #9	CA 2A-17A		No	3CC #9	CA 2A-2A-71A	B71 SCC Only	No	4CC #9	CA 2A-2A-29A-30A	B29 SCC Only	No	50	A 9			
2CC #10	CA_2A-29A(2)	B29 SCC Only	4CC #9	3CC #10	CA_2A-4A-4A		4CC #1	4CC #10	CA 2A-2A-66A-66A		No	50	A 10			
2CC #11	CA 2A-30A		4CC #9	3CC #11	CA_2A-4A-5A		4CC #2	4CC #11	CA 2A-4A-4A-5A		No	50	A 11			
2CC #12	CA 2A-46A	B46 SCC Only	5CC #5	3CC #12	CA_2A-4A-7A		4CC #16	4CC #12	CA 2A-4A-4A-12A	B12 SCC Only	No	50	A 12			
2CC #13	CA 2A-66A		5CC #2	3CC #13	CA 2A-4A-12A		No	4CC #13	CA 2A-4A-5B		No	50	A 13			
2CC #14	CA 2A-71A	B71 SCC Only	3CC #17	3CC #14	CA 2A-4A-13A		No	4CC #14	CA 2A-4A-5A-30A		No	50	A 14			
2CC #15	CA 4A-4A		4CC #1	3CC #15	CA 2A-4A-29A	B29 SCC Only	4CC #20	4CC #15	CA 2A-4A-7C		No	50	A 15			
2CC #16	CA_4A-5A(1)		4CC #2	3CC #16	CA 2A-4A-30A		4CC #19	4CC #16	CA 2A-4A-7A-7A		No	50	A 16			
2CC #17	CA 4A-7A(1)		4CC #16	3CC #17	CA 2A-4A-71A	B71 SCC Only	No	4CC #17	CA 2A-4A-7A-12A	B12 SCC Only	No	50	A 17			
2CC #18	CA 4A-12A(2)		4CC #3	3CC #18	CA 2A-5B		4CC #13	4CC #18	CA 2A-4A-12B	B12 SCC Only	No	50	A 18			
2CC #19	CA 4A-13A		3CC #14	3CC #19	CA 2A-5A-30A		4CC #23	4CC #19	CA 2A-4A-12A-30A	B12 SCC Only	No	50	A 19			
2CC #20	CA 4A-17A	B17 SCC Only	No	3CC #20	CA 2A-5A-66A		4CC #23	4CC #20	CA 2A-4A-29A-30A	B29 SCC Only	No	50	A 20			
2CC #21	CA 4A-29A(2)	B29 SCC Only	4CC #20	3CC #21	CA 2A-7A-7A		4CC #16	4CC #21	CA 2A-5B-30A		5CC #2	50	A 21			
2CC #22	CA 4A-30A		4CC #19	3CC #22	CA 2A-7A-12A		No	4CC #22	CA 2A-5B-66A		5CC #3	50	A 22			
2CC #23	CA 4A-46A	B46 SCC Only	4CC #42	3CC #23	CA 2A-12B		No	4CC #23	CA 2A-5A-30A-66A		No	50	A 23			
2CC #24	CA 4A-71A	B71 SCC Only	3CC #43	3CC #24	CA 2A-12A-30A		No	4CC #24	CA 2A-5A-66B		No	50	A 24			
2CC #25	CA 5B		5CC #2	3CC #25	CA 2A-12A-66A		No	4CC #25	CA 2A-5A-66C		No	50	A 25			
2CC #26	CA 5A-25A		No	3CC #26	CA 2A-13A-66A		4CC #31	4CC #26	CA 2A-5A-66A-66A		No	50	A 26			
2CC #27	CA 5A-30A		4CC #23	3CC #27	CA 2A-14A-30A		No	4CC #27	CA 2A-12A-30A-66A	B12 SCC Only	No		A 27			
2CC #28	CA 5A-66A		4CC #23	3CC #28	CA 2A-29A-30A	B29 SCC Only	4CC #9	4CC #28	CA 2A-12A-66A-66A	B12 SCC Only	No	50	A 28			
	CA 7A-7A(1)		4CC #16	3CC #29	CA 2A-30A-66A		4CC #23	4CC #29	CA 2A-13A-66B		No		A 29			
	CA 7A-12A		3CC #22	3CC #30	CA 2A-46C	B46 SCC Only	5CC #5	4CC #30	CA 2A-13A-66C		No	50	A 30			
2CC #31	CA 7A-46A(1)	B46 SCC Only	No	3CC #31	CA 2A-46A-46A	B46 SCC Only	4CC #34	4CC #31	CA 2A-13A-66A-66A		No	50	A 31			
	CA 12B		3CC #23	3CC #32	CA 2A-46A-66A	B46 SCC Only	4CC #34	4CC #32	CA 2A-46D	B46 SCC Only	5CC #1		A 32			
2CC #33	CA 12A-25A		No	3CC #33	CA 2A-66B		4CC #24	4CC #33	CA 2A-46A-46C	B46 SCC Only	5CC #5		A 33			
2CC #34	CA 12A-30A		4CC #4	3CC #34	CA 2A-66C		4CC #25	4CC #34	CA 2A-46A-46A-66A	B46 SCC Only	No	50	A 34			
	CA 12A-66A(1)		3CC #25	3CC #35	CA 2A-66A-66A		4CC #26		CA 2A-46C-66A	B46 SCC Only	5CC #5		A 35			1
2CC #36	CA 13A-46A	B46 SCC Only	No	3CC #36	CA 2A-66A-71A	B71 SCC Only	No	4CC #36	CA 4A-4A-5B		No		A 36			
2CC #37	CA 13A-66A		4CC #8	3CC #37	CA 4A-4A-5A		4CC #37	4CC #37	CA 4A-4A-5A-30A		No		A 37			1
2CC #38	CA 14A-30A		3CC #27	3CC #38	CA 4A-4A-7A(1)		No	4CC #38	CA 4A-4A-12A-30A	B12 SCC Only	No		A 38			1
2CC #39	CA 14A-66A		3CC #66	3CC #39	CA 4A-4A-12A		4CC #38	4CC #39	CA 4A-4A-29A-30A	B29 SCC Only	No		A 39			1
	CA 25A-25A(1)		4CC #49	3CC #40	CA 4A-4A-13A		No	4CC #40	CA 4A-5B-30A		No		A 40			1

Table E.1.1 Example of Exclusion Table for LTE DL CA



E.2 LTE Downlink Only Carrier Aggregation Test Selection and Setup

SAR test exclusion for LTE downlink Carrier Aggregation is determined by power measurements according to the number of component carriers (CCs) supported by the product implementation. For those configurations required by April 2018 TCBC Workshop Notes, conducted power measurements with LTE Carrier Aggregation (CA) (downlink only) active are made in accordance to KDB Publication 941225 D05Av01r02. The RRC connection is only handled by one cell, the primary component carrier (PCC) for downlink and uplink communications. After making a data connection to the PCC, the UE device adds secondary component carrier(s) (SCC) on the downlink only. All uplink communications and acknowledgements remain identical to specifications when downlink carrier aggregation is inactive on the PCC. Additional conducted output powers are measured with the downlink carrier aggregation active for the configuration with highest measured maximum conducted power with downlink carrier aggregation inactive measured among the channel bandwidth, modulation and RB combinations in each frequency band.

Per FCC KDB Publication 941225 D05Av01r02, no SAR measurements are required for carrier aggregation configurations when the average output power with downlink only carrier aggregation active is not more than 0.25 dB higher than the average output power with downlink only carrier aggregation inactive.

General PCC and SCC configuration selection procedure

PCC uplink channel, channel bandwidth, modulation and RB configurations were selected based on section C)3)b)ii) of KDB 941225 D05v01r02. The downlink PCC channel was paired with the selected PCC uplink channel according to normal configurations without carrier aggregation.

To maximize aggregation bandwidth, highest channel bandwidth available for that CA combination was selected for SCC. For inter-band CA, the SCC downlink channels were selected near the middle of their transmission bands. For contiguous intra-band CA, the downlink channel spacing between the component carriers was set to multiple of 300 kHz less than the nominal channel spacing defined in section 5.4.1A of 3GPP TS 36.521. For non-contiguous intra-band CA, the downlink channel spacing between the component carriers was set to be larger than the nominal channel spacing between the component carriers was set to be larger than the nominal channel spacing between the component carriers.

All selected PCC and SCC(s) remained fully within the uplink/downlink transmission band of the respective component carrier.

When a device supports LTE capabilities with overlapping transmission frequency ranges, the standalone powers from the band with a larger transmission frequency range can be used to select measurement configurations for the band with the fully covered transmission frequency range.

-

E.3 LTE DL Carrier Aggregation Conducted Powers

Below downlink CA configurations were determined based on Manufacturer's information.

	Table E.3.1 CA B	W Class	
Class	ATBC		Maximum
Class	NRB.agg	MHz	number of CC
A	N ≤ 100	20	1
В	25 < N ≤ 100	20	2
С	100 < N ≤ 200	40	2
D	$200 < N \le 300$	60	3
E	300 < N ≤ 400	80	4
F	400 < N ≤ 500	100	5
1	700 < N ≤ 800	160	8

lucilities.	200	Supported Chann	el Bandwidth [MHz]	Destriction	Completely Covered by
Index	200	CC1	CC2	Restriction	Measurement Superset
2CC #1	CA_4A-7A (0)	5, 10	5, 10, 15, 20		No
2CC #2	CA_4A-7A (1)	5, 10, 15, 20	5, 10, 15, 20		No
2CC #3	CA_5A-7A (0)	1.4, 3, 5, 10	10, 15, 20		No
2CC #4	CA_5A-7A (1)	5, 10	10, 15, 20		No
2CC #5	CA_7A-7A (0)	5, 10, 15, 20	10, 15, 20		No
2CC #6	CA_7A-7A (1)	5, 10, 15, 20	5, 10, 15, 20		No
2CC #7	CA_7A-7A (2)	5, 10, 15, 20	5, 10		No
2CC #8	CA_7A-7A (3)	10, 15, 20	10, 15, 20		No
2CC #9	CA_7C (0)	15, 20	15, 20		No
2CC #10	CA_7C (1)	10, 15, 20	10, 15, 20		No
2CC #11	CA_7C (2)	15, 20	10, 15, 20		No
2CC #12	CA_66A-66A (0)	5, 10, 15, 20	5, 10, 15, 20		No
2CC #13	CA_66B (0)	5, 10, 15	5, 10, 15		No
2CC #14	CA 66C (0)	5, 10, 15, 20	5, 10, 15, 20		No

Table E.3.3 LTE Band 4 as PCC

PCC									SCC				Power		
Combination	PCC Band	PCC BW (MHz)	PCC (UL) CH.	PCC (UL) Freq. (MHz)	Mod.	PCC UL# RB	PCC UL RB Offset	PCC (DL) CH.	PCC (DL) Freq. (MHz)	SCC Band	SCC BW (MHz)	SCC (DL) CH.	SCC (DL) Freq. (MHz)	LTE Tx. Power with DL CA Enabled (dBm)	LTE Single Carrier Tx. Power (dBm)
CA_4A-7A (0)	LTE B4	10	20000	1715.0	QPSK	1	25	2000	2115.0	LTE B7	20	3100	2655.0	23.42	23.43
CA_4A-7A (1)	LTE B4	20	20175	1732.5	QPSK	1	50	2175	2132.5	LTE B7	20	3100	2655.0	23.45	23.46

	Table E.3.4 LTE Band 5 as PCC														
PCC								SCC				Power			
Combination	PCC Band	PCC BW (MHz)	PCC (UL) CH.	PCC (UL) Freq. (MHz)	Mod.	PCC UL# RB	PCC UL RB Offset	PCC (DL) CH.	PCC (DL) Freq. (MHz)	SCC Band	SCC BW (MHz)	SCC (DL) CH.	SCC (DL) Freq. (MHz)	LTE Tx. Power with DL CA Enabled (dBm)	LTE Single Carrier Tx. Power (dBm)
CA_5A-7A (0)	LTE B5	10	20525	836.5	QPSK	1	25	2525	881.5	LTE B7	20	3100	2655.0	24.85	24.87
CA_5A-7A (1)	LTE B5	10	20525	836.5	QPSK	1	25	2525	881.5	LTE B7	20	3100	2655.0	24.85	24.87

						Tab	ole E.3.5 LT	E Band 7	as PCC						
	PCC										S	CC		Power	
Combination	PCC Band	PCC BW (MHz)	PCC (UL) CH.	PCC (UL) Freq. (MHz)	Mod.	PCC UL# RB	PCC UL RB Offset	PCC (DL) CH.	PCC (DL) Freq. (MHz)	SCC Band	SCC BW (MHz)	SCC (DL) CH.	SCC (DL) Freq. (MHz)	LTE Tx. Power with DL CA Enabled (dBm)	LTE Single Carrier Tx. Power (dBm)
CA_4A-7A (0)	LTE B7	20	20850	2510.0	QPSK	1	50	2850	2630.0	LTE B4	10	2175	2132.5	23.58	23.60
CA_4A-7A (1)	LTE B7	20	20850	2510.0	QPSK	1	50	2850	2630.0	LTE B4	20	2175	2132.5	23.57	23.60
CA_5A-7A (0)	LTE B7	20	20850	2510.0	QPSK	1	50	2850	2630.0	LTE B5	10	2525	881.5	23.58	23.60
CA_5A-7A (1)	LTE B7	20	20850	2510.0	QPSK	1	50	2850	2630.0	LTE B5	10	2525	881.5	23.58	23.60
CA_7A-7A (0)	LTE B7	20	20850	2510.0	QPSK	1	50	2850	2630.0	LTE B7	20	3350	2680.0	23.56	23.60
CA_7A-7A (1)	LTE B7	20	20850	2510.0	QPSK	1	50	2850	2630.0	LTE B7	20	3350	2680.0	23.56	23.60
CA_7A-7A (2)	LTE B7	20	20850	2510.0	QPSK	1	50	2850	2630.0	LTE B7	10	3400	2685.0	23.54	23.60
CA_7A-7A (3)	LTE B7	20	20850	2510.0	QPSK	1	50	2850	2630.0	LTE B7	20	3350	2680.0	23.56	23.60
CA_7C (0)	LTE B7	20	20850	2510.0	QPSK	1	50	2850	2630.0	LTE B7	20	3048	2649.8	23.52	23.60
CA_7C (1)	LTE B7	20	20850	2510.0	QPSK	1	50	2850	2630.0	LTE B7	20	3048	2649.8	23.52	23.60
CA_7C (2)	LTE B7	20	20850	2510.0	QPSK	1	50	2850	2630.0	LTE B7	20	3048	2649.8	23.52	23.60

						Tab	le E.3.6 LT	E Band 66	as PCC						
PCC								SCC			Power				
Combination	PCC Band	PCC BW (MHz)	PCC (UL) CH.	PCC (UL) Freq. (MHz)	Mod.	PCC UL# RB	PCC UL RB Offset	PCC (DL) CH.	PCC (DL) Freq. (MHz)	SCC Band	SCC BW (MHz)	SCC (DL) CH.	SCC (DL) Freq. (MHz)	LTE Tx. Power with DL CA Enabled (dBm)	LTE Single Carrier Tx. Power (dBm)
CA_66A-66A (0)	LTE B66	20	132072	1720.0	QPSK	1	50	66536	2120.0	LTE B66	20	67036	2170.0	23.40	23.41
CA_66B (0)	LTE B66	15	132597	1772.5	QPSK	1	36	67061	2172.5	LTE B66	5	66968	2163.2	23.28	23.31
CA_66C (0)	LTE B66	20	132072	1720.0	QPSK	1	50	66536	2120.0	LTE B66	20	66734	2139.8	23.38	23.41



Note(s):

The device only supports downlink Carrier Aggregation. Uplink Carrier Aggregation is not supported. The DL carrier aggregation powers were measured according to the April 2018 TCB Workshop Notes (LTE Carrier Aggregation) and FCC KDB Publication 941225 D05Av01102, no further power measurements and SAR measurements are required for DL carrier aggregation configurations when the average output power with downlink only carrier aggregation active is not more than 0.25 dB higher than the average output power with downlink carrier aggregation combinations, PCC uplink channel was selected based on section C.3(b)(ii) of KDB 941225 D05Av01r02.




Figure E.3.1 DL 4CA Power Measurement Setup



APPENDIX F. – Description of Test Equipment

Dt&C

F.1 SAR Measurement Setup

Measurements are performed using the DASY5 automated dosimetric assessment system. The DASY5 is made by Schmid & Partner Engineering AG (SPEAG) in Zurich, Switzerland and consists of high precision robotics system (Staubli), robot controller, desktop computer, near-field probe, probe alignment sensor, and the generic twin phantom containing the brain equivalent material. The robot is a six-axis industrial robot performing precise movements to position the probe to the location (points) of maximum electromagnetic field (EMF) (see Fig. F.1.1).

A cell controller system contains the power supply, robot controller each pendant (Joystick), and a remote control used to drive the robot motors. The PC consists of the Intel Core i7-2600/i7-3770/i7-4770 3.40 GHz desktop computer with Windows 7 system and SAR Measurement Software DASY5,A/D interface card, monitor, mouse, and keyboard. The Staubli Robotis connected to the cell controller to allow software manipulation of the robot. A data acquisition electronic (DAE) circuit that performs the signal amplification, signal multiplexing, AD-conversion, offset measurements, mechanical surface detection, collision detection, etc. is connected to the Electro-optical coupler (EOC). The EOC performs the conversion from the optical into digital electric signal of the DAE and transfers data to the PC plug-in card.




Figure F.1.1 SAR Measurement System Setup

The DAE4 consists of a highly sensitive electrometer-grade preamplifier with auto-zeroing, a channel and gain-switching multiplexer, a fast 16 bit AD-converter and a command decoder and control logic unit. Transmission to the PC-card is accomplished through an optical downlink for data and status information and an optical uplink for commands and clock lines. The mechanical probe mounting device includes two different sensor systems for frontal and sidewise probe contacts. They are also used for mechanical surface detection and probe collision detection. The robot uses its own controller with a built in VME-bus computer. The system is described in detail.

F.2 Probe Specification

Calibration	In air from 10 MHz to 6 GHz In brain and muscle simulating tissue at Frequencies of 750 MHz, 835 MHz, 900 MHz, 1750 MHz, 1900 MHz, 2300 MHz, 2450 MHz, 2600 MHz, 3500 MHz, 3700 MHz, 5200 MHz, 5300 MHz, 5500 MHz, 5600 MHz, 5800 MHz / 2450 MHz, 2600 MHz, 5200 MHz, 5300 MHz, 5500 MHz, 5600 MHz, 5800 MHz / 2450 MHz, 2600 MHz, 3500 MHz, 3700 MHz, 5200 MHz, 5300 MHz, 5500 MHz, 5600 MHz, 5800 MHz						
Frequency	10 MHz to 6 GHz						
Linearity	± 0.2 dB(30 MHz to	6 GHz)					
Dynamic	10 µW/g to > 100 r	mW/g		547			
Range	Linearity :	±0.2dB		A-BEAM			
Dimensions	Overall length :	337 mm		Figure F.2.1 Triangular Probe Configuratior	ıs		
Tip length	20 mm						
Body diameter	12 mm			DT .			
Tip diameter	2.5 mm			Ø			
Distance from pr	obe tip to sensor o	center	1.0 mm				
Application	SAR Dosimetry Te Compliance tests of	-	phones				

Figure F.2.2 Probe Thick-Film Technique



DAE System

The SAR measurements were conducted with the dosimetric probe EX3DV4 designed in the classical triangular configuration(see F.2.1) and optimized for dosimetric evaluation. The probe is constructed using the thick film technique; with printed resistive lines on ceramic substrates. The probe is equipped with an optical multitier line ending at the front of the probe tip. It is connected to the EOC box on the robot arm and provides an automatic detection of the phantom surface. Half of the fibers are connected to a pulsed infrared transmitter, the other half to a synchronized receiver. As the probe approaches the surface, the reflection from the surface produces a coupling from the transmitting to the receiving fibers. This reflection increases first during the approach, reaches maximum and then decreases. If the probe is flatly touching the surface, the coupling is zero. The distance of the coupling maximum to the surface is independent of the surface reflectivity and largely independent of the surface to probe angle. The DASY5 software reads the reflection during a software approach and looks for the maximum using a 2nd order fitting. The approach is stopped at reaching the maximum.



F.3 E-Probe Calibration Process

Dosimetric Assessment Procedure

Each probe is calibrated according to a dosimetric assessment procedure with accuracy better than +/- 10%. The spherical isotropy was evaluated with the procedure and found to be better than +/-0.25dB. The sensitivity parameters (Norm X, Norm Y, Norm Z), the diode compression parameter (DCP) and the conversion factor (Conv F) of the probe is tested.

Free Space Assessment

The free space E-field from amplified probe outputs is determined in a test chamber. This is performed in a TEM cell for frequencies below 1 GHz, and in a waveguide above 1GHz for free space. For the free space calibration, the probe is placed in the volumetric center of the cavity at the proper orientation with the field. The probe is then rotated 360 degrees.

Temperature Assessment *

E-field temperature correlation calibration is performed in a flat phantom filled with the appropriate simulated brain tissue. The measured free space E-field in the medium, correlates to temperature rise in a dielectric medium. For temperature correlation calibration a RF transparent the remits or based temperature probe is used in conjunction with the E-field probe.

$$\mathsf{SAR} = C \frac{\Delta \mathsf{T}}{\Delta t}$$

where:


where:


 Δt = exposure time (30 seconds),

C = heat capacity of tissue (brain or muscle),

 ΔT = temperature increase due to RF exposure.

SAR is proportional to $\Delta T / \Delta t$, the initial rate of tissue heating, before thermal diffusion takes place. Now it's possible to quantify the electric field in the simulated tissue by equating the thermally derived SAR to the E- field;








Figure F.3.2 E-Field and Temperature Measurements at 1800MHz

$$\mathsf{SAR} = \frac{\left|\mathsf{E}\right|^2 \cdot \sigma}{\rho}$$

σ = simulated tissue conductivity,

 ρ = **Tissue** density (1.25 g/cm³ for brain tissue)



F.4 Data Extrapolation

The DASY5 software automatically executes the following procedures to calculate the field units from the microvolt readings at the probe connector. The first step of the evaluation is a linearization of the filtered input signal to account for the compression characteristics of the detector diode. The compensation depends on the input signal, the diode type and the DC-transmission factor from the diode to the evaluation electronics. If the exciting field is pulsed, the crest factor of the signal must be known to correctly compensate for peak power. The formula for each channel can be given like below;

of f	with		= compensated signal of channel i = input signal of channel i	(i=x,y,z) (i=x,y,z)
$V_i = U_i + U_i^2 \cdot \frac{g}{dcp_i}$		cf	= crest factor of exciting field p _i = diode compression point	(DASY parameter) (DASY parameter)

From the compensated input signals the primary field data for each channel can be evaluated:

E-field probes: $E_{i} = \sqrt{\frac{V_{i}}{Norm_{i} \cdot ConvF}}$ with V_{i} = compensated signal of channel i (i = x,y,z) Norm_{i} = sensor sensitivity of channel i (i = x,y,z) $\mu V/(V/m)^{2}$ for E-field probes ConvF = sensitivity of enhancement in solution E_{i} = electric field strength of channel i in V/m

The RSS value of the field components gives the total field strength (Hermetian magnitude):

$$E_{tot} = \sqrt{E_x^2 + E_y^2 + E_z^2}$$

The primary field data are used to calculate the derived field units.

$SAR = E_{tot}^2 \cdot \frac{\sigma}{\rho \cdot 1000}$	with	SAR E _{tox} σ	 = local specific absorption rate in W/g = total field strength in V/m = conductivity in [mho/m] or [Siemens/m] = equivalent tissue density in g/cm³
		P	- educations accord in Gran

The power flow density is calculated assuming the excitation field to be a free space field.

$$P_{pwe} = \frac{E_{bd}^{2}}{3770}$$
 with $P_{pwe} = \text{equivalent power density of a plane wave in W/cm}^{2}$

$$= \text{total electric field strength in V/m}$$

Dt&C

F.5 SAM Twin Phantom

The SAM Twin Phantom V5.0 is constructed of a fiberglass shell integrated in a wooden table. The shape of the shell is based on data from an anatomical study designed to determine the maximum exposure in at least 90% of all users. It enables the dosimetric evaluation of left and right hand phone usage as well as body mounted usage at the flat phantom region. A cover prevents the evaporation of the liquid.

Reference markings on the Phantom allow the complete setup of all predefined phantom positions and measurement grids by manually teaching three points in the robot. (see Fig. F.5.1)

SAM Twin Phantom Specification:

Figure F.5.1 SAM Twin Phantom

Construction	 The shell corresponds to the specifications of the Specific Anthropomorphic Mannequin (SAM) phantom defined in IEEE 1528 and IEC 62209-1. It enables the dosimetric evaluation of left and right hand phone usage as well as body mounted usage at the flat phantom region. A cover prevents evaporation of the liquid. Reference markings on the phantom allow the complete setup of all predefined phantom positions and measurement grids by teaching three points with the robot. Twin SAM V5.0 has the same shell geometry and is manufactured from the same material as Twin SAM V4.0, but has reinforced top structure.
Shell Thickness	2 ± 0.2 mm
Filling Volume	Approx. 25 liters
Dimensions	Length: 1000 mm Width: 500 mm

Height: adjustable feet

Specific Anthropomorphic Mannequin (SAM) Specifications:

The phantom for handset SAR assessment testing is a low-loss dielectric shell, with shape and dimensions derived from the anthropometric data of the 90th percentile adult male head dimensions as tabulated by the US Army. The SAM Twin Phantom shell is bisected along the mid-sagittal plane into right and left halves (see Fig. F.5.2). The perimeter sidewalls of each phantom halves are extended to allow filling with liquid to a depth that is sufficient to minimized reflections from the upper surface. The liquid depth is maintained at a minimum depth of 15cm to minimize reflections from the upper surface.



Figure F.5.2 Sam Twin Phantom shell

F.6 Device Holder for Transmitters

In combination with the Twin SAM Phantom V4.0/V4.0c, V5.0 or ELI4, the Mounting Device enables the rotation of the mounted transmitter device in spherical coordinates. Rotation point is the ear opening point. Transmitter devices can be easily and accurately positioned according to IEC, IEEE, FCC or other specifications. The device holder can be locked for positioning at different phantom sections (left head, right head, flat).

Note: A simulating human hand is not used due to the complex anatomical and geometrical structure of the hand that may produce infinite number of configurations. To produce the worst-case condition (the hand absorbs antenna output power), the hand is omitted during the tests.

Figure F.6.1 Mounting Device







F.7 Automated Test System Specifications

Positioner

Robot Repeatability No. of axis	Stäubli Unimation Corp. Robot Model: TX90XL 0.02 mm 6
Data Acquisition Electro Cell Controller Processor Clock Speed Operating System Data Card	Intel Core i7-2600/ Intel Core i7-3770/Intel Core i7-4770 3.40 GHz Windows 7 Professional DASY5 PC-Board
<u>Data Converter</u> Features Software Connecting Lines	Signal, multiplexer, A/D converter. & control logic DASY5 Optical downlink for data and status info Optical uplink for commands and clock
<u>PC Interface Card</u> Function	24 bit (64 MHz) DSP for real time processing Link to DAE 4 16 bit A/D converter for surface detection system serial link to robot direct emergency stop output for robot
<u>E-Field Probes</u> Model Construction Frequency Linearity	EX3DV4 S/N: 3933, 3916, 3930 Triangular core fiber optic detection system 10 MHz to 6 GHz ± 0.2 dB (30 MHz to 6 GHz)
<u>Phantom</u> Phantom Shell Material Thickness	SAM Twin Phantom (V5.0) Composite 2.0 ± 0.2 mm



Figure F.7.1 DASY5 Test System