


Date: 2019-6-13 Electronics: DAE4 Sn1525 Medium: Body 1900 MHz Medium parameters used: f = 1900 MHz; $\sigma = 1.545$ S/m; $\varepsilon_r = 52.21$; $\rho = 1000$ kg/m³ Ambient Temperature: 22.9°C Liquid Temperature: 22.5°C Communication System: CW Frequency: 1900 MHz Duty Cycle: 1:1 Probe: EX3DV4 – SN7514 ConvF(7.53, 7.53, 7.53)

System Validation/Area Scan (81x121x1):Interpolated grid: dx=1.000 mm, dy=1.000 mm Reference Value = 93.6 V/m; Power Drift = -0.06 dB Fast SAR: SAR(1 g) = 10.4 W/kg; SAR(10 g) = 5.52 W/kg Maximum value of SAR (interpolated) = 12.4 W/kg

System Validation/Zoom Scan (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 93.6 V/m; Power Drift = -0.06 dB Peak SAR (extrapolated) = 18.92 W/kg SAR(1 g) = 10.3 W/kg; SAR(10 g) = 5.44 W/kg Maximum value of SAR (measured) = 12.3 W/kg



0 dB = 12.3 W/kg = 10.90 dB W/kg






Date: 2019-6-16 Electronics: DAE4 Sn1525 Medium: Head 2450 MHz Medium parameters used: f = 2450 MHz; $\sigma = 1.798$ mho/m; $\epsilon_r = 38.89$; $\rho = 1000$ kg/m³ Ambient Temperature: 22.9°C Liquid Temperature: 22.5°C Communication System: CW Frequency: 2450 MHz Duty Cycle: 1:1 Probe: EX3DV4 – SN7514 ConvF(6.95, 6.95, 6.95)

System Validation /Area Scan (61x81x1): Interpolated grid: dx=1.000 mm, dy=1.000 mm Reference Value = 82.01 V/m; Power Drift = -0.01 dB SAR(1 g) = 12.9 W/kg; SAR(10 g) = 6.05 W/kg Maximum value of SAR (interpolated) = 16.1 W/kg

System Validation /Zoom Scan (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 82.01 V/m; Power Drift = -0.01 dB Peak SAR (extrapolated) = 26.76 W/kg SAR(1 g) = 12.7 W/kg; SAR(10 g) = 5.92 W/kg Maximum value of SAR (measured) = 15.9 W/kg



0 dB = 15.9 W/kg = 12.01 dBW/kg






Date: 2019-6-16 Electronics: DAE4 Sn1525 Medium: Body 2450 MHz Medium parameters used: f = 2450 MHz; $\sigma = 1.923$ S/m; $\varepsilon_r = 51.88$; $\rho = 1000$ kg/m³ Ambient Temperature: 22.9°C Liquid Temperature: 22.5°C Communication System: CW Frequency: 2450 MHz Duty Cycle: 1:1 Probe: EX3DV4 – SN7514 ConvF(7.13, 7.13, 7.13)

System Validation/Area Scan (81x101x1): Interpolated grid: dx=1.000 mm, dy=1.000 mm Reference Value = 86.05 V/m; Power Drift = 0.04 dB SAR(1 g) = 12.4 W/kg; SAR(10 g) = 5.72 W/kg Maximum value of SAR (interpolated) = 14 W/kg

System Validation/Zoom Scan (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 86.05 V/m; Power Drift = 0.04 dBPeak SAR (extrapolated) = 24.13 W/kgSAR(1 g) = 12.6 W/kg; SAR(10 g) = 5.89 W/kgMaximum value of SAR (measured) = 14.2 W/kg



0 dB = 14.2 W/kg = 11.52 dB W/kg

Fig.B.10 validation 2450MHz 250mW




Date: 2019-6-17 Electronics: DAE4 Sn1525 Medium: Head 2600 MHz Medium parameters used: f = 2600 MHz; $\sigma = 1.957$ mho/m; $\epsilon_r = 38.58$; $\rho = 1000$ kg/m³ Ambient Temperature: 22.9°C Liquid Temperature: 22.5°C Communication System: CW Frequency: 2600 MHz Duty Cycle: 1:1 Probe: EX3DV4 – SN7514 ConvF(6.92, 6.92, 6.92)

System Validation/Area Scan(81x81x1): Interpolated grid: dx=1.000 mm, dy=1.000 mm Reference Value = 77.36 V/m; Power Drift = 0.02 dB SAR(1 g) = 14.3 W/kg; SAR(10 g) = 6.48 W/kg Maximum value of SAR (interpolated) = 21.9 W/kg

System Validation /Zoom Scan (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 77.36 V/m; Power Drift = 0.02 dB Peak SAR (extrapolated) = 30.51 W/kg SAR(1 g) = 14.1 W/kg; SAR(10 g) = 6.33 W/kg Maximum value of SAR (measured) = 21.7 W/kg



0 dB = 21.7 W/kg = 13.36 dBW/kg






Date: 2019-6-17 Electronics: DAE4 Sn1525 Medium: Body 2600 MHz Medium parameters used: f = 2600 MHz; $\sigma = 2.21$ mho/m; $\epsilon_r = 51.95$; $\rho = 1000$ kg/m³ Ambient Temperature: 22.9°C Liquid Temperature: 22.5°C Communication System: CW Frequency: 2600 MHz Duty Cycle: 1:1 Probe: EX3DV4 – SN7514 ConvF(7.06, 7.06, 7.06)

System Validation /Area Scan(81x121x1):Interpolated grid: dx=1.000 mm, dy=1.000 mm Reference Value = 79.11 V/m; Power Drift = -0.01 dB Fast SAR: SAR(1 g) = 14 W/kg; SAR(10 g) = 6.36 W/kg Maximum value of SAR (interpolated) = 22.1 W/kg

System Validation /Zoom Scan (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm Reference Value = 79.11 V/m; Power Drift = -0.01 dB Peak SAR (extrapolated) = 30.82 W/kg

SAR(1 g) = 13.9 W/kg; SAR(10 g) = 6.28 W/kg

Maximum value of SAR (measured) = 22 W/kg



0 dB = 22 W/kg = 13.42 dB W/kg

Fig.B.12 validation 2600MHz 250mW



The SAR system verification must be required that the area scan estimated 1-g SAR is within 3% of the zoom scan 1-g SAR.

	•		Area scan	Zoom scan		
Date	Band	Position	(1g)	(1g)	Drift (%)	
2010 0 1 1	750	Head	2.09	2.07	0.97	
2019-6-14	750	Body	2.16	2.19	-1.37	
2010 0 12	835	Head	2.32	2.29	1.31	
2019-6-12	835	Body	2.4	2.43	-1.23	
0040.0.45	1750	Head	9.08	9.18	-1.09	
2019-6-15	1750	Body	9.51	9.42	0.96	
2010 0 12	1900	Head	10.3	10.2	0.98	
2019-6-13	1900	Body	10.4	10.3	0.97	
0040.0.40	2450	Head	12.9	12.7	1.57	
2019-6-16	2450	Body	12.4	12.6	-1.59	
2010 0 17	2600	Head	14.3	14.1	1.42	
2019-6-17	2600	Body	14	13.9	0.72	


Table B.1 Comparison between area scan and zoom scan for system verification



ANNEX C SAR Measurement Setup

C.1 Measurement Set-up

The Dasy4 or DASY5 system for performing compliance tests is illustrated above graphically. This system consists of the following items:



Picture C.1SAR Lab Test Measurement Set-up

- A standard high precision 6-axis robot (StäubliTX=RX family) with controller, teach pendant and software. An arm extension for accommodating the data acquisition electronics (DAE).
- An isotropic field probe optimized and calibrated for the targeted measurement.
- A data acquisition electronics (DAE) which performs the signal amplification, signal multiplexing, AD-conversion, offset measurements, mechanical surface detection, collision detection, etc. The unit is battery powered with standard or rechargeable batteries. The signal is optically transmitted to the EOC.
- The Electro-optical converter (EOC) performs the conversion from optical to electrical signals for the digital communication to the DAE. To use optical surface detection, a special version of the EOC is required. The EOC signal is transmitted to the measurement server.
- The function of the measurement server is to perform the time critical tasks such as signal filtering, control of the robot operation and fast movement interrupts.
- The Light Beam used is for probe alignment. This improves the (absolute) accuracy of the probe positioning.
- A computer running WinXP and the DASY4 or DASY5 software.
- Remote control and teach pendant as well as additional circuitry for robot safety such as
- warning lamps, etc.
- The phantom, the device holder and other accessories according to the targeted measurement.



C.2 Dasy4 or DASY5 E-field Probe System

The SAR measurements were conducted with the dosimetric probe designed in the classical triangular configuration and optimized for dosimetric evaluation. The probe is constructed using the thick film technique; with printed resistive lines on ceramic substrates. The probe is equipped with an optical multifiber line ending at the front of the probe tip. It is connected to the EOC box on the robot arm and provides an automatic detection of the phantom surface. Half of the fibers are connected to a pulsed infrared transmitter, the other half to a synchronized receiver. As the probe approaches the surface, the reflection from the surface produces a coupling from the transmitting to the receiving fibers. This reflection increases first during the approach, reaches maximum and then decreases. If the probe is flatly touching the surface, the coupling is zero. The distance of the coupling maximum to the surface is independent of the surface reflectivity and largely independent of the surface to probe angle. The DASY4 or DASY5 software reads the reflection durning a software approach and looks for the maximum using 2nd ord curve fitting. The approach is stopped at reaching the maximum.

Probe Specifications:

Model:	ES3DV3, EX3DV4
Frequency	10MHz — 6.0GHz(EX3DV4)
Range:	10MHz — 4GHz(ES3DV3)
Calibration:	In head and body simulating tissue at
	Frequencies from 835 up to 5800MHz
Linearity:	± 0.2 dB(30 MHz to 6 GHz) for EX3DV4
± 0.2 dB(30 MHz	to 4 GHz) for ES3DV3
DynamicRange:	10 mW/kg — 100W/kg
Probe Length:	330 mm
Probe Tip	
Length:	20 mm
Body Diameter:	12 mm
Tip Diameter:	2.5 mm (3.9 mm for ES3DV3)
Tip-Center:	1 mm (2.0mm for ES3DV3)
Application:SAF	R Dosimetry Testing
	Compliance tests of mobile phones
	Dosimetry in strong gradient fields
Picture C.3E-fiel	d Probe



Picture C.2Near-field Probe



C.3 E-field Probe Calibration

Each E-Probe/Probe Amplifier combination has unique calibration parameters. A TEM cell calibration procedure is conducted to determine the proper amplifier settings to enter in the probe parameters. The amplifier settings are determined for a given frequency by subjecting the probe to a known E-field density (1 mW/cm²) using an RF Signal generator, TEM cell, and RF Power Meter.

The free space E-field from amplified probe outputs is determined in a test chamber. This calibration can be performed in a TEM cell if the frequency is below 1 GHz and inn a waveguide or other methodologies above 1 GHz for free space. For the free space calibration, the probe is placed ©Copyright. All rights reserved by CTTL.



in the volumetric center of the cavity and at the proper orientation with the field. The probe is then rotated 360 degrees until the three channels show the maximum reading. The power density readings equates to 1 mW/cm².

E-field temperature correlation calibration is performed in a flat phantom filled with the appropriate simulated brain tissue. The E-field in the medium correlates with the temperature rise in the dielectric medium. For temperature correlation calibration a RF transparent thermistor-based temperature probe is used in conjunction with the E-field probe.

$$SAR = C \frac{\Delta T}{\Delta t}$$

Where:

 Δt = Exposure time (30 seconds), C = Heat capacity of tissue (brain or muscle), ΔT = Temperature increase due to RF exposure.

$$SAR = \frac{\left|E\right|^2 \cdot \sigma}{\rho}$$

Where: σ = Simulated tissue conductivity, ρ = Tissue density (kg/m³).

C.4 Other Test Equipment

C.4.1 Data Acquisition Electronics(DAE)

The data acquisition electronics consist of a highly sensitive electrometer-grade preamplifier with auto-zeroing, a channel and gain-switching multiplexer, a fast 16 bit AD-converter and a command decoder with a control logic unit. Transmission to the measurement server is accomplished through an optical downlink for data and status information, as well as an optical uplink for commands and the clock.

The mechanical probe mounting device includes two different sensor systems for frontal and sideways probe contacts. They are used for mechanical surface detection and probe collision detection.

The input impedance of the DAE is 200 MOhm; the inputs are symmetrical and floating. Common mode rejection is above 80 dB.



PictureC.4: DAE

No.I18Z60993-SEM03 Page 144 of 214



C.4.2 Robot

The SPEAG DASY system uses the high precision robots (DASY4: RX90XL; DASY5: RX160L) type from Stäubli SA (France). For the 6-axis controller system, the robot controller version from Stäubli is used. The Stäubli robot series have many features that are important for our application:

- High precision (repeatability 0.02mm)
- High reliability (industrial design)
- > Low maintenance costs (virtually maintenance free due to direct drive gears; no belt drives)
- Jerk-free straight movements (brushless synchron motors; no stepper motors)
- > Low ELF interference (motor control fields shielded via the closed metallic construction shields)



Picture C.5DASY 4

Picture C.6DASY 5

C.4.3 Measurement Server

The Measurement server is based on a PC/104 CPU broad with CPU (dasy4: 166 MHz, Intel Pentium; DASY5: 400 MHz, Intel Celeron), chipdisk (DASY4: 32 MB; DASY5: 128MB), RAM (DASY4: 64 MB, DASY5: 128MB). The necessary circuits for communication with the DAE electronic box, as well as the 16 bit AD converter system for optical detection and digital I/O interface are contained on the DASY I/O broad, which is directly connected to the PC/104 bus of the CPU broad.

The measurement server performs all real-time data evaluation of field measurements and surface detection, controls robot movements and handles safety operation. The PC operating system cannot interfere with these time critical processes. All connections are supervised by a watchdog, and disconnection of any of the cables to the measurement server will automatically disarm the robot and disable all program-controlled robot movements. Furthermore, the measurement server is equipped with an expansion port which is reserved for future applications. Please note that this expansion port does not have a standardized pinout, and therefore only devices provided by SPEAG can be connected. Devices from any other supplier could seriously damage the measurement server.



Picture C.7 Server for DASY 4

Picture C.8 Server for DASY 5

C.4.4 Device Holder for Phantom

The SAR in the phantom is approximately inversely proportional to the square of the distance between the source and the liquid surface. For a source at 5mm distance, a positioning uncertainty of ± 0.5 mm would produce a SAR uncertainty of $\pm 20\%$. Accurate device positioning is therefore crucial for accurate and repeatable measurements. The positions in which the devices must be measured are defined by the standards.

The DASY device holder is designed to cope with the different positions given in the standard. It has two scales for device rotation (with respect to the body axis) and device inclination (with respect to the line between the ear reference points). The rotation centers for both scales are the ear reference point (ERP). Thus the device needs no repositioning when changing the angles.

The DASY device holder is constructed of low-loss POM material having the following dielectric parameters: relative permittivity $\ell = 3$ and loss tangent $\delta = 0.02$. The amount of dielectric material has been reduced in the closest vicinity of the device, since measurements have suggested that the influence of the clamp on the test results could thus be lowered.

<Laptop Extension Kit>

The extension is lightweight and made of POM, acrylic glass and foam. It fits easily on the upper part of the Mounting Device in place of the phone positioner. The extension is fully compatible with the Twin-SAM and ELI phantoms.



Picture C.9-1: Device Holder



Picture C.9-2: Laptop Extension Kit



No.I18Z60993-SEM03 Page 146 of 214

C.4.5 Phantom

The SAM Twin Phantom V4.0 is constructed of a fiberglass shell integrated in a table. The shape of the shell is based on data from an anatomical study designed to

Represent the 90th percentile of the population. The phantom enables the dissymmetric evaluation of SAR for both left and right handed handset usage, as well as body-worn usage using the flat phantom region. Reference markings on the Phantom allow the complete setup of all predefined phantom positions and measurement grids by manually teaching three points in the robot. The shell phantom has a 2mm shell thickness (except the ear region where shell thickness increases to 6 mm).

Shell Thickness:2±0. 2 mmFilling Volume:Approx. 25 litersDimensions:810 x 1000 x 500 mm (H x L x W)Available:Special




Picture C.10: SAM Twin Phantom




ANNEX D Position of the wireless device in relation to the phantom

D.1 General considerations

This standard specifies two handset test positions against the head phantom – the "cheek" position and the "tilt" position.




Picture D.1-a Typical "fixed" case handset Picture D.1-b Typical "clam-shell" case handset



Picture D.2 Cheek position of the wireless device on the left side of SAM






Picture D.3 Tilt position of the wireless device on the left side of SAM

D.2 Body-worn device

A typical example of a body-worn device is a mobile phone, wireless enabled PDA or other battery operated wireless device with the ability to transmit while mounted on a person's body using a carry accessory approved by the wireless device manufacturer.




Picture D.4Test positions for body-worn devices

D.3 Desktop device

A typical example of a desktop device is a wireless enabled desktop computer placed on a table or desk when used.

The DUT shall be positioned at the distance and in the orientation to the phantom that corresponds to the intended use as specified by the manufacturer in the user instructions. For devices that employ an external antenna with variable positions, tests shall be performed for all antenna positions specified. Picture8.5 show positions for desktop device SAR tests. If the intended use is not specified, the device shall be tested directly against the flat phantom.





Picture D.5 Test positions for desktop devices



D.4 DUT Setup Photos

Picture D.6



ANNEX E Equivalent Media Recipes

The liquid used for the frequency range of 800-3000 MHz consisted of water, sugar, salt, preventol, glycol monobutyl and Cellulose. The liquid has been previously proven to be suited for worst-case. The Table E.1 shows the detail solution. It's satisfying the latest tissue dielectric parameters requirements proposed by the IEEE 1528 and IEC 62209.

Frequency	835Head	025Dady	1900	1900	2450	2450	5800	5800
(MHz)	osoneau	835Body	Head	Body	Head	Body	Head	Body
Ingredients (% by weight)								
Water	41.45	52.5	55.242	69.91	58.79	72.60	65.53	65.53
Sugar	56.0	45.0	١	١	١	١	١	\
Salt	1.45	1.4	0.306	0.13	0.06	0.18	١	\
Preventol	0.1	0.1	١	١	١	١	١	\
Cellulose	1.0	1.0	١	١	١	١	١	/
Glycol	1	1	44.452	29.96	41.15	27.22	1	١
Monobutyl	١	١	44.452	29.90	41.15	21.22	١	١
Diethylenglycol	1	1	1	N	1	1	17.24	17.24
monohexylether	١	١	١	١	λ	١	17.24	17.24
Triton X-100	١	١	١	١	١	١	17.24	17.24
Dielectric	c=11 E	ε=55.2	c=10.0	ε=53.3	c=20.2	c=50.7	c=25.2	c=10.0
Parameters	ε=41.5		ε=40.0		ε=39.2	ε=52.7	ε=35.3	ε=48.2
Target Value	σ=0.90	σ=0.97	σ=1.40	σ=1.52	σ=1.80	σ=1.95	σ=5.27	σ=6.00

TableE.1: Composition of the Tissue Equivalent Matter

Note: There are a little adjustment respectively for 750, 1750, 2600, 5200, 5300 and 5600 based on the recipe of closest frequency in table E.1.



ANNEX F System Validation

The SAR system must be validated against its performance specifications before it is deployed. When SAR probes, system components or software are changed, upgraded or recalibrated, these must be validated with the SAR system(s) that operates with such components.

	Validation data		Status (OK ar Nat)
			Status (OK or Not)
	,		OK
			OK
Head 1900MHz	Sep.11,2018		OK
		2000 MHz	OK
Head 2100MHz	Sep.11,2018	2100 MHz	OK
Head 2300MHz	Sep.11,2018	2300 MHz	OK
Head 2450MHz	Sep.11,2018	2450 MHz	OK
Head 2600MHz	Sep.12,2018	2600 MHz	OK
Head 3500MHz	Sep.12,2018	3500 MHz	OK
Head 3700MHz	Sep.12,2018	3700 MHz	OK
Head 5200MHz	Sep.12,2018	5250 MHz	OK
Head 5500MHz	Sep.12,2018	5600 MHz	OK
Head 5800MHz	Sep.12,2018	5800 MHz	OK
Body 750MHz	Sep.12,2018	750 MHz	OK
Body 850MHz	Sep.9,2018	835 MHz	OK
Body 900MHz	Sep.9,2018	900 MHz	OK
Body 1750MHz	Sep.9,2018	1750 MHz	OK
Body 1810MHz	Sep.9,2018	1810 MHz	OK
Body 1900MHz	Sep.9,2018	1900 MHz	OK
Body 2000MHz	Sep.13,2018	2000 MHz	OK
Body 2100MHz	Sep.13,2018	2100 MHz	OK
Body 2300MHz	Sep.13,2018	2300 MHz	OK
Body 2450MHz	Sep.13,2018	2450 MHz	OK
Body 2600MHz	Sep.13,2018	2600 MHz	OK
Body 3500MHz	Sep.8,2018	3500 MHz	OK
Body 3700MHz	Sep.8,2018	3700 MHz	OK
Body 5200MHz	Sep.8,2018	5250 MHz	OK
Body 5500MHz	Sep.8,2018	5600 MHz	OK
Body 5800MHz	Sep.8,2018	5800 MHz	OK
	Liquid name Head 750MHz Head 850MHz Head 900MHz Head 1750MHz Head 1750MHz Head 1900MHz Head 2000MHz Head 2000MHz Head 2300MHz Head 2450MHz Head 2600MHz Head 3500MHz Head 5500MHz Head 5500MHz Body 750MHz Body 750MHz Body 1750MHz Body 1750MHz Body 1750MHz Body 1750MHz Body 1900MHz Body 1900MHz Body 2000MHz Body 2000MHz Body 2300MHz Body 2450MHz Body 2450MHz Body 2500MHz Body 3700MHz Body 3700MHz	Liquid name Validation date Head 750MHz Sep.10,2018 Head 850MHz Sep.10,2018 Head 900MHz Sep.10,2018 Head 1750MHz Sep.10,2018 Head 1750MHz Sep.10,2018 Head 1810MHz Sep.10,2018 Head 1900MHz Sep.11,2018 Head 2000MHz Sep.11,2018 Head 2100MHz Sep.11,2018 Head 2300MHz Sep.11,2018 Head 2450MHz Sep.11,2018 Head 2600MHz Sep.12,2018 Head 3700MHz Sep.12,2018 Head 5200MHz Sep.12,2018 Head 5500MHz Sep.12,2018 Head 5500MHz Sep.12,2018 Head 5500MHz Sep.12,2018 Body 750MHz Sep.12,2018 Body 750MHz Sep.9,2018 Body 850MHz Sep.9,2018 Body 1750MHz Sep.9,2018 Body 1900MHz Sep.9,2018 Body 1900MHz Sep.13,2018 Body 2000MHz Sep.13,2018 Body 2100MHz Sep.13,2018 Bo	Liquid name Validation date Frequency point Head 750MHz Sep.10,2018 750 MHz Head 850MHz Sep.10,2018 835 MHz Head 900MHz Sep.10,2018 900 MHz Head 1750MHz Sep.10,2018 1750 MHz Head 1750MHz Sep.10,2018 1810 MHz Head 1810MHz Sep.10,2018 1810 MHz Head 1900MHz Sep.11,2018 1900 MHz Head 2000MHz Sep.11,2018 2000 MHz Head 2100MHz Sep.11,2018 2000 MHz Head 2300MHz Sep.11,2018 2450 MHz Head 2450MHz Sep.12,2018 2600 MHz Head 2600MHz Sep.12,2018 3500 MHz Head 3700MHz Sep.12,2018 3700 MHz Head 5200MHz Sep.12,2018 5600 MHz Head 5500MHz Sep.12,2018 5600 MHz Head 5800MHz Sep.12,2018 5800 MHz Body 750MHz Sep.9,2018 835 MHz Body 850MHz Sep.9,2018 1750 MHz Body 1750MHz Sep.9,2018

Table F.1: System Validation for 7514



ANNEX G Probe Calibration Certificate

Probe 7514 Calibration Certificate

Chmid & Partner Engineering AG eughausstrasse 43, 8004 Zur	Dry of	S C S	Schweizerischer Kalibrierdienst Service suisse d'étalonnage Servizio svizzero di taratura Swiss Calibration Service
ccredited by the Swiss Accredine Swiss Accredine Swiss Accreditation Servi	ice is one of the signatories	to the EA	reditation No.: SCS 0108
ultilateral Agreement for the ient CTTL-BJ (Aud	_		EX3-7514_Aug18
CALIBRATION	CERTIFICATE		
Dbject	EX3DV4 - SN:751	4	
Calibration procedure(s)	QA CAL-25.v6	A CAL-12.v9, QA CAL-14.v4, QA lure for dosimetric E-field probes	CAL-23.v5,
Calibration date:	August 27, 2018		
The measurements and the und	certainties with confidence pro	hal standards, which realize the physical units bability are given on the following pages and a facility: environment temperature $(22 \pm 3)^{\circ}$ C a	are part of the certificate.
The measurements and the une All calibrations have been cond Calibration Equipment used (M Primary Standards	certainties with confidence pro	bability are given on the following pages and	are part of the certificate.
The measurements and the une All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP	certainties with confidence pro lucted in the closed laboratory &TE critical for calibration) ID SN: 104778	bability are given on the following pages and a facility: environment temperature (22 ± 3)°C a Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673)	are part of the certificate. and humidity < 70%. Scheduled Calibration Apr-19
The measurements and the une All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-Z91	certainties with confidence pro lucted in the closed laboratory &TE critical for calibration) ID SN: 104778 SN: 103244	Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672)	are part of the certificate. and humidity < 70%. Scheduled Calibration Apr-19 Apr-19
The measurements and the une All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91	certainties with confidence pro lucted in the closed laboratory &TE critical for calibration) ID SN: 104778 SN: 104244 SN: 103245	bability are given on the following pages and a facility: environment temperature (22 ± 3)°C a Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672)/02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02673)	are part of the certificate. and humidity < 70%. Scheduled Calibration Apr-19 Apr-19 Apr-19
The measurements and the une All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator	certainties with confidence pro lucted in the closed laboratory &TE critical for calibration) ID SN: 104778 SN: 103244 SN: 103245 SN: S5277 (20x)	Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02673)	are part of the certificate. and humidity < 70%. Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19 Apr-19
The measurements and the une All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Reference Probe ES3DV2	certainties with confidence pro lucted in the closed laboratory &TE critical for calibration) ID SN: 104778 SN: 104244 SN: 103245	bability are given on the following pages and a facility: environment temperature (22 ± 3)°C a Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672)/02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02673)	are part of the certificate. and humidity < 70%. Scheduled Calibration Apr-19 Apr-19 Apr-19
The measurements and the une All calibrations have been cond Calibration Equipment used (M Primary Standards Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Reference Probe ES3DV2 DAE4	Certainties with confidence pro lucted in the closed laboratory &TE critical for calibration) ID SN: 104778 SN: 103244 SN: 103245 SN: S5277 (20x) SN: 3013	Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02673) 04-Apr-17 (No. ES3-3013_Dec17)	are part of the certificate. and humidity < 70%. Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19 Apr-19 Dec-18
The measurements and the una All calibrations have been cond Calibration Equipment used (M Primary Standards Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Reference Probe ES3DV2 DAE4 Secondary Standards	Certainties with confidence producted in the closed laboratory &TE critical for calibration) ID SN: 104778 SN: 103244 SN: 103245 SN: 55277 (20x) SN: 3013 SN: 660	Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) 04-Apr-17 (No. 217-02682) 30-Dec-17 (No. ES3-3013_Dec17) 21-Dec-17 (No. DAE4-660_Dec17)	are part of the certificate. and humidity < 70%. Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19 Dec-18 Dec-18
The measurements and the una All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Reference Probe ES3DV2 DAE4 Secondary Standards Power meter E4419B	certainties with confidence pro lucted in the closed laboratory &TE critical for calibration) ID SN: 104778 SN: 103244 SN: 103245 SN: 55277 (20x) SN: 3013 SN: 660 ID	Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) 04-Apr-17 (No. ES3-3013_Dec17) 21-Dec-17 (No. DAE4-660_Dec17) Check Date (in house)	are part of the certificate. and humidity < 70%. Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19 Dec-18 Dec-18 Dec-18 Scheduled Check
The measurements and the une All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Reference Probe ES3DV2 DAE4 Secondary Standards Power meter E4419B Power sensor E4412A	certainties with confidence pro lucted in the closed laboratory &TE critical for calibration) ID SN: 104778 SN: 103244 SN: 103245 SN: 55277 (20x) SN: 3013 SN: 660 ID SN: GB41293874	Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) 04-Apr-17 (No. ES3-3013_Dec17) 21-Dec-17 (No. DAE4-660_Dec17) Check Date (in house) 06-Apr-16 (in house check Jun-18)	are part of the certificate. and humidity < 70%. Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19 Dec-18 Dec-18 Dec-18 Scheduled Check In house check: Jun-20
The measurements and the unexall calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Reference Probe ES3DV2 DAE4 Secondary Standards Power meter E44198 Power sensor E4412A Power sensor E4412A RF generator HP 8648C	certainties with confidence producted in the closed laboratory &TE critical for calibration) ID SN: 104778 SN: 103244 SN: 103245 SN: 3013 SN: 660 ID SN: 6841293874 SN: W141498087 SN: 000110210 SN: U33642U01700	Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02673) 04-Apr-17 (No. ES3-3013_Dec17) 21-Dec-17 (No. DAE4-660_Dec17) Check Date (in house) 06-Apr-16 (in house check Jun-18)	are part of the certificate. and humidity < 70%. Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19 Dec-18 Dec-18 Dec-18 Scheduled Check In house check: Jun-20 In house check: Jun-20 In house check: Jun-20 In house check: Jun-20
The measurements and the une All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Reference Probe ES3DV2 DAE4 Secondary Standards Power meter E4419B Power sensor E4412A Power sensor E4412A RF generator HP 8648C	certainties with confidence producted in the closed laboratory &TE critical for calibration) ID SN: 104778 SN: 103244 SN: 103245 SN: 3013 SN: 660 ID SN: 680 SN: 6841293874 SN: 3013 SN: 6841293874 SN: 3010210	Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) 04-Apr-17 (No. E33-3013_Dec17) 21-Dec-17 (No. DAE4-660_Dec17) Check Date (in house) 06-Apr-16 (in house check Jun-18) 06-Apr-16 (in house check Jun-18)	are part of the certificate. and humidity < 70%. Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19 Dec-18 Dec-18 Scheduled Check In house check: Jun-20 In house check: Jun-20
The measurements and the unexall calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Reference Probe ES3DV2 DAE4 Secondary Standards Power meter E44198 Power sensor E4412A Power sensor E4412A RF generator HP 8648C	certainties with confidence producted in the closed laboratory &TE critical for calibration) ID SN: 104778 SN: 103244 SN: 103245 SN: 3013 SN: 660 ID SN: 6841293874 SN: W141498087 SN: 000110210 SN: U33642U01700	Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02673) 04-Apr-17 (No. ES3-3013_Dec17) 21-Dec-17 (No. DAE4-660_Dec17) Check Date (in house) 06-Apr-16 (in house check Jun-18)	are part of the certificate. and humidity < 70%. Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19 Dec-18 Dec-18 Dec-18 Scheduled Check In house check: Jun-20 In house check: Jun-20 In house check: Jun-20 In house check: Jun-20
The measurements and the unerselectron of the measurements and the unerselectron of the unerselectron of the measurement of the unerselectron	ID ID SN: 104778 SN: 103244 SN: 103245 SN: 3013 SN: 3013 SN: GB41293874 SN: 30110210 SN: US3642U01700 SN: US41080477	Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) 05-Apr-16 (No. DAE4-660_Dec17) 21-Dec-17 (No. DAE4-660_Dec17) Check Date (in house) 06-Apr-16 (in house check Jun-18) 06-Apr-16 (in house check Jun-18) 06-Apr-16 (in house check Jun-18) 04-Aug-99 (in house check Jun-18) 31-Mar-14 (in house check Oct-17)	are part of the certificate. and humidity < 70%. Scheduled Calibration Apr-19 Apr-19 Apr-19 Dec-18 Dec-18 Dec-18 Scheduled Check In house check: Jun-20 In house check: Cot-18
The measurements and the une All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Reference Probe ES3DV2 DAE4 Secondary Standards Power meter E44198 Power sensor E4412A RF generator HP 8648C Network Analyzer E8358A Calibrated by:	certainties with confidence producted in the closed laboratory &TE critical for calibration) ID SN: 104778 SN: 103244 SN: 103245 SN: 55277 (20x) SN: 3013 SN: 660 ID SN: GB41293874 SN: 000110210 SN: US3642U01700 SN: US41080477	Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02673) 05-Apr-16 (in house check Jun-18) 06-Apr-16 (in house check Jun-18) 06-Apr-16 (in house check Jun-18) 04-Aug-99 (in house check Jun-18) 04-Aug-99 (in house check Oct-17) Function	are part of the certificate. and humidity < 70%. Scheduled Calibration Apr-19 Apr-19 Apr-19 Dec-18 Dec-18 Dec-18 Scheduled Check In house check: Jun-20 In house check: Cot-18
The measurements and the uno	ID ID ID SN: 104778 SN: 103244 SN: 103244 SN: 103245 SN: 3013 SN: 660 ID ID SN: 660 SN: 00110210 SN: 00110210 SN: 000110210 SN: US3642U01700 SN: US41080477 SN: US41080477	Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02682) 30-Dec-17 (No. ES3-3013_Dec17) 21-Dec-17 (No. DAE4-660_Dec17) Check Date (in house) 06-Apr-16 (in house check Jun-18) 06-Apr-16 (in house check Jun-18) 06-Apr-16 (in house check Jun-18) 06-Apr-16 (in house check Jun-18) 04-Aug-99 (in house check Jun-18) 31-Mar-14 (in house check Oct-17) Function Laboratory Technician	are part of the certificate. and humidity < 70%. Scheduled Calibration Apr-19 Apr-19 Apr-19 Dec-18 Dec-18 Dec-18 Scheduled Check In house check: Jun-20 In house check: Cot-18



No.I18Z60993-SEM03 Page 153 of 214

Calibration Laboratory of Schmid & Partner **Engineering AG** Zeughausstrasse 43, 8004 Zurich, Switzerland



Schweizerischer Kalibrierdienst S Service suisse d'étalonnage С

Servizio svizzero di taratura S Swiss Calibration Service

Accreditation No.: SCS 0108

Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Glossary:

TSL	tissue simulating liquid
NORMx,y,z	sensitivity in free space
ConvF	sensitivity in TSL / NORMx,y,z
DCP	diode compression point
CF	crest factor (1/duty_cycle) of the RF signal
A, B, C, D	modulation dependent linearization parameters
Polarization φ	φ rotation around probe axis
Polarization 9	9 rotation around an axis that is in the plane normal to probe axis (at measurement center),
	i.e., $\vartheta = 0$ is normal to probe axis
Connector Angle	information used in DASY system to align probe sensor X to the robot coordinate system

Calibration is Performed According to the Following Standards:

- a) IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013 IEC 62209-1, ", "Measurement procedure for the assessment of Specific Absorption Rate (SAR) from hand-
- b) held and body-mounted devices used next to the ear (frequency range of 300 MHz to 6 GHz)", July 2016
- IEC 62209-2, "Procedure to determine the Specific Absorption Rate (SAR) for wireless communication devices c) used in close proximity to the human body (frequency range of 30 MHz to 6 GHz)", March 2010
- d) KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

Methods Applied and Interpretation of Parameters:

- NORMx, y,z: Assessed for E-field polarization $\vartheta = 0$ (f ≤ 900 MHz in TEM-cell; f > 1800 MHz: R22 waveguide). NORMx,y,z are only intermediate values, i.e., the uncertainties of NORMx,y,z does not affect the E²-field uncertainty inside TSL (see below ConvF).
- NORM(f)x,y,z = NORMx,y,z * frequency_response (see Frequency Response Chart). This linearization is implemented in DASY4 software versions later than 4.2. The uncertainty of the frequency response is included in the stated uncertainty of ConvF.
- DCPx,y,z: DCP are numerical linearization parameters assessed based on the data of power sweep with CW signal (no uncertainty required). DCP does not depend on frequency nor media.
- PAR: PAR is the Peak to Average Ratio that is not calibrated but determined based on the signal characteristics
- Ax,y,z; Bx,y,z; Cx,y,z; Dx,y,z; VRx,y,z: A, B, C, D are numerical linearization parameters assessed based on the data of power sweep for specific modulation signal. The parameters do not depend on frequency nor media. VR is the maximum calibration range expressed in RMS voltage across the diode.
- ConvF and Boundary Effect Parameters: Assessed in flat phantom using E-field (or Temperature Transfer Standard for f ≤ 800 MHz) and inside waveguide using analytical field distributions based on power measurements for f > 800 MHz. The same setups are used for assessment of the parameters applied for boundary compensation (alpha, depth) of which typical uncertainty values are given. These parameters are used in DASY4 software to improve probe accuracy close to the boundary. The sensitivity in TSL corresponds to NORMx,y,z * ConvF whereby the uncertainty corresponds to that given for ConvF. A frequency dependent ConvF is used in DASY version 4.4 and higher which allows extending the validity from ± 50 MHz to ± 100 MHz.
- Spherical isotropy (3D deviation from isotropy): in a field of low gradients realized using a flat phantom exposed by a patch antenna.
- Sensor Offset: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required.
- Connector Angle: The angle is assessed using the information gained by determining the NORMx (no uncertainty required).

Certificate No: EX3-7514_Aug18

Page 2 of 39



EX3DV4 - SN:7514

August 27, 2018

Probe EX3DV4

SN:7514

Manufactured: Calibrated: November 13, 2017 August 27, 2018

Calibrated for DASY/EASY Systems (Note: non-compatible with DASY2 system!)

Certificate No: EX3-7514_Aug18

Page 3 of 39



EX3DV4-SN:7514

August 27, 2018

DASY/EASY - Parameters of Probe: EX3DV4 - SN:7514

Basic Calibration Parameters

10 Westerlage	Sensor X	Sensor Y	Sensor Z	Unc (k=2) ± 10.1 %	
Norm $(\mu V/(V/m)^2)^A$	0.46	0.44	0.39		
DCP (mV) ^B	96.5	101.1	97.9		

Modulation Calibration Parameters

UID	Communication System Name		A dB	B dB√μV	С	D dB	VR mV	Unc ^t (k=2)
0	CW	X	0.0	0.0	1.0	0.00	179.1	±3.5 %
	VN 200 200	Y	0.0	0.0	1.0		177.3	
		Z	0.0	0.0	1.0		158.1	

Note: For details on UID parameters see Appendix.

Sensor Model Parameters

	C1 fF	C2 fF	α V ⁻¹	T1 ms.V⁻²	T2 ms.V ⁻¹	T3 ms	T4 V⁻²	T5 V ⁻¹	Т6
Х	31.17	241.1	37.77	3.625	0.025	5.031	0.000	0.325	1.005
Y	34.86	259.7	35.41	7.412	0.000	5.026	0.323	0.291	1.002
Z	33.14	259.6	38.65	3.827	0.264	5.046	0.000	0.373	1.008

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

^A The uncertainties of Nom X,Y,Z do not affect the E²-field uncertainty inside TSL (see Pages 5 and 6).
 ^B Numerical linearization parameter: uncertainty not required.
 ^E Uncertainty is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

Certificate No: EX3-7514_Aug18

Page 4 of 39



EX3DV4- SN:7514

August 27, 2018

DASY/EASY - Parameters of Probe: EX3DV4 - SN:7514

Calibration Parameter Determined in Head Tissue Simulating Media

f (MHz) ^c	Relative Permittivity ^F	Conductivity (S/m) ^F	ConvF X	ConvF Y	ConvF Z	Alpha ^G	Depth ^G (mm)	Unc (k=2)
150	52.3	0.76	12.79	12.79	12.79	0.00	1.00	± 13.3 %
300	45.3	0.87	11.57	11.57	11.57	0.07	1.20	± 13.3 %
450	43.5	0.87	10.68	10.68	10.68	0.14	1.20	± 13.3 %
750	41.9	0.89	9.47	9.47	9.47	0.45	0.89	± 12.0 %
835	41.5	0.90	9.09	9.09	9.09	0.53	0.85	± 12.0 %
900	41.5	0.97	9.03	9.03	9.03	0.49	0.85	± 12.0 %
1450	40.5	1.20	8.24	8.24	8.24	0.35	0.80	± 12.0 %
1640	40.2	1.31	8.22	8.22	8.22	0.38	0.81	± 12.0 %
1750	40.1	1.37	8.10	8.10	8.10	0.36	0.83	± 12.0 %
1810	40.0	1.40	7.82	7.82	7.82	0.35	0.81	± 12.0 %
1900	40.0	1.40	7.73	7.73	7.73	0.31	0.80	± 12.0 %
2000	40.0	1.40	7.64	7.64	7.64	0.30	0.84	± 12.0 %
2100	39.8	1.49	7.57	7.57	7.57	0.27	0.85	± 12.0 %
2300	39.5	1.67	7.42	7.42	7.42	0.31	0.80	± 12.0 %
2450	39.2	1.80	6.95	6.95	6.95	0.38	0.98	± 12.0 %
2600	39.0	1.96	6.92	6.92	6.92	0.25	1.05	± 12.0 %
3500	37.9	2.91	6.78	6.78	6.78	0.79	0.64	± 13.1 %
3700	37.7	3.12	6.61	6.61	6.61	0.42	0.93	± 13.1 %
5200	36.0	4.66	5.05	5.05	5.05	0.40	1.80	± 13.1 %
5250	35.9	4.71	5.02	5.02	5.02	0.40	1.80	± 13.1 %
5300	35.9	4.76	4.99	4.99	4.99	0.40	1.80	± 13.1 %
5500	35.6	4.96	4.59	4.59	4.59	0.40	1.80	± 13.1 %
5600	35.5	5.07	4.41	4.41	4.41	0.40	1.80	± 13.1 %
5750	35.4	5.22	4.47	4.47	4.47	0.40	1.80	± 13.1 %
5800	35.3	5.27	4.42	4.42	4.42	0.40	1.80	± 13.1 %

^C Frequency validity above 300 MHz of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to ± 50 MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below at 150 MHz is ± 50 MHz. Above 5 GHz frequency validity can be extended to ± 110 MHz.
^F At frequencies below 3 GHz, the validity of tissue parameters (ε and σ) can be relaxed to ± 10% if liquid compensation formula is applied to measured SAR values. At frequencies above 3 GHz, the validity of tissue parameters (ε and σ) is restricted to ± 5%. The uncertainty is the RSS of the ConvF uncertainty is the RSS of

The ConvE uncertainty for indicated target tissue parameters. ^G Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than \pm 1% for frequencies below 3 GHz and below \pm 2% for frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.

Certificate No: EX3-7514_Aug18

Page 5 of 39



EX3DV4-SN:7514

August 27, 2018

DASY/EASY - Parameters of Probe: EX3DV4 - SN:7514

Calibration Parameter Determined in Body Tissue Simulating Media

f (MHz) ^c	Relative Permittivity ^F	Conductivity (S/m) ^F	ConvF X	ConvF Y	ConvF Z	Alpha ^G	Depth ^G (mm)	Unc (k=2)
150	61.9	0.80	12.43	12.43	12.43	0.00	1.00	± 13.3 %
300	58.2	0.92	11.39	11.39	11.39	0.05	1.20	± 13.3 %
450	56.7	0.94	11.34	11.34	11.34	0.08	1.20	± 13.3 %
750	55.5	0.96	9.68	9.68	9.68	0.31	1.04	± 12.0 %
835	55.2	0.97	9.47	9.47	9.47	0.46	0.80	± 12.0 %
900	55.0	1.05	9.34	9.34	9.34	0.46	0.83	± 12.0 %
1450	54.0	1.30	8.02	8.02	8.02	0.31	0.80	± 12.0 %
1640	53.7	1.42	7.85	7.85	7.85	0.42	0.81	± 12.0 %
1750	53.4	1.49	7.82	7.82	7.82	0.39	0.83	± 12.0 %
1810	53.3	1.52	7.69	7.69	7.69	0.32	0.92	± 12.0 %
1900	53.3	1.52	7.53	7.53	7.53	0.35	0.83	± 12.0 %
2000	53.3	1.52	7.45	7.45	7.45	0.39	0.80	± 12.0 %
2100	53.2	1.62	7.39	7.39	7.39	0.32	0.94	± 12.0 %
2300	52.9	1.81	7.25	7.25	7.25	0.37	0.85	± 12.0 %
2450	52.7	1.95	7.13	7.13	7.13	0.32	0.97	± 12.0 %
2600	52.5	2.16	7.06	7.06	7.06	0.24	1.10	± 12.0 %
3500	51.3	3.31	6.85	6.85	6.85	0.00	1.00	± 13.1 % _
3700	51.0	3.55	6.75	6.75	6.75	0.00	1.00	± 13.1 %
5200	49.0	5.30	4.59	4.59	4.59	0.50	1.90	<u>± 13.1 %</u>
5250	48.9	5.36	4.54	4.54	4.54	0.50	1.90	± 13.1 %
5300	48.9	5.42	4.49	4.49	4.49	0.50	1.90	± 13.1 %
5500	48.6	5.65	4.17	4.17	4.17	0.50	1.90	± 13.1 %
5600	48.5	5.77	4.00	4.00	4.00	0.50	1.90	± 13.1 %
5750	48.3	5.94	3.98	3.98	3.98	0.50	1.90	± 13.1 %
5800	48.2	6.00	3.94	3.94	3.94	0.50	1.90	± 13.1 %

^C Frequency validity above 300 MHz of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to ± 50 MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below at 150 MHz is ± 50 MHz. Above 5 GHz frequency validity can be extended to ± 110 MHz.
^F At frequencies below 3 GHz, the validity of tissue parameters (e and σ) can be relaxed to ± 10% if liquid compensation formula is applied to measured SAR values. At frequencies above 3 GHz, the validity of tissue parameters (e and σ) is restricted to ± 5%. The uncertainty is the RSS of the ConvF uncertainty for indicated target tissue parameters.
^G Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than ± 1% for frequencies below 3 GHz and below ± 2% for frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.

Certificate No: EX3-7514_Aug18


Page 6 of 39



EX3DV4- SN:7514

August 27, 2018

Frequency Response of E-Field (TEM-Cell:ifi110 EXX, Waveguide: R22)



Uncertainty of Frequency Response of E-field: ± 6.3% (k=2)


Certificate No: EX3-7514_Aug18

Page 7 of 39



EX3DV4- SN:7514

August 27, 2018



Receiving Pattern (ϕ), $\vartheta = 0^{\circ}$

Uncertainty of Axial Isotropy Assessment: ± 0.5% (k=2)


Certificate No: EX3-7514_Aug18

Page 8 of 39



EX3DV4-- SN:7514

August 27, 2018




Certificate No: EX3-7514_Aug18

Page 9 of 39



EX3DV4- SN:7514

August 27, 2018



Certificate No: EX3-7514_Aug18

Page 10 of 39