

# RF System Check Test Report

Measurement performed on April 25

## Device Under Test

| Manufacturer | Model | Dimensions [mm] | Speaker Position [mm] |
|--------------|-------|-----------------|-----------------------|
|              |       |                 |                       |

## Hardware Setup

| Probe Name      | Probe Calibration Date | DAE Name    | DAE Calibration Date |
|-----------------|------------------------|-------------|----------------------|
| EF3DV3 - SN4066 | July 12, 2023          | DAE4 Sn1343 | June 30, 2023        |

## Communication Systems


| Band Name | Communication Systems Name | Channel | Frequency [MHz] |
|-----------|----------------------------|---------|-----------------|
| CD835     | CW                         | 50      | 835.0           |

## Grid Settings

| Extent X [mm] | Extent Y [mm] | Step X [mm] | Step Y [mm] | Distance [mm] |
|---------------|---------------|-------------|-------------|---------------|
| 20.0          | 180.0         | 5.0         | 5.0         | 15.0          |

## Results

| Dipole Type | Dipole Serial Number | E <sub>max</sub> [V/m] | Drift [dB] |
|-------------|----------------------|------------------------|------------|
| CD835       | 1000                 | 105.5                  | -0.03      |



# RF System Check Test Report

Measurement performed on April 25

## Device Under Test

| Manufacturer | Model | Dimensions [mm] | Speaker Position [mm] |
|--------------|-------|-----------------|-----------------------|
|              |       |                 |                       |

## Hardware Setup

| Probe Name      | Probe Calibration Date | DAE Name    | DAE Calibration Date |
|-----------------|------------------------|-------------|----------------------|
| EF3DV3 - SN4066 | July 12, 2023          | DAE4 Sn1343 | June 30, 2023        |

## Communication Systems


| Band Name | Communication Systems Name | Channel | Frequency [MHz] |
|-----------|----------------------------|---------|-----------------|
| CD1880    | CW                         | 50      | 1880.0          |

## Grid Settings

| Extent X [mm] | Extent Y [mm] | Step X [mm] | Step Y [mm] | Distance [mm] |
|---------------|---------------|-------------|-------------|---------------|
| 20.0          | 90.0          | 5.0         | 5.0         | 15.0          |

## Results

| Dipole Type | Dipole Serial Number | E <sub>max</sub> [V/m] | Drift [dB] |
|-------------|----------------------|------------------------|------------|
| CD1880      | 1000                 | 88.05                  | -0.03      |



# RF System Check Test Report

Measurement performed on April 25

## Device Under Test

| Manufacturer | Model | Dimensions [mm] | Speaker Position [mm] |
|--------------|-------|-----------------|-----------------------|
|              |       |                 |                       |

## Hardware Setup

| Probe Name      | Probe Calibration Date | DAE Name    | DAE Calibration Date |
|-----------------|------------------------|-------------|----------------------|
| EF3DV3 - SN4066 | July 12, 2023          | DAE4 Sn1343 | June 30, 2023        |

## Communication Systems


| Band Name | Communication Systems Name | Channel | Frequency [MHz] |
|-----------|----------------------------|---------|-----------------|
| CD2600V3  | CW                         | 50      | 2600.0          |

## Grid Settings

| Extent X [mm] | Extent Y [mm] | Step X [mm] | Step Y [mm] | Distance [mm] |
|---------------|---------------|-------------|-------------|---------------|
| 20.0          | 90.0          | 5.0         | 5.0         | 15.0          |

## Results

| Dipole Type | Dipole Serial Number | E <sub>max</sub> [V/m] | Drift [dB] |
|-------------|----------------------|------------------------|------------|
| CD2600      | 1009                 | 87.2                   | 0.03       |



# RF System Check Test Report

Measurement performed on April 25

## Device Under Test

| Manufacturer | Model | Dimensions [mm] | Speaker Position [mm] |
|--------------|-------|-----------------|-----------------------|
|              |       |                 |                       |

## Hardware Setup

| Probe Name      | Probe Calibration Date | DAE Name    | DAE Calibration Date |
|-----------------|------------------------|-------------|----------------------|
| EF3DV3 - SN4066 | July 12, 2023          | DAE4 Sn1343 | June 30, 2023        |

## Communication Systems

| Band Name | Communication Systems Name | Channel | Frequency [MHz] |
|-----------|----------------------------|---------|-----------------|
| CD3500V3  | CW                         | 50      | 3500.0          |

## Grid Settings

| Extent X [mm] | Extent Y [mm] | Step X [mm] | Step Y [mm] | Distance [mm] |
|---------------|---------------|-------------|-------------|---------------|
| 20.0          | 60.0          | 5.0         | 5.0         | 15.0          |

## Results

| Dipole Type | Dipole Serial Number | E <sub>max</sub> [V/m] | Drift [dB] |
|-------------|----------------------|------------------------|------------|
| CD3500      | 1011                 | 85.4                   | 0.03       |

