

# NORTHWEST EMC

**Medtronic Inc.**

**MyCareLink Monitor, Model 24950B**

**FCC 2.1091:2016**

**Report # MDTR0462.5 Rev 01**



NVLAP Lab Code: 200630-0

*This report must not be used to claim product certification, approval, or endorsement by NVLAP, NIST, or any agency of the federal government of the United States of America. This Report may only be duplicated in its entirety*

# CERTIFICATE OF EVALUATION



Last Date of Evaluation: June 21, 2016  
Medtronic Inc.  
Model: MyCareLink Monitor, Model 24950

## Radio Equipment Evaluation

### Standards

| Specification   | Method  |
|-----------------|---|
| FCC 2.1091:2016 | FCC 447498 D01 General RF Exposure Guidance v06 |

### Results

| Method Clause | Test Description             | Applied | Results | Comments |
|---------------|------------------------------|---------|---------|----------|
| 7.2           | Maximum Permissible Exposure | Yes     | Pass    |          |

### Deviations From Test Standards

None

### Approved By:

Donald Facteau, IT Manager

*Product compliance is the responsibility of the client; therefore, the tests and equipment modes of operation represented in this report were agreed upon by the client, prior to testing. The results of this test pertain only to the sample(s) tested. The specific description is noted in each of the individual sections of the test report supporting this certificate of test. This report reflects only those tests from the referenced standards shown in the certificate of test. It does not include inspection or verification of labels, identification, marking or user information.*

# REVISION HISTORY

| Revision Number | Description                       | Date    | Page Number |
|-----------------|-----------------------------------|---------|-------------|
| 00              | Corrected MICS antenna gain value | 2/15/17 | 9           |

# ACCREDITATIONS AND AUTHORIZATIONS

---

## United States

---

**FCC** - Designated by the FCC as a Telecommunications Certification Body (TCB). Certification chambers, Open Area Test Sites, and conducted measurement facilities are listed with the FCC.

**A2LA** - Accredited by A2LA to ISO / IEC 17065 as a product certifier. This allows Northwest EMC to certify transmitters to FCC and IC specifications.

**NVLAP** - Each laboratory is accredited by NVLAP to ISO 17025

---

## Canada

---

**IC** - Recognized by Industry Canada as a Certification Body (CB). Certification chambers and Open Area Test Sites are filed with IC.

---

## European Union

---

**European Commission** – Validated by the European Commission as a Conformity Assessment Body (CAB) under the EMC directive and as a Notified Body under the R&TTE Directive.

---

## Australia/New Zealand

---

**ACMA** - Recognized by ACMA as a CAB for the acceptance of test data.

---

## Korea

---

**MSIP / RRA** - Recognized by KCC's RRA as a CAB for the acceptance of test data.

---

## Japan

---

**VCCI** - Associate Member of the VCCI. Conducted and radiated measurement facilities are registered.

---

## Taiwan

---

**BSMI** – Recognized by BSMI as a CAB for the acceptance of test data.

**NCC** - Recognized by NCC as a CAB for the acceptance of test data.

---

## Singapore

---

**IDA** – Recognized by IDA as a CAB for the acceptance of test data.

---

## Israel

---

**MOC** – Recognized by MOC as a CAB for the acceptance of test data.

---

## Hong Kong

---

**OFCA** – Recognized by OFCA as a CAB for the acceptance of test data.

---

## Vietnam

---

**MIC** – Recognized by MIC as a CAB for the acceptance of test data.

---


## SCOPE

---

For details on the Scopes of our Accreditations, please visit:

<http://www.nwemc.com/accreditations/>  
<http://gsi.nist.gov/global/docs/cabs/designations.html>

# FACILITIES



|   |   |  |   |  |   |
|---|---|--|---|--|---|
| <b>California</b><br>Labs OC01-13<br>41 Tesla<br>Irvine, CA 92618<br>(949) 861-8918 | <b>Minnesota</b><br>Labs MN01-08, MN10<br>9349 W Broadway Ave.<br>Brooklyn Park, MN 55445<br>(612)-638-5136 | <b>New York</b><br>Labs NY01-04<br>4939 Jordan Rd.<br>Elbridge, NY 13060<br>(315) 554-8214 | <b>Oregon</b><br>Labs EV01-12<br>22975 NW Evergreen Pkwy<br>Hillsboro, OR 97124<br>(503) 844-4066 | <b>Texas</b><br>Labs TX01-09<br>3801 E Plano Pkwy<br>Plano, TX 75074<br>(469) 304-5255 | <b>Washington</b><br>Labs NC01-05<br>19201 120 <sup>th</sup> Ave NE<br>Bothell, WA 98011<br>(425)984-6600 |
| <b>NVLAP</b>  |   |  |   |  |   |
| NVLAP Lab Code: 200676-0  | NVLAP Lab Code: 200881-0  | NVLAP Lab Code: 200761-0   | NVLAP Lab Code: 200630-0  | NVLAP Lab Code:201049-0  | NVLAP Lab Code: 200629-0  |
| <b>Industry Canada</b>  |   |  |   |  |   |
| 2834B-1, 2834B-3  | 2834E-1   | N/A  | 2834D-1, 2834D-2  | 2834G-1  | 2834F-1   |
| <b>BSMI</b>   |   |  |   |  |   |
| SL2-IN-E-1154R  | SL2-IN-E-1152R  | N/A  | SL2-IN-E-1017   | SL2-IN-E-1158R   | SL2-IN-E-1153R  |
| <b>VCCI</b>   |   |  |   |  |   |
| A-0029  | A-0109  | N/A  | A-0108  | A-0201   | A-0110  |
| <b>Recognized Phase I CAB for ACMA, BSMI, IDA, KCC/RRR, MIC, MOC, NCC, OFCA</b>     |   |  |   |  |   |
| US0158  | US0175  | N/A  | US0017  | US0191   | US0157  |



# PRODUCT DESCRIPTION

## Client and Equipment Under Evaluation (EUT) Information

|                           |                                  |
|---------------------------|----------------------------------|
| <b>Company Name:</b>      | Medtronic Inc.                   |
| <b>Address:</b>           | 710 Medtronic Parkway            |
| <b>City, State, Zip:</b>  | Minneapolis, MN 55432            |
| <b>Test Requested By:</b> | Joel Peltier                     |
| <b>Model:</b>             | MyCareLink Monitor, Model 24950B |
| <b>Evaluation Date:</b>   | June 21, 2016                    |

## Information Provided by the Party Requesting the Evaluation

### Functional Description of the EUT:

The MyCareLink Monitor, Model 24950B, is a Medical instrument designed for remote monitoring of Medtronic implantable medical devices. The MyCareLink Monitor is made up of a base unit and a handset.

The base unit communicates to an implanted device via the MICS/MEDS radio or the BTLE radio, and then relays the downloaded data to the MyCareLink network via the cellular dongle.

The base unit contains a Bluetooth (BT) radio module (Classic & Low-Energy) manufactured by Murata, a MICS/MEDS Radio module (Medtronic proprietary), and a pre-approved cellular dongle (Huawei Technologies Co., Ltd., FCC ID: QISMS2131I-8). The MICS radio was originally certified as part of the previous version of this product under FCC ID: LF524950. The Cellular and Bluetooth radio transceivers are mobile transmitters that each operate through their own antenna. The radios can transmit simultaneously.

### Objective:

To demonstrate compliance of the MyCareLink Monitor, Model 24950B, to the requirements of FCC 2.1091: Mobile Devices.



# MAXIMUM PERMISSIBLE EXPOSURE (MPE)



## OVERVIEW

Human exposure to RF emissions from mobile devices (47 CFR §2.1091) may be evaluated based on the MPE limits adopted by the FCC for electric and magnetic field strength and/or power density, as appropriate, since exposures are assumed to occur at distances of 20 cm or more from persons. ANSI C95.1:2005 + Amd 1:2010 specifies a minimum separation distance of 20 cm for performing reliable field measurements to determine adherence to MPE limits. If the minimum separation distance between a transmitter and nearby persons is more than 20 cm under normal operating conditions, compliance with MPE limits may be determined at such distance from the transmitter. When applicable, operation instructions and prominent warning labels may be used to alert the exposed persons to maintain a specified distance from the transmitter or to limit their exposure durations and usage conditions to ensure compliance. If the use of warning labels on a transmitter is not effective or desirable, the alternative of performing SAR evaluation with the device at its closest range to persons under normal operating conditions may be used. The field strength and power density limits adopted by the FCC are based on whole-body averaged exposure and the assumption of RF field levels relate most accurately to estimating whole-body averaged SAR. This means some local values of exposures exceeding the stated field strength and power density limits may not necessarily imply non-compliance if the spatial average of spatially averaged RF fields over the exposed portions of a person's body does not exceed the limits.

## COMPLIANCE WITH FCC 2.1091

*"Mobile devices that operate in the Cellular Radiotelephone Service, the Personal Communications Services, the Satellite Communications Services, the General Wireless Communications Service, the Wireless Communications Service, the Maritime Services and the Specialized Mobile Radio Service authorized under subpart H of part 22 of this chapter, parts 24, 25, 26 and 27 of this chapter, part 80 of this chapter (ship earth stations devices only) and part 90 of this chapter are subject to routine environmental evaluation for RF exposure prior to equipment authorization or use if they operate at frequencies of 1.5 GHz or below and their effective radiated power (ERP) is 1.5 watts or more, or if they operate at frequencies above 1.5 GHz and their ERP is 3 watts or more. Unlicensed personal communications service devices, unlicensed millimeter wave devices and unlicensed NII devices authorized under §§15.253, 15.255, and 15.257, and subparts D and E of part 15 of this chapter are also subject to routine environmental evaluation for RF exposure prior to equipment authorization or use if their ERP is 3 watts or more or if they meet the definition of a portable device as specified in §2.1093(b) requiring evaluation under the provisions of that section. **All other mobile and unlicensed transmitting devices are categorically excluded from routine environmental evaluation for RF exposure prior to equipment authorization or use, except as specified in §§1.1307(c) and 1.1307(d) of this chapter. Applications for equipment authorization of mobile and unlicensed transmitting devices subject to routine environmental evaluation must contain a statement confirming compliance with the limits specified in paragraph (d) of this section as part of their application.**"*

**The device will only be used with a separation distance between the antenna and the body of the user or nearby persons as shown in the table below and can therefore be considered a mobile transmitter per 47 CFR 2.1091(b).**

## COMPLIANCE WITH FCC KDB 447498 D01 General RF Exposure Guidance v06

"KDB 447498 D01 General RF Exposure Guidance v06" provides the procedures, requirements, and authorization policies for mobile and portable devices.

Devices operating in standalone mobile device exposure conditions may contain a single transmitter or multiple transmitters that do not transmit simultaneously are covered in section 7.1.

Devices containing multiple transmitters capable of simultaneous transmissions are covered in section 7.2.

# MAXIMUM PERMISSIBLE EXPOSURE (MPE)



## LIMITS

Limits for General Population /Uncontrolled Exposure: 47 CFR 1.1310

| Frequency Range<br>(MHz) | Electric Field<br>Strength<br>(V/m) | Magnetic Field<br>Strength<br>(A/m) | Power Density<br>(mW/cm <sup>2</sup> ) | Averaging Time<br>(minutes) |
|--------------------------|-------------------------------------|-------------------------------------|--|-----------------------------|
| 0.3 - 1.34               | 614                                 | 1.63                                | *(100)                                 | 30                          |
| 1.34 - 30                | 824/f                               | 2.19/f                              | *(180/f <sup>2</sup> )                 | 30                          |
| 30 - 300                 | 27.5                                | 0.073                               | 0.2                                    | 30                          |
| 300 - 1500               |                                     |                                     | f/1500                                 | 30                          |
| 1500 - 100000            |                                     |                                     | 1                                      | 30                          |

f = frequency in MHz

\* = Plane-wave equivalent power density

## ASSESSMENT

The exposure level for the radio is evaluated at a 20 cm distance from the radio's transmitting antenna using the general equation:

$$S = \frac{P * G}{4 * \pi * R^2}$$

Where: S = power density (mW/cm<sup>2</sup>)

P = power input to the antenna (mW)

G = numeric power gain relative to an isotropic radiator

R = distance to the center of the radiation of the antenna (20 cm = limit for MPE estimates)

P\*G = EIRP

Solving for S, the maximum power density 20 cm from the transmitting antenna is determined. This level is then compared to the applicable limit for the transmit frequency. If limits were not met at the 20 cm boundary the evaluation distance is increased until the limit is met as shown in the table below.

For co-located radios, the ratio of the calculated level to the limit is determined. The ratios for each co-located radio are summed. If the sum is less than or equal to one, then the device is excluded from testing and is deemed compliant.



# MAXIMUM PERMISSIBLE EXPOSURE (MPE)



The standalone MPE and summed MPE ratios are summarized in the following table:

| Radio     | Transmit Frequency (MHz) | Measured Conducted Output Power (mW) | Duty Cycle | Highest Antenna Gain (dBi) | Minimum Antenna Cable Loss (dB) | Minimum Separation Distance (cm) | Power Density (mW/cm <sup>2</sup> ) | Limit (mW/cm <sup>2</sup> ) | Ratio    |
|-----------|--------------------------|--------------------------------------|------------|----------------------------|---------------------------------|----------------------------------|-------------------------------------|-----------------------------|----------|
| MICS      | 404.9                    | 0.061                                | 1          | -5.2                       | 0                               | 20                               | 0.0000037                           | 0.3                         | 0.000014 |
| Max Ratio |                          |                                      |            |                            |                                 |                                  |                                     |                             | 0.000014 |

| Radio     | Transmit Frequency (MHz) | Measured Radiated Output Power (dBm) | Duty Cycle | Minimum Separation Distance (cm) | Power Density (mW/cm <sup>2</sup> ) | Limit (mW/cm <sup>2</sup> ) | Ratio  |
|-----------|--------------------------|--------------------------------------|------------|----------------------------------|-------------------------------------|-----------------------------|--------|
| Cellular  | 824.2                    | 31.48                                | 0.5        | 20                               | 0.229                               | 0.5                         | 0.4174 |
| Cellular  | 1850.2                   | 32.42                                | 0.5        | 20                               | 0.174                               | 1.0                         | 0.1737 |
| Max Ratio |                          |                                      |            |                                  |                                     |                             | 0.4174 |

| Radio     | Transmit Frequency (MHz) | Measured Conducted Output Power (mW) | Duty Cycle | Highest Antenna Gain (dBi) | Minimum Antenna Cable Loss (dB) | Minimum Separation Distance (cm) | Power Density (mW/cm <sup>2</sup> ) | Limit (mW/cm <sup>2</sup> ) | Ratio   |
|-----------|--------------------------|--------------------------------------|------------|----------------------------|---------------------------------|----------------------------------|-------------------------------------|-----------------------------|---------|
| BT        | 2402                     | 1.97                                 | 1          | 2.8                        | 0                               | 20                               | 0.00075                             | 1.0                         | 0.00075 |
| Max Ratio |                          |                                      |            |                            |                                 |                                  |                                     |                             | 0.00075 |

| Sum of Maximum Ratios | Limit | Compliant |
|-----------------------|-------|-----------|
| 0.4182                | 1     | Yes       |