

RF exposure compliance issues:

This device is only used with a handheld computer such as a palm pilot with no provision to connect to any other type of device. However, the maximum rf EIRP power output from the module is less than 1 milliwatt. If the entire rf power output were absorbed by 1 gram of tissue (not possible considering typical rf circuits), the SAR limit would still not be exceeded. Further, with a separation distance of 20 cm or greater, the MPE limits are well above the potential a 1 milliwatt device is capable of producing.