

**APPENDIX C.**

**Calibration Certificate(s)**

## Calibration Certificate

### Dosimetric E-Field Probe

Type:

**ET3DV6R**

Serial Number:

**1429**

Place of Calibration:

**Zurich**

Date of Calibration:

**April 25, 2002**

Calibration Interval:

**12 months**

Schmid & Partner Engineering AG hereby certifies, that this device has been calibrated on the date indicated above. The calibration was performed in accordance with specifications and procedures of Schmid & Partner Engineering AG.

Wherever applicable, the standards used in the calibration process are traceable to international standards. In all other cases the standards of the Laboratory for EMF and Microwave Electronics at the Swiss Federal Institute of Technology (ETH) in Zurich, Switzerland have been applied.

Calibrated by:

*D. Vella*

Approved by:

*Oliver Kotz*

**Probe ET3DV6R**

**SN:1429**

|                   |                   |
|-------------------|-------------------|
| Manufactured:     | May 7, 2001       |
| Last calibration: | September 4, 2001 |
| Recalibrated:     | April 25, 2002    |

**Calibrated for System DASY3**

## DASY3 - Parameters of Probe: ET3DV6R SN:1429

### Sensitivity in Free Space

|       |   |
|-------|---|
| NormX | <b>2.18</b> $\mu\text{V}/(\text{V}/\text{m})^2$ |
| NormY | <b>2.11</b> $\mu\text{V}/(\text{V}/\text{m})^2$ |
| NormZ | <b>2.33</b> $\mu\text{V}/(\text{V}/\text{m})^2$ |

### Diode Compression

|       |           |    |
|-------|-----------|----|
| DCP X | <b>98</b> | mV |
| DCP Y | <b>98</b> | mV |
| DCP Z | <b>98</b> | mV |

### Sensitivity in Tissue Simulating Liquid

|         |                              |                             |                                       |
|---------|------------------------------|-----------------------------|---------------------------------------|
| Head    | <b>900 MHz</b>               | $\epsilon_r = 41.5 \pm 5\%$ | $\sigma = 0.97 \pm 5\% \text{ mho/m}$ |
| Head    | <b>835 MHz</b>               | $\epsilon_r = 41.5 \pm 5\%$ | $\sigma = 0.90 \pm 5\% \text{ mho/m}$ |
| ConvF X | <b>6.2</b> $\pm 9.5\%$ (k=2) | Boundary effect:            |                                       |
| ConvF Y | <b>6.2</b> $\pm 9.5\%$ (k=2) | Alpha                       | <b>0.81</b>                           |
| ConvF Z | <b>6.2</b> $\pm 9.5\%$ (k=2) | Depth                       | <b>1.52</b>                           |
| Head    | <b>1800 MHz</b>              | $\epsilon_r = 40.0 \pm 5\%$ | $\sigma = 1.40 \pm 5\% \text{ mho/m}$ |
| Head    | <b>1900 MHz</b>              | $\epsilon_r = 40.0 \pm 5\%$ | $\sigma = 1.40 \pm 5\% \text{ mho/m}$ |
| ConvF X | <b>5.0</b> $\pm 9.5\%$ (k=2) | Boundary effect:            |                                       |
| ConvF Y | <b>5.0</b> $\pm 9.5\%$ (k=2) | Alpha                       | <b>0.48</b>                           |
| ConvF Z | <b>5.0</b> $\pm 9.5\%$ (k=2) | Depth                       | <b>2.37</b>                           |

### Boundary Effect


|                       |                              |                                   |             |
|-----------------------|------------------------------|-----------------------------------|-------------|
| Head                  | <b>900 MHz</b>               | Typical SAR gradient: 5 % per mm  |             |
| Probe Tip to Boundary |                              | <b>1 mm</b>                       | <b>2 mm</b> |
| SAR <sub>pe</sub> [%] | Without Correction Algorithm | 7.6                               | 3.8         |
| SAR <sub>pe</sub> [%] | With Correction Algorithm    | 0.0                               | 0.1         |
| Head                  | <b>1800 MHz</b>              | Typical SAR gradient: 10 % per mm |             |
| Probe Tip to Boundary |                              | <b>1 mm</b>                       | <b>2 mm</b> |
| SAR <sub>pe</sub> [%] | Without Correction Algorithm | 11.2                              | 7.4         |
| SAR <sub>pe</sub> [%] | With Correction Algorithm    | 0.2                               | 0.2         |

### Sensor Offset


|                            |            |    |
|----------------------------|------------|----|
| Probe Tip to Sensor Center | <b>2.7</b> | mm |
|----------------------------|------------|----|

## Receiving Pattern ( $\phi$ ), $\theta = 0^\circ$


**f = 30 MHz, TEM cell if110**




**f = 100 MHz, TEM cell if110**





**f = 300 MHz, TEM cell if110**




**f = 900 MHz, TEM cell if110**






**Isotropy Error ( $\phi$ ),  $\theta = 0^\circ$**





## Frequency Response of E-Field

( TEM-Cell:ifi110, Waveguide R22)




## Dynamic Range f(SAR<sub>brain</sub>) ( Waveguide R22 )






## Conversion Factor Assessment



|      |          |                              |                                       |
|------|----------|------------------------------|---------------------------------------|
| Head | 900 MHz  | $\epsilon_r = 41.5 \pm 5\%$  | $\sigma = 0.97 \pm 5\% \text{ mho/m}$ |
| Head | 835 MHz  | $\epsilon_r = 41.5 \pm 5\%$  | $\sigma = 0.90 \pm 5\% \text{ mho/m}$ |
|      | ConvF X  | <b>6.2</b> $\pm 9.5\%$ (k=2) | Boundary effect:                      |
|      | ConvF Y  | <b>6.2</b> $\pm 9.5\%$ (k=2) | Alpha <b>0.81</b>                     |
|      | ConvF Z  | <b>6.2</b> $\pm 9.5\%$ (k=2) | Depth <b>1.52</b>                     |
| Head | 1800 MHz | $\epsilon_r = 40.0 \pm 5\%$  | $\sigma = 1.40 \pm 5\% \text{ mho/m}$ |
| Head | 1900 MHz | $\epsilon_r = 40.0 \pm 5\%$  | $\sigma = 1.40 \pm 5\% \text{ mho/m}$ |
|      | ConvF X  | <b>5.0</b> $\pm 9.5\%$ (k=2) | Boundary effect:                      |
|      | ConvF Y  | <b>5.0</b> $\pm 9.5\%$ (k=2) | Alpha <b>0.48</b>                     |
|      | ConvF Z  | <b>5.0</b> $\pm 9.5\%$ (k=2) | Depth <b>2.37</b>                     |


Conversion Factor Assessment



|      |          |                              |                                       |
|------|----------|------------------------------|---------------------------------------|
| Body | 900 MHz  | $\epsilon_r = 55.0 \pm 5\%$  | $\sigma = 1.05 \pm 5\% \text{ mho/m}$ |
| Body | 835 MHz  | $\epsilon_r = 55.2 \pm 5\%$  | $\sigma = 0.97 \pm 5\% \text{ mho/m}$ |
|      | ConvF X  | <b>5.9</b> $\pm 9.5\%$ (k=2) | Boundary effect:                      |
|      | ConvF Y  | <b>5.9</b> $\pm 9.5\%$ (k=2) | Alpha <b>0.58</b>                     |
|      | ConvF Z  | <b>5.9</b> $\pm 9.5\%$ (k=2) | Depth <b>1.86</b>                     |
| Body | 1800 MHz | $\epsilon_r = 53.3 \pm 5\%$  | $\sigma = 1.52 \pm 5\% \text{ mho/m}$ |
| Body | 1900 MHz | $\epsilon_r = 53.3 \pm 5\%$  | $\sigma = 1.52 \pm 5\% \text{ mho/m}$ |
|      | ConvF X  | <b>4.7</b> $\pm 9.5\%$ (k=2) | Boundary effect:                      |
|      | ConvF Y  | <b>4.7</b> $\pm 9.5\%$ (k=2) | Alpha <b>0.60</b>                     |
|      | ConvF Z  | <b>4.7</b> $\pm 9.5\%$ (k=2) | Depth <b>2.26</b>                     |

## Deviation from Isotropy in HSL

Error ( $\theta, \phi$ ),  $f = 900$  MHz



## Calibration Certificate

### 900 MHz System Validation Dipole

Type:

**D900V2**

Serial Number:

**160**

Place of Calibration:

**Zurich**

Date of Calibration:

**August 29, 2002**

Calibration Interval:

**24 months**

Schmid & Partner Engineering AG hereby certifies, that this device has been calibrated on the date indicated above. The calibration was performed in accordance with specifications and procedures of Schmid & Partner Engineering AG.

Wherever applicable, the standards used in the calibration process are traceable to international standards. In all other cases the standards of the Laboratory for EMF and Microwave Electronics at the Swiss Federal Institute of Technology (ETH) in Zurich, Switzerland have been applied.

Calibrated by:

**D. Vellodi**

Approved by:

**Philippe Käfer**

**DASY**

**Dipole Validation Kit**

**Type: D900V2**

**Serial: 160**

**Manufactured: April 30, 2002**  
**Calibrated: August 29, 2002**

## **1. Measurement Conditions**

The measurements were performed in the flat section of the SAM twin phantom filled with head simulating solution of the following electrical parameters at 900 MHz:

|                        |            |      |
|------------------------|------------|------|
| Relative Dielectricity | 41.7       | ± 5% |
| Conductivity           | 0.97 mho/m | ± 5% |

The DASY3 System (Software version 3.1d) with a dosimetric E-field probe ET3DV6 (SN:1507, Conversion factor 6.5 at 900 MHz) was used for the measurements.

The dipole was mounted on the small tripod so that the dipole feedpoint was positioned below the center marking of the flat phantom section and the dipole was oriented parallel to the body axis (the long side of the phantom). The standard measuring distance was 15mm from dipole center to the solution surface. The included distance holder was used during measurements for accurate distance positioning.

The coarse grid with a grid spacing of 20mm was aligned with the dipole. The 5x5x7 fine cube was chosen for cube integration. Probe isotropy errors were cancelled by measuring the SAR with normal and 90° turned probe orientations and averaging.

The dipole input power (forward power) was 250mW ± 3 %. The results are normalized to 1W input power.

### **2.1. SAR Measurement with DASY3 System**

Standard SAR-measurements were performed according to the measurement conditions described in section 1. The results (see figure supplied) have been normalized to a dipole input power of 1W (forward power). The resulting averaged SAR-values measured with the dosimetric probe ET3DV6 SN:1507 and applying the worst-case extrapolation are:

|  |           |
|--|-----------|
| averaged over 1 cm <sup>3</sup> (1 g) of tissue:   | 11.1 mW/g |
| averaged over 10 cm <sup>3</sup> (10 g) of tissue: | 7.04 mW/g |

### **2.2 SAR Measurement with DASY4 System**

Standard SAR-measurements were performed according to the measurement conditions described in section 1. The results (see figure supplied) have been normalized to a dipole input power of 1W (forward power). The resulting averaged SAR-values measured with the dosimetric probe ET3DV6 SN:1507 and applying the advanced extrapolation are:

|  |           |
|--|-----------|
| averaged over 1 cm <sup>3</sup> (1 g) of tissue:   | 10.4 mW/g |
| averaged over 10 cm <sup>3</sup> (10 g) of tissue: | 6.72 mW/g |

### 3. Dipole Impedance and Return Loss

The impedance was measured at the SMA-connector with a network analyzer and numerically transformed to the dipole feedpoint. The transformation parameters from the SMA-connector to the dipole feedpoint are:

|                      |                 |                                       |
|----------------------|-----------------|---------------------------------------|
| Electrical delay:    | <b>1.401 ns</b> | (one direction)                       |
| Transmission factor: | <b>0.991</b>    | (voltage transmission, one direction) |

The dipole was positioned at the flat phantom sections according to section 1 and the distance holder was in place during impedance measurements.

|                                 |                                |
|---------------------------------|--------------------------------|
| Feedpoint impedance at 900 MHz: | $\text{Re}\{Z\} = 50.4 \Omega$ |
|                                 | $\text{Im}\{Z\} = -5.2 \Omega$ |
| Return Loss at 900 MHz          | <b>-25.7 dB</b>                |

### 4. Measurement Conditions

The measurements were performed in the flat section of the SAM twin phantom filled with body simulating solution of the following electrical parameters at 900 MHz:

|                        |                   |           |
|------------------------|-------------------|-----------|
| Relative Dielectricity | <b>54.8</b>       | $\pm 5\%$ |
| Conductivity           | <b>1.02 mho/m</b> | $\pm 5\%$ |

The DASY3 System (Software version 3.1d) with a dosimetric E-field probe ET3DV6 (SN:1507, Conversion factor 6.2 at 900 MHz) was used for the measurements.

The dipole was mounted on the small tripod so that the dipole feedpoint was positioned below the center marking of the flat phantom section and the dipole was oriented parallel to the body axis (the long side of the phantom). The standard measuring distance was 15mm from dipole center to the solution surface. The included distance holder was used during measurements for accurate distance positioning.

The coarse grid with a grid spacing of 20mm was aligned with the dipole. The 5x5x7 fine cube was chosen for cube integration. Probe isotropy errors were cancelled by measuring the SAR with normal and 90° turned probe orientations and averaging.

The dipole input power (forward power) was 250mW  $\pm 3\%$ . The results are normalized to 1W input power.

### **5.1. SAR Measurement with DASY3 System**

Standard SAR-measurements were performed according to the measurement conditions described in section 4. The results (see figure supplied) have been normalized to a dipole input power of 1W (forward power). The resulting averaged SAR-values measured with the dosimetric probe ET3DV6 SN:1507 and applying the worst-case extrapolation are:

averaged over 1 cm<sup>3</sup> (1 g) of tissue: 11.1 mW/g

averaged over 10 cm<sup>3</sup> (10 g) of tissue: 7.12 mW/g

### **5.2 SAR Measurement with DASY4 System**

Standard SAR-measurements were performed according to the measurement conditions described in section 4. The results (see figure supplied) have been normalized to a dipole input power of 1W (forward power). The resulting averaged SAR-values measured with the dosimetric probe ET3DV6 SN:1507 and applying the advanced extrapolation are:

averaged over 1 cm<sup>3</sup> (1 g) of tissue: 10.1 mW/g

averaged over 10 cm<sup>3</sup> (10 g) of tissue: 6.68 mW/g

## **6. Dipole Impedance and Return Loss**

The dipole was positioned at the flat phantom sections according to section 4 and the distance holder was in place during impedance measurements.

Feedpoint impedance at 900 MHz:  $\text{Re}\{Z\} = 46.7 \Omega$

$\text{Im}\{Z\} = -6.2 \Omega$

Return Loss at 900 MHz -22.8 dB



## **7. Handling**

Do not apply excessive force to the dipole arms, because they might bend. Bending of the dipole arms stresses the soldered connections near the feedpoint leading to a damage of the dipole.

## **8. Design**

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals.

## **9. Power Test**

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

# Validation Dipole D900V2 SN:160, d=15 mm

Frequency: 900 MHz; Antenna Input Power: 250 [mW]

SAM Phantom; Flat Section; Grid Spacing: Dx = 20.0, Dy = 20.0, Dz = 10.0



Probe: ET3DV6 - SN1507; ConvF(6.50,6.50,6.50) at 900 MHz; IEEE 1528 900 MHz:  $\sigma = 0.97$  mho/m  $\epsilon_r = 41.7$   $\rho = 1.00$  g/cm<sup>3</sup>

Cubes (2): Peak: 4.39 mW/g  $\pm 0.03$  dB, SAR (1g): 2.78 mW/g  $\pm 0.02$  dB, SAR (10g): 1.76 mW/g  $\pm 0.02$  dB, (Worst-case extrapolation)

Penetration depth: 11.6 (10.7, 12.8) [mm]



Powerdrift: -0.00 dB

SAR<sub>tot</sub> [mW/g]



# Validation Dipole D900V2 SN:160, d=15 mm

Frequency: 900 MHz; Antenna Input Power: 250 [mW]  
SAM Phantom; Flat Section; Grid Spacing: Dx = 20.0, Dy = 20.0, Dz = 10.0  
Probe: ET3DV6 - SN1507; ConvF(6.50,6.50) at 900 MHz; IEEE 1528 900 MHz:  $\sigma = 0.97 \text{ mho/m}$ ,  $\epsilon_r = 41.7$ ,  $\rho = 1.00 \text{ g/cm}^3$   
Cubes (2): Peak:  $3.90 \text{ mW/g} \pm 0.03 \text{ dB}$ , SAR (1g):  $2.60 \text{ mW/g} \pm 0.02 \text{ dB}$ , SAR (10g):  $1.68 \text{ mW/g} \pm 0.02 \text{ dB}$ , (Advanced extrapolation)  
Penetration depth:  $12.5 (12.2, 12.9) [\text{mm}]$   
Powerdrift:  $-0.00 \text{ dB}$



27 Aug 2002 09:36:10

CH1 S11 1 U FS

1: 50.412  $\Omega$  -5.2148  $\Omega$  33.911 pF

900.000 000 MHz

De1


PRm

Cor

Avg

16

↑



CH2 S11 LOG

5 dB/REF 0 dB


1: -25.675 dB

900.000 000 MHz

PRm



Cor

↑



Validation Dipole D900V2 SN:160, d=15 mm



Frequency: 900 MHz; Antenna Input Power: 250 [mW]  
SAM Phantom; Flat Section; Grid Spacing: Dx = 20.0, Dy = 20.0, Dz = 10.0  
Probe: ET3DV6 - SN1507; ConvF(6,20,6,20,6,20) at 900 MHz; IEEE 1528 900 MHz:  $\sigma = 1.02 \text{ mho/m}$   $\epsilon_r = 54.8$   $\rho = 1.00 \text{ g/cm}^3$   
Cubes (2): Peak:  $4.34 \text{ mW/g} \pm 0.01 \text{ dB}$ , SAR (1g):  $2.78 \text{ mW/g} \pm 0.02 \text{ dB}$ , SAR (10g):  $1.78 \text{ mW/g} \pm 0.02 \text{ dB}$ , (Worst-case extrapolation)  
Penetration depth:  $12.1 (11.2, 13.4) [\text{mm}]$   
Powerdrift:  $-0.01 \text{ dB}$



# Validation Dipole D900V2 SN:160, d=15 mm

Frequency: 900 MHz; Antenna Input Power: 250 [mW]  
SAM Phantom; Flat Section; Grid Spacing: Dx = 20.0, Dy = 20.0, Dz = 10.0  
Probe: ET3DV6 - SN1507; ConvF(6,20,6,20) at 900 MHz; IEEE 1528 900 MHz:  $\sigma = 1.02 \text{ mho/m}$   $\epsilon_r = 54.8$   $\rho = 1.00 \text{ g/cm}^3$   
Cubes (2): Peak:  $3.67 \text{ mW/g} \pm 0.01 \text{ dB}$ , SAR (1g):  $2.53 \text{ mW/g} \pm 0.02 \text{ dB}$ , SAR (10g):  $1.67 \text{ mW/g} \pm 0.02 \text{ dB}$ , (Advanced extrapolation)  
Penetration depth: 13.6 (13.5, 13.7) [mm]  
Powerdrift: -0.01 dB

SAR<sub>Tot</sub> [mW/g]



Mosdie

[CH1] S11 1 U FS

1: 46.680  $\Omega$  -6.1777  $\Omega$  28.525 pF

29 Aug 2002 09:44:25


900.000 000 MHz

Del

PRm

Cor  
Avg  
16

↑



CH2

S11

LOG


5 dB/REF 0 dB

1:-22.802 dB

900.000 000 MHz

PRm  
Cor

↑



START 700.000 000 MHz

STOP 1 100.000 000 MHz

## Calibration Certificate

### 1800 MHz System Validation Dipole

Type:

**D1800V2**

Serial Number:

**230**

Place of Calibration:

**Zurich**

Date of Calibration:

**October 25, 2001**


Calibration Interval:

**24 months**

Schmid & Partner Engineering AG hereby certifies, that this device has been calibrated on the date indicated above. The calibration was performed in accordance with specifications and procedures of Schmid & Partner Engineering AG.

Wherever applicable, the standards used in the calibration process are traceable to international standards. In all other cases the standards of the Laboratory for EMF and Microwave Electronics at the Swiss Federal Institute of Technology (ETH) in Zurich, Switzerland have been applied.

Calibrated by:



Approved by:





**DASY**

**Dipole Validation Kit**

**Type: D1800V2**

**Serial: 230**

**Manufactured: February 26, 1998**

**Calibrated: October 25, 2001**

## **1. Measurement Conditions**

The measurements were performed in the flat section of the new generic twin phantom filled with head simulating glycol solution of the following electrical parameters at 1800 MHz:

|                        |                   |           |
|------------------------|-------------------|-----------|
| Relative Dielectricity | <b>40.7</b>       | $\pm 5\%$ |
| Conductivity           | <b>1.35 mho/m</b> | $\pm 5\%$ |

The DASY3 System (Software version 3.1d) with a dosimetric E-field probe ET3DV6 (SN:1507, Conversion factor 5.57 at 1800 MHz) was used for the measurements.

The dipole was mounted on the small tripod so that the dipole feedpoint was positioned below the center marking of the flat phantom section and the dipole was oriented parallel to the body axis (the long side of the phantom). The standard measuring distance was 10mm from dipole center to the solution surface. The included distance holder was used during measurements for accurate distance positioning.

The coarse grid with a grid spacing of 15mm was aligned with the dipole. The 5x5x7 fine cube was chosen for cube integration. Probe isotropy errors were cancelled by measuring the SAR with normal and 90° turned probe orientations and averaging.

The dipole input power (forward power) was 250mW  $\pm 3\%$ . The results are normalized to 1W input power.

## **2. SAR Measurement**

Standard SAR-measurements were performed with the phantom according to the measurement conditions described in section 1. The results have been normalized to a dipole input power of 1W (forward power). The resulting averaged SAR-values are:

|  |                  |
|--|------------------|
| averaged over 1 cm <sup>3</sup> (1 g) of tissue:   | <b>37.4 mW/g</b> |
| averaged over 10 cm <sup>3</sup> (10 g) of tissue: | <b>19.7 mW/g</b> |

Note: If the liquid parameters for validation are slightly different from the ones used for initial calibration, the SAR-values will be different as well. The estimated sensitivities of SAR-values and penetration depths to the liquid parameters are listed in the DASY Application Note 4: 'SAR Sensitivities'.

### **3. Dipole Impedance and Return Loss**

The impedance was measured at the SMA-connector with a network analyzer and numerically transformed to the dipole feedpoint. The transformation parameters from the SMA-connector to the dipole feedpoint are:

|                      |                 |                                       |
|----------------------|-----------------|---------------------------------------|
| Electrical delay:    | <b>1.213 ns</b> | (one direction)                       |
| Transmission factor: | <b>0.990</b>    | (voltage transmission, one direction) |

The dipole was positioned at the flat phantom sections according to section 1 and the distance holder was in place during impedance measurements.

|                                  |   |
|----------------------------------|---|
| Feedpoint impedance at 1800 MHz: | $\text{Re}\{Z\} = $ <b>49.3 <math>\Omega</math></b> |
|----------------------------------|---|

|  |   |
|--|---|
|  | $\text{Im}\{Z\} = $ <b>-6.2 <math>\Omega</math></b> |
|--|---|

|                         |                |
|-------------------------|----------------|
| Return Loss at 1800 MHz | <b>-24.0dB</b> |
|-------------------------|----------------|

### **4. Measurement Conditions**

The measurements were performed in the flat section of the new generic twin phantom filled with muscle simulating glycol solution of the following electrical parameters at 1800 MHz:

|                        |                   |           |
|------------------------|-------------------|-----------|
| Relative Dielectricity | <b>53.5</b>       | $\pm 5\%$ |
| Conductivity           | <b>1.45 mho/m</b> | $\pm 5\%$ |

The DASY3 System (Software version 3.1d) with a dosimetric E-field probe ET3DV6 (SN:1507, Conversion factor 4.85 at 1800 MHz) was used for the measurements.

The dipole was mounted on the small tripod so that the dipole feedpoint was positioned below the center marking of the flat phantom section and the dipole was oriented parallel to the body axis (the long side of the phantom). The standard measuring distance was 10mm from dipole center to the solution surface. The included distance holder was used during measurements for accurate distance positioning.

The coarse grid with a grid spacing of 15mm was aligned with the dipole. The 5x5x7 fine cube was chosen for cube integration. Probe isotropy errors were cancelled by measuring the SAR with normal and 90° turned probe orientations and averaging.

The dipole input power (forward power) was 250mW  $\pm 3\%$ . The results are normalized to 1W input power.

## **5. SAR Measurement**

Standard SAR-measurements were performed with the phantom according to the measurement conditions described in section 4. The results have been normalized to a dipole input power of 1W (forward power). The resulting averaged SAR-values are:

averaged over 1 cm<sup>3</sup> (1 g) of tissue:                    **40.8 mW/g**

averaged over 10 cm<sup>3</sup> (10 g) of tissue:                    **21.4 mW/g**

Note: If the liquid parameters for validation are slightly different from the ones used for initial calibration, the SAR-values will be different as well. The estimated sensitivities of SAR-values and penetration depths to the liquid parameters are listed in the DASY Application Note 4: ‘SAR Sensitivities’.

## **6. Dipole Impedance and Return Loss**

The dipole was positioned at the flat phantom sections according to section 4 and the distance holder was in place during impedance measurements.

Feedpoint impedance at 1800 MHz:                     $\text{Re}\{Z\} = 44.7 \Omega$

$\text{Im}\{Z\} = -6.5 \Omega$

Return Loss at 1800 MHz                                    **-21.1 dB**

## **7. Handling**



The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals.

Do not apply excessive force to the dipole arms, because they might bend. If the dipole arms have to be bent back, take care to release stress to the soldered connections near the feedpoint; they might come off.

After prolonged use with 40W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

Validation Dipole D1800V2 SN:230, d = 10 mm

Frequency: 1800 MHz; Antenna Input Power: 250 [mW]  
SAM Phantom; Flat Section; Grid Spacing: Dx = 20.0, Dy = 20.0, Dz = 10.0  
Probe: ET3DV6 - SN1507; ConvF(5.57,5.57,5.57) at 1800 MHz; IEEE1528 1800 MHz;  $\sigma = 1.35 \text{ mho/m}$   $\epsilon_r = 40.7$   $\rho = 1.00 \text{ g/cm}^3$   
Cubes (2): Peak: 17.5 mW/g  $\pm 0.02 \text{ dB}$ , SAR (1g): 9.36 mW/g  $\pm 0.01 \text{ dB}$ , SAR (10g): 4.92 mW/g  $\pm 0.02 \text{ dB}$ , (Worst-case extrapolation)  
Penetration depth: 8.5 (7.9, 9.6) [mm]  
Powerdrift: -0.03 dB



24 Oct 2001 16:31:05

S11 1 U F8 J: 49.246  $\Omega$  -6.1252  $\Omega$  14.272 pF 1 800.000 000 MHz


↑

Del

PRm

Cor


↑



CH2 S11 LOG 5 dB/REF 0 dB 1:-24.044 dB 1 800.000 000 MHz

PRm  
Cor

↑





START 1 400.000 000 MHz

STOP 2 200.000 000 MHz

Validation Dipole D1800V2 SN:230, d = 10 mm

Frequency: 1800 MHz; Antenna Input Power: 250 [mW]  
SAM Phantom; Flat Section; Grid Spacing: Dx = 20.0, Dy = 20.0, Dz = 10.0  
Probe: ET3DV6 - SN1507; ConvF(4.85,4.85,4.85) at 1800 MHz; Muscle 1800 MHz;  $\sigma = 1.45 \text{ mho/m}$   $\epsilon_r = 53.5$   $\rho = 1.00 \text{ g/cm}^3$   
Cubes (2): Peak: 19.2 mW/g  $\pm 0.01 \text{ dB}$ , SAR (1g): 10.2 mW/g  $\pm 0.02 \text{ dB}$ , SAR (10g): 5.34 mW/g  $\pm 0.02 \text{ dB}$ , (Worst-case extrapolation)  
Penetration depth: 8.8 (7.9, 10.3) [mm]  
Powerdrift: -0.03 dB



24 Oct 2021 20:24:30

[S11] S11 1 U F3 1: 44.738 n -5.5410 n 13.518 pF 1 800.000 000 MHz


?

Del

PRm

Cor

↑




CH2 S11 LOG S dB/REF 0 dB 1:-21.069 dB 1 800.000 000 MHz

PRm

Cor

↑



START 1 400.000 000 MHz

STOP 2 200.000 000 MHz



1.6 GHz



Validation Dipole D1800V2 SN:230, d = 10 mm

Frequency: 1800 MHz; Antenna Input Power: 250 [mW]  
SAM Phantom; Flat Section; Grid Spacing: Dx = 20.0, Dy = 20.0, Dz = 10.0  
Probe: ET3DV6 - SN1507; ConvF(4.85,4.85,4.85) at 1800 MHz; Muscle 1800 MHz;  $\sigma = 1.45$  mho/m  $\epsilon_r = 53.5$   $\rho = 1.00$  g/cm<sup>3</sup>  
Cubes (2): Peak: 19.2 mW/g  $\pm$  0.01 dB, SAR (1g): 10.2 mW/g  $\pm$  0.02 dB, SAR (10g): 5.34 mW/g  $\pm$  0.02 dB, (Worst-case extrapolation)  
Penetration depth: 8.8 (7.9, 10.3) [mm]  
Powerdrift: -0.03 dB

SAR<sub>Tot</sub> [mW/g]



CH1 S11 1 U F8 1: 44.738 n -6.5410 n 13.515 pF 1 000.000 000 MHz


y

Del

PRm

Cor

↑




CH2 S11 LOG 5 dB/REF 0 dB 1: -21.053 dB 1 000.000 000 MHz

PRm

Cor

↑



START 1 400.000 000 MHz

STOP 2 200.000 000 MHz