



American Telecommunications Certification Body Inc.  
6731 Whittier Ave, McLean, VA 22101

April 21, 2005

RE: Medtronic Inc.

FCC ID: LF537751

After a review of the submitted information, I have a few comments on the above referenced Application.

- 1) Please kindly consider correcting the 175 kHz report to properly discuss peak and average emissions. Information regarding this does not appear to be provided in the Revision B Report.
- 2) Regarding the 9 kHz report, it appears that the extrapolation of this device would be 52.4 dB per decade and not 48 dB per decade as given on page 11. Note that page 22 mentions 50 dB. Please review and correct as necessary.
- 3) The results on page 11 of the 9 kHz report do not appear to match the data on page 22. Please explain and/or correct as necessary.

Timothy R. Johnson  
Examining Engineer

[mailto: tjohnson@AmericanTCB.com](mailto:tjohnson@AmericanTCB.com)

The items indicated above must be submitted before processing can continue on the above referenced application. Failure to provide the requested information may result in application termination. Correspondence should be considered part of the permanent submission and may be viewed from the Internet after a Grant of Equipment Authorization is issued.

Please do not respond to this correspondence using the email reply button. In order for your response to be processed expeditiously, you must submit your documents through the AmericanTCB.com website. Also, please note that partial responses increase processing time and should not be submitted.

Any questions about the content of this correspondence should be directed to the sender.