

## TachNT2 SAR Statement (G0M-1509-5054)

In FCC KDB 447498 General RF Exposure Guidance D01 v05r01, Section 4.2.4, the FCC provides a power threshold, below which implanted devices excluded from SAR measurement or numerical simulation:

"4.2.4 Transmitters implanted in the body of a user

When the aggregate of the maximum power available at the antenna port and radiating structures of an implanted transmitter, under all operating circumstances, is  $\leq 1.0$  mW, SAR test exclusion may be applied. The maximum available output power requirement and worst case operating conditions must be supported by power measurement results and fully justified in a SAR analysis report, in lieu of SAR measurement or numerical simulation, according to design and implementation requirements of the device."

Since both the radiated output power and conducted output power are well below 1.0 mW, this device is excluded from SAR measurement or numerical simulation.

FCC ID: QRITACHNT2 and  
IC: 4708A-TACHNT2 are compliant with FCC RF Exposure requirements.

**Signature:**



**Date: 3/9/2016**

Mark Briesemeister  
Manager Regulatory Affairs  
BIOTRONIK SE & Co. KG  
Woermannkehre 1  
12359 Berlin  
Germany