

RF EXPOSURE REPORT

Applicant	FLIR Commercial Systems, Inc.
Address 9 Townsend West Nashua, NH 03063, USA	

Manufacturer or Supplier	Shenzhen Everbest Machinery Industry Co., Ltd					
Address	19th Building, 5th Region, Baiwangxin Industrial Park, SongBai Rd., Baimang, Xili, Nanshan, Shenzhen China					
Product	Coating Thickness Tester					
Brand Name	EXTECH					
Model	CG304					
Additional Model & Model Difference	N/A					
Date of tests	Sep. 13, 2018 ~ Sep. 29, 2018					

- **KDB 447498 D01**
- **◯** IEEE C95.1

CONCLUSION: The submitted sample was found to **COMPLY** with the test requirement

Tested by Ryan Lu Project Engineer / EMC Department	Approved by Glyn He Supervisor / EMC Department	
Ryan	A	
	Date:Apr. 11, 2019	

This report is governed by, and incorporates by reference, CPS Conditions of Service as posted at the date of issuance of this report at http://www.bureauveritas.com/home/about-us/our-business/cps/about-us/lerms-conditions/and is intended for your exclusive use. Any copying or replication of this report to or for any other person or entity, or use of our name or trademark, is permitted only with our prior written permission. This report sets forth our characteristics of the lot from which a test samples identified herein. The results set forth in this report are not indicative or representative of the quality or characteristics of the lot from which a test sample was taken or any similar or identical product unless specifically and expressly noted. Our report includes all of the tests requested by you and the results thereof based upon the information that you provided to us. Measurement uncertainty is only provided upon requiset for accredited tests. You have 60 days from date of issuance of this report to notify us of any material error or omission caused by our negligence or if you require measurement uncertainty; provided, however, that such notice shall be in writing and shall specifically address the issue you wish to raise. A failure to raise such issue within the prescribed time shall constitute you unqualified acceptance of the completeness of this report, the tests conducted and the correctness of the report contents.



TABLE OF CONTENTS

RELE	ASE CONTROL RECORD	3
1.	CERTIFICATION	4
	RF EXPOSURE DEFINE	
3.	CLASSIFICATION	5
4	SAR TEST EXCLUSION THRESHOLDS	6

Tel: +86 769 8998 2098 Fax: +86 769 8593 1080

Email: customerservice.dg@cn.bureauveritas.com



RELEASE CONTROL RECORD

ISSUE NO.	REASON FOR CHANGE	DATE ISSUED	
FM180913N015	Original release	Nov. 08, 2018	
FM190404N018	Based on the original report FM180913N015 changed the information of applicant, Brand Name and model No., but it doesn't need to be retested.	Apr. 11, 2019	

Tel: +86 769 8998 2098 Fax: +86 769 8593 1080

Email: customerservice.dg@cn.bureauveritas.com



1. CERTIFICATION

FCC ID:	IWK-CG304A		
PRODUCT:	Coating Thickness Tester		
BRAND NAME:	EXTECH		
MODEL NO.: CG304			
ADDITIONAL NO.:	.: N/A		
TEST SAMPLE:	Engineering Sample		
APPLICANT:	FLIR Commercial Systems, Inc.		
STANDARDS:	FCC Part 2 (Section 2.1093)		
	KDB 447498 D01		
	IEEE C95.1		

Tel: +86 769 8998 2098 Fax: +86 769 8593 1080

Email: customerservice.dg@cn.bureauveritas.com



2. RF EXPOSURE DEFINE

The corresponding SAR Exclusion Threshold condition, listed below:

1) The 1-g and 10-g SAR test exclusion thresholds for 100 MHz to 6 GHz at test separation distances ≤ 50 mm are determined by:

[(max. power of channel, including tune-up tolerance, mW)/(min. test separation distance, mm)] $\cdot [\sqrt{f(GHz)}] \le 3.0$ for 1-g SAR and ≤ 7.5 for 10-g extremity SAR,16 where

- > f(GHz) is the RF channel transmit frequency in GHz
- Power and distance are rounded to the nearest mW and mm before calculation
- The result is rounded to one decimal place for comparison

The test exclusions are applicable only when the minimum test separation distance is \leq 50 mm and for transmission frequencies between 100 MHz and 6 GHz. When the minimum test separation distance is < 5 mm, a distance of 5 mm is applied to determine SAR test exclusion.

- 2) At 100 MHz to 6 GHz and for test separation distances > 50 mm, the SAR test exclusion threshold is determined according to the following:
- a) [Threshold at 50 mm in step 1) + (test separation distance 50 mm)-(f(MHz)/150)] mW, at 100MHz to 1500 MHz
- b) [Threshold at 50 mm in step 1) + (test separation distance 50 mm)·10] mW at > 1500 MHz and ≤ 6 GHz
- 3) At frequencies below 100 MHz, the following may be considered for SAR test exclusion.
 - a) The threshold at the corresponding test separation distance at 100 MHz in step 2) is multiplied by $[1 + \log(100/f(MHz))]$ for test separation distances > 50 mm and < 200 mm.
 - b) The threshold determined by the equation in a) for 50 mm and 100 MHz is multiplied by ½ for test separation distances ≤ 50 mm.
 - c) SAR measurement procedures are not established below 100 MHz. When SAR test exclusion cannot be applied, a KDB inquiry is required to determine SAR evaluation requirements for any test results to be acceptable.

3. CLASSIFICATION

The antenna of this product, under normal use condition, is at less than 20cm away from the body of the user. So, this device is classified as **Portable Device.**

Tel: +86 769 8998 2098 Fax: +86 769 8593 1080

Email: <u>customerservice.dg@cn.bureauveritas.com</u>



4. SAR TEST EXCLUSION THRESHOLDS

The tuned conducted Average Power (declared by client)

Mode	Frequency (MHz)	Power (dBm) Tole		Lower Tolerance (dBm)	Upper Tolerance (dBm)
BT-LE(GFSK)	2402-2480	-6	+-1	-7	-5

The measured conducted Average Power

Mode	Frequency (MHz)	Averaged Power (dBm)	
BT-LE(GFSK)	2402	-6.46	

SAR Test Exclusion Thresholds

Frequency (MHz)	Maximum source-based time averaged conducted output power (dBm)	Minimum separation distance (mm)	Result of Eq. 1	Limit for 1-g SAR	Limit for 10-g extremity SAR	Verdict
2402-2480	-5	5	0.098	3.0	7.5	Exempt from SAR

Conclusion

Therefore this device complies with FCC's RF radiation exposure limits for general population without SAR evaluation.

Tel: +86 769 8998 2098

Fax: +86 769 8593 1080

Email: <u>customerservice.dg@cn.bureauveritas.com</u>

Page 6 of 6