


REVISION			
REV	DESCRIPTION	ECO NUMBER	DATE
01	INITIAL RELEASE	20-00720	10/12/2020
02	ADDED FCC INFO	20-00XXX	5/11/2020

Added

FCC INFO TEXT ADDED



- (CTQ-1)
1. MATERIAL: 5 MIL LEXAN 8835V FILM WITH 3M 200MP REAR ADHESIVE (2 MIL). MATERIAL CERTIFICATION TO BE PROVIDED WITH EACH LOT.
 2. SEE ART FILE FOR COMPLETE GEOMETRY DEFINITION.
 3. CTQ ARE CRITICAL INSPECTION DIMENSIONS FOR SUPPLIER (2).
 4. UNLESS OTHERWISE SPECIFIED GENERAL TOLERANCES APPLY: CUT-TO-CUT & PRINT-TO CUT .020, PRINT TO PRINT .010.
 5. THE PARTS SHALL BE PACKAGED IN A MANNER THAT PREVENTS DAMAGE OR CONTAMINATION FROM FOREIGN MATERIAL. EACH MANUFACTURING LOT SHALL BE LABELED IN ACCORDANCE WITH INSPIRE SPEC 700-006-000.
 6. UPON APPROVAL BY INSPIRE MEDICAL SYSTEMS OF THE INITIAL DESIGN, ANY PROCESS CHANGES, DESIGN CHANGES, PACKAGING CHANGES, RAW MATERIAL CHANGES, OR DEVIATIONS CONSIDERED BY THE MANUFACTURER MUST BE SUBMITTED TO INSPIRE IN WRITING FOR PRE-APPROVAL. THE INFORMATION SUBMITTED SHALL INCLUDE A COMPLETE DESCRIPTION OF THE CHANGE AND THE EFFECT THE CHANGE WILL HAVE ON ALL THE CHARACTERISTICS OF THE DEVICE/MATERIAL. UPON REQUEST, THE MANUFACTURER SHALL SUBMIT SAMPLES OF THE PROPOSED DEVICE/MATERIAL FOR EVALUATION AND APPROVAL BY INSPIRE MEDICAL SYSTEMS.

PROPERTY AND CONFIDENTIALITY NOTICE:
This drawing is the property of Inspire Medical Systems, Inc. and is to be used for the sole purpose of manufacturing the part or parts shown hereon. No part of this drawing may be reproduced in any form or by any means without the written permission of Inspire. This permission is strictly prohibited.

Inspire Medical Systems, Inc.
Golden Valley, MN USA

TITLE
LABEL, TEST POINT COVER,
MODEL 2580

ECO NUMBER	DATE	APPROVED BY	DATE
R. ROBERTS	10/12/2020	K. LARSON	10/12/2020
B	200-403-000	02	1

IMS P/N	REV	DESCRIPTION
200-403-001	02	LABEL, TEST POINT COVER, MODEL 2580

WAS "01"

WAS "01"